



### Analytical Summary Report

Analysis Method: 9045C  
Parameter: pH  
Run Number: LB91645  
BalanceID: WC SC-4

Analyst By : Jignesh  
Supervisor Review By : apatel  
Temp.: 24.2 °C  
SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/16/2017	16:35
2	CAL2	1	Water			7.01	11/16/2017	16:36
3	CAL3	1	Water			10.02	11/16/2017	16:37
4	ICV	1	Water			7.00	11/16/2017	16:39
5	CCV1	1	Water			2.01	11/16/2017	17:10
6	I6513-01	1	Solid	20.02	20	7.00	11/16/2017	17:11
7	I6513-01DUP	1	Solid	20.03	20	7.01	11/16/2017	17:12
8	CCV2	1	Water			12.01	11/16/2017	17:13

2391645

### WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i6513

WorkList ID : 105588

Date : 11/16/2017 4:25:09 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/15/2017	Solid	I6513-01 -D	pH	Cool 4 deg C	AREC01	J11	SP-1	11/15/2017	9045C

Date/Time 11.16.17 4:25 PM  
 Received by: JA  
 Relinquished by: CS

Date/Time 11.16.17 5:30 PM  
 Received by: CP  
 Relinquished by: JP