

7470A


LB91649

INSTRUMENT ID: CY2

Linear

 μ Abs.:

34248



A= 0.0000e+000

B= 2.8734e-004

C= 6.3324e-002

Rho= 0.9998381

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.012	-0.012	-263	0.000	-263				
0.2	0.200	0.186	-0.014	427	0.0 %	427				
2.5	2.500	2.500	-0.000	8479	0.0 %	8479				
5.0	5.000	4.993	-0.007	17156	0.0 %	17156				
7.5	7.500	7.630	0.130	26333	0.0 %	26333				
10.0	10.000	9.904	-0.096	34248	0.0 %	34248				

LB91649 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Nov 2017 10:02:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	-263	-	0.0000	7470A	PPB	17 Nov 2017 10:19:14	S	Std
0.2 - 1	S0.2	427	-	0.2000	7470A	PPB	17 Nov 2017 10:21:23	S	Std
2.5 - 1	S2.5	8479	-	2.5000	7470A	PPB	17 Nov 2017 10:23:32	S	Std
5.0 - 1	S5	17156	-	5.0000	7470A	PPB	17 Nov 2017 10:25:40	S	Std
7.5 - 1	S7.5	26333	-	7.5000	7470A	PPB	17 Nov 2017 10:27:48	S	Std
10.0 - 1	S10	34248	-	10.0000	7470A	PPB	17 Nov 2017 10:29:56	S	Std
ICV68 - 1	ICV68	13147	3.8409	-7470A	PPB	PPB	17 Nov 2017 10:32:30	U	SMPL
ICB68 - 1	ICB68	-409	-0.0542	-7470A	PPB	PPB	17 Nov 2017 10:34:37	U	SMPL
CCV32 - 1	CCV32	17111	4.9799	-7470A	PPB	PPB	17 Nov 2017 10:36:44	U	SMPL
CCB32 - 1	CCB32	-403	-0.0525	-7470A	PPB	PPB	17 Nov 2017 10:38:50	U	SMPL
CRA - 1	CRA	405	0.1797	-7470A	PPB	PPB	17 Nov 2017 10:40:58	U	SMPL
HighStd - 1	HighStd	35008	10.1224	-7470A	PPB	PPB	17 Nov 2017 10:43:04	U	SMPL
ChkStd - 1	ChkStd	24865	7.2079	-7470A	PPB	PPB	17 Nov 2017 10:45:11	U	SMPL
PB104338BL - 1	PBW	-462	-0.0694	-7470A	PPB	PPB	17 Nov 2017 10:47:20	U	SMPL
PB104338BS - 1	LCSW	13025	3.8059	-7470A	PPB	PPB	17 Nov 2017 10:49:28	U	SMPL
I6501-01 - 1	T71-MW6	-402	-0.0522	-7470A	PPB	PPB	17 Nov 2017 10:51:36	U	SMPL
I6501-01DUP - 1	T71-MW6DUP	-317	-0.0278	-7470A	PPB	PPB	17 Nov 2017 10:53:44	U	SMPL
I6501-01MS - 1	T71-MW6MS	13486	3.9383	-7470A	PPB	PPB	17 Nov 2017 10:55:53	U	SMPL
I6501-01MSD - 1	T71-MW6MSD	13258	3.8728	-7470A	PPB	PPB	17 Nov 2017 10:58:01	U	SMPL
I6501-02 - 1	MW-19	-346	-0.0361	-7470A	PPB	PPB	17 Nov 2017 11:00:08	U	SMPL
CCV33 - 1	CCV33	17321	5.0403	-7470A	PPB	PPB	17 Nov 2017 11:02:15	U	SMPL
CCB33 - 1	CCB33	-412	-0.0551	-7470A	PPB	PPB	17 Nov 2017 11:04:22	U	SMPL
I6501-03 - 1	FB111317	-312	-0.0263	-7470A	PPB	PPB	17 Nov 2017 11:06:29	U	SMPL
I6501-05 - 1	MW-25	-281	-0.0174	-7470A	PPB	PPB	17 Nov 2017 11:08:35	U	SMPL
I6501-06 - 1	MW-23	-300	-0.0229	-7470A	PPB	PPB	17 Nov 2017 11:10:42	U	SMPL
I6501-07 - 1	MW-22D	-297	-0.0220	-7470A	PPB	PPB	17 Nov 2017 11:12:49	U	SMPL
I6501-08 - 1	MW-22S	-295	-0.0214	-7470A	PPB	PPB	17 Nov 2017 11:14:57	U	SMPL
I6501-09 - 1	MW-9	-288	-0.0194	-7470A	PPB	PPB	17 Nov 2017 11:17:04	U	SMPL
I6501-10 - 1	FB111417	-316	-0.0275	-7470A	PPB	PPB	17 Nov 2017 11:19:13	U	SMPL
I6501-01LX5 - 1		-288	-0.0194	-7470A	PPB	PPB	17 Nov 2017 11:26:48	U	SMPL
I6501-01A - 1		14314	4.1762	-7470A	PPB	PPB	17 Nov 2017 11:28:56	U	SMPL
CCV34 - 1	CCV34	17378	5.0566	-7470A	PPB	PPB	17 Nov 2017 11:31:03	U	SMPL
CCB34 - 1	CCB34	-415	-0.0559	-7470A	PPB	PPB	17 Nov 2017 11:33:10	U	SMPL