

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB91660

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP56414
calibration std. phosphate 1 ppm	WP59707
calibration std. phosphate 0.5 ppm	WP59706
calibration std. phosphate 0.3 ppm	WP59705
calibration std. 0 ppm	WP59702
phosphate ICV std.	WP59709
phosphate CCV std.	WP59708
calibration std. phosphate 0.1 ppm	WP59704
calibration std. phosphate 0.05 ppm	WP59703
Combined reagent	WP59712
11 N sulfuric acid	WP57045
Phenolphthalein indicator	WP57038
Sodium hydroxide, 1N	WP57036

Intercept: -0.0006Slope: 0.6516Regression: 0.999822

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	11/17/2017	11:30
2	CAL2	0.05	1	50	50	0.034	11/17/2017	11:30
3	CAL3	0.10	1	50	50	0.066	11/17/2017	11:31
4	CAL4	0.30	1	50	50	0.186	11/17/2017	11:31
5	CAL5	0.50	1	50	50	0.330	11/17/2017	11:32
6	CAL6	1.00	1	50	50	0.651	11/17/2017	11:32



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB91660

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.326	0.501	11/17/2017	11:33
2	ICB		1	50	50	0.000	0.001	11/17/2017	11:34
3	CCV1	0.50	1	50	50	0.329	0.506	11/17/2017	11:34
4	CCB1		1	50	50	0.000	0.001	11/17/2017	12:50
5	PB104371BL		1	50	50	0.000	0.001	11/17/2017	11:35
6	PB104371BS	0.50	1	50	50	0.327	0.503	11/17/2017	11:36
7	I6507-01		1	50	50	0.691	1.061	11/17/2017	11:36
8	I6507-01DUP		1	50	50	0.736	1.130	11/17/2017	11:37
9	I6507-01MS	0.50	1	50	50	0.863	1.325	11/17/2017	11:37
10	I6507-01MSD	0.50	1	50	50	0.860	1.321	11/17/2017	11:38
11	I6507-01		5	50	50	0.193	0.297	11/17/2017	11:55
12	I6507-01DUP		5	50	50	0.205	0.316	11/17/2017	11:55
13	I6507-01MS	0.50	5	50	50	0.272	0.418	11/17/2017	11:56
14	I6507-01MSD	0.50	5	50	50	0.269	0.414	11/17/2017	11:56
15	CCV2	0.50	1	50	50	0.329	0.506	11/17/2017	11:57
16	CCB2		1	50	50	0.000	0.001	11/17/2017	11:57