



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91677

Temp.: 19.3 °C

SLOPE.: 97.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/17/2017	16:40
2	CAL2	1	Water			7.00	11/17/2017	16:41
3	CAL3	1	Water			10.03	11/17/2017	16:42
4	ICV	1	Water			7.00	11/17/2017	16:43
5	CCV1	1	Water			2.01	11/17/2017	16:45
6	I6535-05	1	Water			7.68	11/17/2017	16:46
7	I6535-05DUP	1	Water			7.69	11/17/2017	16:47
8	I6537-02	1	Water			7.80	11/17/2017	16:48
9	CCV2	1	Water			12.01	11/17/2017	16:49

WORKLIST(Hardcopy Internal Chain)

LB 01677

WorkList Name : CORROSIONITY I6535 WorkList ID : 105656 Date : 11/17/2017 4:31:51 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/17/2017	Water	I6535-05-17	Corrosivity	Cool 4 deg C	PSEG01	K61	RT-4252-RT-3025-RT-2342	11/17/2017	9040C
11/17/2017	Water	I6537-02 - 17	Corrosivity	Cool 4 deg C	PSEG01	K61	SV22482L	11/17/2017	9040C

Date/Time 11-17-17 4:30 PM
 Received by: JP
 Relinquished by: JP

Date/Time 11-17-17 6:00 PM
 Received by: JP
 Relinquished by: JP