



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB91678

Temp.: 19.3 °C

SLOPE.: 97.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/17/2017	16:40
2	CAL2	1	Water			7.00	11/17/2017	16:41
3	CAL3	1	Water			10.03	11/17/2017	16:42
4	ICV	1	Water			7.00	11/17/2017	16:43
5	CCV1	1	Water			2.01	11/17/2017	16:45
6	I6535-04	1	Water			7.68	11/17/2017	16:46
7	I6535-04DUP	1	Water			7.69	11/17/2017	16:47
8	I6537-01	1	Water			7.80	11/17/2017	16:48
9	CCV2	1	Water			12.01	11/17/2017	16:49

LB 91678

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i6535

WorkList ID : 105657

Date : 11/17/2017 4:32:25 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/17/2017	Water	I6535-04 - 1	pH	Cool 4 deg C	PSEG01	K61	RT-4252-RT-3025-RT-2342	11/17/2017	SM 4500-PH B
11/17/2017	Water	I6537-01 - 1	pH	Cool 4 deg C	PSEG01	K61	SV22482L	11/17/2017	SM 4500-PH B

Date/Time 11-17-17 11:30 AM
Received by: JP
Relinquished by: CP

Date/Time 11-17-17 6:00 PM
Received by: CP
Relinquished by: JP