



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91679

Temp.: 23.4 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/17/2017	17:00
2	CAL2	1	Water			7.01	11/17/2017	17:01
3	CAL3	1	Water			10.04	11/17/2017	17:03
4	ICV	1	Water			6.99	11/17/2017	17:05
5	CCV1	1	Water			2.01	11/17/2017	17:06
6	I6300-06	1	Solid	20.02	20	9.10	11/17/2017	17:09
7	I6300-06DUP	1	Solid	20.03	20	9.11	11/17/2017	17:10
8	I6306-02	1	Solid	20.02	20	7.23	11/17/2017	17:12
9	I6311-06	1	Solid	20.05	20	8.56	11/17/2017	17:14
10	I6535-03	1	Solid	20.05	20	8.26	11/17/2017	17:15
11	I6538-02	1	Solid	20.01	20	7.72	11/17/2017	17:16
12	CCV2	1	Water			12.01	11/17/2017	17:18

WORKLIST(Hardcopy Internal Chain)

2B 9/16/79

WorkList Name : CORROSIONITY I6538

WorkList ID : 105655

Date : 11/17/2017 4:30:58 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/17/2017	Solid	I6300-06	Corrosivity	Cool 4 deg C	PSEG05	K61	OR-01-111717	11/17/2017	9045C
11/17/2017	Solid	I6306-02	Corrosivity	Cool 4 deg C	PSEG05	K61	NB-08-111717	11/17/2017	9045C
11/17/2017	Solid	I6311-06	Corrosivity	Cool 4 deg C	PSEG05		TR-04-111717	11/17/2017	9045C
11/17/2017	Solid	I6535-03	Corrosivity	Cool 4 deg C	PSEG01	K61	RT-4252-RT-3025-RT-2342	11/17/2017	9045C
11/17/2017	Solid	I6538-02	Corrosivity	Cool 4 deg C	PSEG02	K61	OILY-STONES	11/17/2017	9045C

Date/Time 11-17-17
 Received by: JB
 Relinquished by: _____

Date/Time 11-17-17
 Received by: _____
 Relinquished by: JD