

245.1


LB91681

INSTRUMENT ID: CV1

Linear

 μ Abs.:

33491



A= 0.0000e+000

B= 2.9666e-004

C= 3.0460e-002

Rho= 0.9999609

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.026	-0.026	-189	0.000	-189				
0.2	0.200	0.171	-0.029	475	0.0 %	475				
2.5	2.500	2.542	0.042	8467	0.0 %	8467				
5.0	5.000	5.044	0.044	16901	0.0 %	16901				
7.5	7.500	7.502	0.002	25184	0.0 %	25184				
10.0	10.000	9.966	-0.034	33491	0.0 %	33491				

LB91681 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 17 Nov 2017 17:32:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	-189	-	0.0000	245.1	PPB	17 Nov 2017 17:33:11	S	Std
0.2 - 1	SO2	475	-	0.2000	245.1	PPB	17 Nov 2017 17:35:19	S	Std
2.5 - 1	52.5	8467	-	2.5000	245.1	PPB	17 Nov 2017 17:37:28	S	Std
5.0 - 1	55	16901	-	5.0000	245.1	PPB	17 Nov 2017 17:39:36	S	Std
7.5 - 1	57.5	25184	-	7.5000	245.1	PPB	17 Nov 2017 17:41:44	S	Std
10.0 - 1	S10	33491	-	10.0000	245.1	PPB	17 Nov 2017 17:43:52	S	Std
ICV70 - 1	ICV70	13093	3.9147	-245.1		PPB	17 Nov 2017 17:46:33	U	SMPL
ICB70 - 1	ICB70	-297	-0.0576	-245.1		PPB	17 Nov 2017 17:48:40	U	SMPL
CCV42 - 1	CCV42	16562	4.9438	-245.1		PPB	17 Nov 2017 17:50:47	U	SMPL
CCB42 - 1	CCB42	-313	-0.0624	-245.1		PPB	17 Nov 2017 17:52:53	U	SMPL
CRA - 1	CRA	471	0.1702	-245.1		PPB	17 Nov 2017 17:55:01	U	SMPL
HighStd - 1	HighStd	33927	10.0953	-245.1		PPB	17 Nov 2017 17:57:07	U	SMPL
ChkStd - 1	ChkStd	24116	7.1848	-245.1		PPB	17 Nov 2017 17:59:14	U	SMPL
PB104315BL - 1	PBW	-341	-0.0707	-245.1		PPB	17 Nov 2017 18:01:23	U	SMPL
PB104315BS - 1	LCSW	12536	3.7494	-245.1		PPB	17 Nov 2017 18:03:31	U	SMPL
I6490-01 - 1	111417-JETT-E-D-B	-256	-0.0455	-245.1		PPB	17 Nov 2017 18:05:39	U	SMPL
I6490-02 - 1	111417-JETT-E-D-S	-187	-0.0250	-245.1		PPB	17 Nov 2017 18:07:47	U	SMPL
I6490-03 - 1	111417-JETT-E-U-B	-171	-0.0203	-245.1		PPB	17 Nov 2017 18:09:55	U	SMPL
I6490-04 - 1	111417-JETT-E-U-S	-170	-0.0200	-245.1		PPB	17 Nov 2017 18:12:04	U	SMPL
I6490-05 - 1	111417-TIDO-E-D-B	-185	-0.0244	-245.1		PPB	17 Nov 2017 18:14:10	U	SMPL
CCV43 - 1	CCV43	16694	4.9829	-245.1		PPB	17 Nov 2017 18:16:17	U	SMPL
CCB43 - 1	CCB43	-316	-0.0633	-245.1		PPB	17 Nov 2017 18:18:24	U	SMPL
I6490-06 - 1	111417-TIDO-E-D-S	-169	-0.0197	-245.1		PPB	17 Nov 2017 18:20:31	U	SMPL
I6490-07 - 1	111417-TIDO-E-U-B	-190	-0.0259	-245.1		PPB	17 Nov 2017 18:22:38	U	SMPL
I6490-08 - 1	111417-TIDO-E-U-S	-190	-0.0259	-245.1		PPB	17 Nov 2017 18:24:45	U	SMPL
I6491-01 - 1	111317-TIDO-E-D-B	-184	-0.0241	-245.1		PPB	17 Nov 2017 18:26:52	U	SMPL
I6491-02 - 1	111317-TIDO-E-D-S	-186	-0.0247	-245.1		PPB	17 Nov 2017 18:29:00	U	SMPL
I6491-03 - 1	111317-TIDO-E-U-B	-189	-0.0256	-245.1		PPB	17 Nov 2017 18:31:07	U	SMPL
I6491-04 - 1	111317-TIDO-E-U-S	-190	-0.0259	-245.1		PPB	17 Nov 2017 18:33:15	U	SMPL
I6491-05 - 1	111317-SIDI-E-D-B	-176	-0.0218	-245.1		PPB	17 Nov 2017 18:35:24	U	SMPL
I6491-06 - 1	111317-SIDI-E-D-S	-185	-0.0244	-245.1		PPB	17 Nov 2017 18:37:32	U	SMPL
I6491-07 - 1	111317-SIDI-E-U-B	-184	-0.0241	-245.1		PPB	17 Nov 2017 18:39:39	U	SMPL
CCV44 - 1	CCV44	16663	4.9737	-245.1		PPB	17 Nov 2017 18:41:47	U	SMPL
CCB44 - 1	CCB44	-313	-0.0624	-245.1		PPB	17 Nov 2017 18:43:54	U	SMPL
I6491-08 - 1	111317-SIDI-E-U-S	-194	-0.0271	-245.1		PPB	17 Nov 2017 18:46:02	U	SMPL
I6491-08DUP - 1	111317-SIDI-E-U-SDUP	-192	-0.0265	-245.1		PPB	17 Nov 2017 18:48:09	U	SMPL
I6491-08MS - 1	111317-SIDI-E-U-SMS	15110	4.5130	-245.1		PPB	17 Nov 2017 18:50:16	U	SMPL
I6491-08MSD - 1	111317-SIDI-E-U-SMSD	14576	4.3546	-245.1		PPB	17 Nov 2017 18:52:23	U	SMPL
I6491-08LX5 - 1		-291	-0.0559	-245.1		PPB	17 Nov 2017 18:54:31	U	SMPL
I6491-08A - 1		14391	4.2997	-245.1		PPB	17 Nov 2017 18:56:39	U	SMPL
CCV45 - 1	CCV45	16499	4.9251	-245.1		PPB	17 Nov 2017 18:58:47	U	SMPL
CCB45 - 1	CCB45	-291	-0.0559	-245.1		PPB	17 Nov 2017 19:00:55	U	SMPL