

7470A


LB91692

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34125



A= 0.0000e+000

B= 2.9186e-004

C= 3.7571e-002

Rho= 0.9999756

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.003	-0.003	-139	0.000	-139				
0.2	0.200	0.173	-0.027	464	0.0 %	464				
2.5	2.500	2.522	0.022	8514	0.0 %	8514				
5.0	5.000	5.042	0.042	17146	0.0 %	17146				
7.5	7.500	7.469	-0.031	25461	0.0 %	25461				
10.0	10.000	9.997	-0.003	34125	0.0 %	34125				

LB91692 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Nov 2017 10:28:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-139	-	0.0000	7470A	PPB	20 Nov 2017 10:35:17	S	Std
0.2 - 1	50.2	464	-	0.2000	7470A	PPB	20 Nov 2017 10:37:26	S	Std
2.5 - 1	52.5	8514	-	2.5000	7470A	PPB	20 Nov 2017 10:39:34	S	Std
5.0 - 1	55	17146	-	5.0000	7470A	PPB	20 Nov 2017 10:41:42	S	Std
7.5 - 1	57.5	25461	-	7.5000	7470A	PPB	20 Nov 2017 10:43:50	S	Std
10.0 - 1	510	34125	-	10.0000	7470A	PPB	20 Nov 2017 10:45:57	S	Std
ICV71 - 1	ICV71	12765	3.7631	-	7470A	PPB	20 Nov 2017 10:48:32	U	SMPL
ICB71 - 1	ICB71	-320	-0.0558	-	7470A	PPB	20 Nov 2017 10:50:39	U	SMPL
CCV46 - 1	CCV46	17242	5.0698	-	7470A	PPB	20 Nov 2017 10:52:46	U	SMPL
CCB46 - 1	CCB46	-328	-0.0582	-	7470A	PPB	20 Nov 2017 10:54:53	U	SMPL
CRA - 1	CRA	457	0.1710	-	7470A	PPB	20 Nov 2017 10:56:59	U	SMPL
HighStd - 1	HighStd	34723	10.1718	-	7470A	PPB	20 Nov 2017 10:59:06	U	SMPL
ChkStd - 1	ChkStd	24929	7.3133	-	7470A	PPB	20 Nov 2017 11:01:13	U	SMPL
PB104376BL - 1	PBW	-389	-0.0760	-	7470A	PPB	20 Nov 2017 11:03:22	U	SMPL
PB104376BS - 1	LCSW	13711	4.0392	-	7470A	PPB	20 Nov 2017 11:05:31	U	SMPL
PB104376TB - 1	PB104333TB	-303	-0.0509	-	7470A	PPB	20 Nov 2017 11:07:39	U	SMPL
I6299-06 - 1	OK-01-111617	-207	-0.0228	-	7470A	PPB	20 Nov 2017 11:09:47	U	SMPL
I6304-13 - 1	CL-01-111617-A	-189	-0.0176	-	7470A	PPB	20 Nov 2017 11:11:55	U	SMPL
I6304-15 - 1	CL-01-111617-B	-189	-0.0176	-	7470A	PPB	20 Nov 2017 11:14:04	U	SMPL
I6304-17 - 1	CL-01-111617-C	-194	-0.0190	-	7470A	PPB	20 Nov 2017 11:16:11	U	SMPL
CCV47 - 1	CCV47	17122	5.0348	-	7470A	PPB	20 Nov 2017 11:18:18	U	SMPL
CCB47 - 1	CCB47	-321	-0.0561	-	7470A	PPB	20 Nov 2017 11:20:24	U	SMPL
I6304-19 - 1	CL-01-111617-D	-190	-0.0179	-	7470A	PPB	20 Nov 2017 11:22:32	U	SMPL
I6304-21 - 1	CL-01-111617-E	-190	-0.0179	-	7470A	PPB	20 Nov 2017 11:24:38	U	SMPL
I6309-02 - 1	BU-02-111517	-169	-0.0118	-	7470A	PPB	20 Nov 2017 11:26:45	U	SMPL
I6512-03 - 1	WC-1	-179	-0.0147	-	7470A	PPB	20 Nov 2017 11:28:52	U	SMPL
I6517-03 - 1	HP-WC(COMP)	-181	-0.0153	-	7470A	PPB	20 Nov 2017 11:31:00	U	SMPL
I6521-04 - 1	MH-(0-12)	-176	-0.0138	-	7470A	PPB	20 Nov 2017 11:33:08	U	SMPL
I6521-08 - 1	TP-1-(0-5)	-181	-0.0153	-	7470A	PPB	20 Nov 2017 11:35:16	U	SMPL
I6521-08DUP - 1	TP-1-(0-5)DUP	-182	-0.0155	-	7470A	PPB	20 Nov 2017 11:37:24	U	SMPL
I6521-08MS - 1	TP-1-(0-5)MS	13024	3.8387	-	7470A	PPB	20 Nov 2017 11:39:33	U	SMPL
I6521-08MSD - 1	TP-1-(0-5)MSD	12839	3.7847	-	7470A	PPB	20 Nov 2017 11:41:40	U	SMPL
CCV48 - 1	CCV48	17198	5.0569	-	7470A	PPB	20 Nov 2017 11:43:46	U	SMPL
CCB48 - 1	CCB48	-315	-0.0544	-	7470A	PPB	20 Nov 2017 11:45:54	U	SMPL
I6521-12 - 1	TP-2-(0-5)	-194	-0.0190	-	7470A	PPB	20 Nov 2017 11:48:01	U	SMPL
PB104394BL - 1	PBW	-180	-0.0150	-	7470A	PPB	20 Nov 2017 11:50:08	U	SMPL
PB104394BS - 1	LCSW	13473	3.9698	-	7470A	PPB	20 Nov 2017 11:52:16	U	SMPL
I6505-01 - 1	LW-01	-278	-0.0436	-	7470A	PPB	20 Nov 2017 11:54:23	U	SMPL
I6505-01DUP - 1	LW-01DUP	-198	-0.0202	-	7470A	PPB	20 Nov 2017 11:56:30	U	SMPL
I6505-02 - 1	I6505-01MS	12571	3.7065	-	7470A	PPB	20 Nov 2017 11:58:38	U	SMPL
I6505-03 - 1	I6505-01MSD	12465	3.6756	-	7470A	PPB	20 Nov 2017 12:00:46	U	SMPL
I6505-04 - 1	LW-03	-309	-0.0526	-	7470A	PPB	20 Nov 2017 12:02:54	U	SMPL
I6505-05 - 1	LW-D	-205	-0.0223	-	7470A	PPB	20 Nov 2017 12:05:03	U	SMPL
I6505-06 - 1	LW-06	-214	-0.0249	-	7470A	PPB	20 Nov 2017 12:07:10	U	SMPL
CCV49 - 1	CCV49	17239	5.0689	-	7470A	PPB	20 Nov 2017 12:09:17	U	SMPL
CCB49 - 1	CCB49	-294	-0.0482	-	7470A	PPB	20 Nov 2017 12:11:24	U	SMPL
I6505-07 - 1	LW-04	-210	-0.0237	-	7470A	PPB	20 Nov 2017 12:13:32	U	SMPL
I6505-08 - 1	LW-05	-167	-0.0112	-	7470A	PPB	20 Nov 2017 12:15:40	U	SMPL
I6521-08LX5 - 1		-164	-0.0103	-	7470A	PPB	20 Nov 2017 12:17:47	U	SMPL
I6521-08A - 1		13449	3.9628	-	7470A	PPB	20 Nov 2017 12:19:54	U	SMPL
I6505-01LX5 - 1		-291	-0.0474	-	7470A	PPB	20 Nov 2017 12:22:02	U	SMPL
I6505-01A - 1		13333	3.9289	-	7470A	PPB	20 Nov 2017 12:24:09	U	SMPL
CCV50 - 1	CCV50	16904	4.9711	-	7470A	PPB	20 Nov 2017 12:26:18	U	SMPL
CCB50 - 1	CCB50	-288	-0.0465	-	7470A	PPB	20 Nov 2017 12:28:26	U	SMPL