


7471A

LB 91696
INSTRUMENT ID: CV 1

Linear

 μ Abs.:

44219



A= 0.0000e+000

B= 2.2513e-004

C= 3.9440e-002

Rho= 0.9999518

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.030	0.030	-42	0.000	-42				
0.2	0.200	0.201	0.001	718	0.0 %	718				
2.5	2.500	2.483	-0.017	10855	0.0 %	10855				
5.0	5.000	4.938	-0.062	21759	0.0 %	21759				
7.5	7.500	7.553	0.053	33375	0.0 %	33375	0			
10.0	10.000	9.994	-0.006	44219	0.0 %	44219				

LB91696 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 20 Nov 2017 13:38:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-42	-	0.00007471A	PPB		20 Nov 2017 14:47:15	S	Std
0.2 - 1	50.2	718	-	0.20007471A	PPB		20 Nov 2017 14:49:32	S	Std
2.5 - 1	52.5	10855	-	2.50007471A	PPB		20 Nov 2017 14:51:50	S	Std
5.0 - 1	55	21759	-	5.00007471A	PPB		20 Nov 2017 14:54:07	S	Std
7.5 - 1	57.5	33375	-	7.50007471A	PPB		20 Nov 2017 14:56:25	S	Std
10.0 - 1	510	44219	-	10.00007471A	PPB		20 Nov 2017 14:58:42	S	Std
ICV72 - 1	ICV72	17306	3.9355	-7471A	PPB		20 Nov 2017 15:01:52	U	SMPL
ICB72 - 1	ICB72	-223	-0.0108	-7471A	PPB		20 Nov 2017 15:04:07	U	SMPL
CCV51 - 1	CCV51	21560	4.8933	-7471A	PPB		20 Nov 2017 15:06:23	U	SMPL
CCB51 - 1	CCB51	-242	-0.0150	-7471A	PPB		20 Nov 2017 15:08:39	U	SMPL
CRA - 1	CRA	698	0.1966	-7471A	PPB		20 Nov 2017 15:10:55	U	SMPL
HighStd - 1	HighStd	44884	10.1442	-7471A	PPB		20 Nov 2017 15:13:10	U	SMPL
ChkStd - 1	ChkStd	31088	7.0383	-7471A	PPB		20 Nov 2017 15:15:27	U	SMPL
PB104378BL - 1	PBS	-413	-0.0535	-7471A	PPB		20 Nov 2017 15:17:45	U	SMPL
PB104378BS - 1	LCSS	15889	3.6165	-7471A	PPB		20 Nov 2017 15:20:03	U	SMPL
I6299-05 - 1	OK-01-111617	12976	2.9607	-7471A	PPB		20 Nov 2017 15:22:20	U	SMPL
I6300-05 - 1	OR-01-111717	214	0.0876	-7471A	PPB		20 Nov 2017 15:24:38	U	SMPL
I6304-12 - 1	CL-01-111617-A	1186	0.3064	-7471A	PPB		20 Nov 2017 15:26:55	U	SMPL
I6304-14 - 1	CL-01-111617-B	1567	0.3922	-7471A	PPB		20 Nov 2017 15:29:13	U	SMPL
I6304-16 - 1	CL-01-111617-C	1537	0.3855	-7471A	PPB		20 Nov 2017 15:31:29	U	SMPL
CCV52 - 1	CCV52	21762	4.9387	-7471A	PPB		20 Nov 2017 15:33:46	U	SMPL
CCB52 - 1	CCB52	-243	-0.0153	-7471A	PPB		20 Nov 2017 15:36:01	U	SMPL
I6304-18 - 1	CL-01-111617-D	1543	0.3868	-7471A	PPB		20 Nov 2017 15:38:17	U	SMPL
I6304-20 - 1	CL-01-111617-E	2421	0.5845	-7471A	PPB		20 Nov 2017 15:40:33	U	SMPL
I6306-01 - 1	NB-08-111717	649	0.1855	-7471A	PPB		20 Nov 2017 15:42:49	U	SMPL
I6311-05 - 1	TR-04-111717	983	0.2607	-7471A	PPB		20 Nov 2017 15:45:06	U	SMPL
I6512-01 - 1	WC-1	1525	0.3828	-7471A	PPB		20 Nov 2017 15:47:22	U	SMPL
I6513-01 - 1	SP-1	4344	1.0174	-7471A	PPB		20 Nov 2017 15:49:40	U	SMPL
I6517-01 - 1	HP-WC(COMP)	29555	6.6932	-7471A	PPB		20 Nov 2017 15:51:57	U	SMPL
I6521-01 - 1	MH-(0-12)	168	0.0773	-7471A	PPB		20 Nov 2017 16:04:52	U	SMPL
I6521-01DUP - 1	MH-(0-12)DUP	144	0.0719	-7471A	PPB		20 Nov 2017 16:07:09	U	SMPL
I6521-01MS - 1	MH-(0-12)MS	15948	3.6298	-7471A	PPB		20 Nov 2017 16:09:25	U	SMPL
CCV53 - 1	CCV53	21865	4.9619	-7471A	PPB		20 Nov 2017 16:11:42	U	SMPL
CCB53 - 1	CCB53	-268	-0.0209	-7471A	PPB		20 Nov 2017 16:13:58	U	SMPL
I6521-01MSD - 1	MH-(0-12)MSD	15874	3.6132	-7471A	PPB		20 Nov 2017 16:16:14	U	SMPL
I6521-05 - 1	TP-1-(0-5)	2697	0.6466	-7471A	PPB		20 Nov 2017 16:18:31	U	SMPL
I6521-09 - 1	TP-2-(0-5)	4323	1.0127	-7471A	PPB		20 Nov 2017 16:20:47	U	SMPL
I6528-01 - 1	S-1-BACKYARD	14386	3.2782	-7471A	PPB		20 Nov 2017 16:23:04	U	SMPL
I6528-02 - 1	S-2-BACKYARD	758	0.2101	-7471A	PPB		20 Nov 2017 16:25:20	U	SMPL
I6535-01 - 1	RT-4252-RT-3025-RT-2342	1641	0.4089	-7471A	PPB		20 Nov 2017 16:27:37	U	SMPL
I6538-01 - 1	OILY-STONES	-191	-0.0036	-7471A	PPB		20 Nov 2017 16:29:55	U	SMPL
I6521-01LX5 - 1		-47	0.0289	-7471A	PPB		20 Nov 2017 16:32:12	U	SMPL
I6521-01A - 1		16778	3.8167	-7471A	PPB		20 Nov 2017 16:34:30	U	SMPL
CCV54 - 1	CCV54	21578	4.8973	-7471A	PPB		20 Nov 2017 16:36:47	U	SMPL
CCB54 - 1	CCB54	-244	-0.0155	-7471A	PPB		20 Nov 2017 16:39:03	U	SMPL