



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91707

Temp.: 24.4 °C

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/20/2017	17:30
2	CAL2	1	Water			7.00	11/20/2017	17:31
3	CAL3	1	Water			10.01	11/20/2017	17:32
4	ICV	1	Water			6.99	11/20/2017	17:34
5	CCV1	1	Water			2.01	11/20/2017	17:35
6	I6558-02	1	Water			11.67	11/20/2017	17:36
7	I6558-02DUP	1	Water			11.68	11/20/2017	17:38
8	I6558-04	1	Water			7.50	11/20/2017	17:39
9	CCV2	1	Water			12.01	11/20/2017	17:40

2891707

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6578 WorkList ID : 105726 Date : 11/20/2017 5:12:06 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/20/2017	Water	I6558-02-18	Corrosivity	Cool 4 deg C	PSEG01	M52	A1243	11/20/2017	9040C
11/20/2017	Water	I6558-04-18	Corrosivity	Cool 4 deg C	PSEG01	M52	SUMP-WELL-1	11/20/2017	9040C

Date/Time 11-20-17 5:12 PM
Received by: Ji
Relinquished by: CP

Date/Time 11-20-17 6:20 PM
Received by: CP
Relinquished by: Ji