



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB91708

Temp.: 24.4 °C

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/20/2017	17:30
2	CAL2	1	Water			7.00	11/20/2017	17:31
3	CAL3	1	Water			10.01	11/20/2017	17:32
4	ICV	1	Water			6.99	11/20/2017	17:34
5	CCV1	1	Water			2.01	11/20/2017	17:35
6	I6558-01	1	Water			11.67	11/20/2017	17:36
7	I6558-01DUP	1	Water			11.68	11/20/2017	17:38
8	I6558-03	1	Water			7.50	11/20/2017	17:39
9	I6558-05	1	Water			10.03	11/20/2017	17:41
10	CCV2	1	Water			12.02	11/20/2017	17:42

2991708

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I6578

WorkList ID : 105727

Date : 11/20/2017 5:12:33 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/20/2017	Water	I6558-01 -1-	pH	Cool 4 deg C	PSEG01	M52	A1243	11/20/2017	SM 4500-PH B
11/20/2017	Water	I6558-03 -1-	pH	Cool 4 deg C	PSEG01	M52	SUMP-WELL-1	11/20/2017	SM 4500-PH B
11/20/2017	Water	I6558-05 ✓	pH	Cool 4 deg C	PSEG01	M52	A5972OT-REAR-TANK	11/20/2017	SM 4500-PH B

Date/Time 11-20-17 5:12 PM
Received by: JP
Relinquished by: JP

Date/Time 11-20-17 6:20 PM
Received by: JP
Relinquished by: JP