


7470A

LB 91715
INSTRUMENT ID: CV4

Linear

 μ Abs.:

35741



A= 0.0000e+000

B= 2.7782e-004

C= 1.1640e-001

Rho= 0.9996132

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.069	0.069	-169	0.000	-169				
0.2	0.200	0.239	0.039	442	0.0 %	442				
2.5	2.500	2.484	-0.016	8521	0.0 %	8521				
5.0	5.000	4.781	-0.219	16790	0.0 %	16790				
7.5	7.500	7.581	0.081	26867	0.0 %	26867				
10.0	10.000	10.046	0.046	35741	0.0 %	35741				

LB91715 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Nov 2017 10:49:57

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-169	-	0.0000	7470A	PPB	21 Nov 2017 10:54:16	S	Std
0.2 - 1	50.2	442	-	0.2000	7470A	PPB	21 Nov 2017 10:56:24	S	Std
2.5 - 1	32.5	8521	-	2.5000	7470A	PPB	21 Nov 2017 10:58:33	S	Std
5.0 - 1	55	16790	-	5.0000	7470A	PPB	21 Nov 2017 11:00:41	S	Std
7.5 - 1	57.5	26867	-	7.5000	7470A	PPB	21 Nov 2017 11:02:49	S	Std
10.0 - 1	510	35741	-	10.0000	7470A	PPB	21 Nov 2017 11:04:57	S	Std
ICV73 - 1	ICV73	12596	3.6158	-	7470A	PPB	21 Nov 2017 11:07:37	U	SMPL
ICB73 - 1	ICB73	-297	0.0339	-	7470A	PPB	21 Nov 2017 11:09:43	U	SMPL
CCV55 - 1	CCV55	16817	4.7885	-	7470A	PPB	21 Nov 2017 11:11:51	U	SMPL
CCB55 - 1	CCB55	-301	0.0328	-	7470A	PPB	21 Nov 2017 11:13:57	U	SMPL
CRA - 1	CRA	425	0.2345	-	7470A	PPB	21 Nov 2017 11:16:05	U	SMPL
HighStd - 1	HighStd	35652	10.0213	-	7470A	PPB	21 Nov 2017 11:18:12	U	SMPL
ChkStd - 1	ChkStd	25104	7.0908	-	7470A	PPB	21 Nov 2017 11:20:18	U	SMPL
PB104393BL - 1	PBW	-390	0.0081	-	7470A	PPB	21 Nov 2017 11:22:27	U	SMPL
PB104393BS - 1	LCSW	13305	3.8128	-	7470A	PPB	21 Nov 2017 11:24:37	U	SMPL
I6525-01 - 1	MW-21	-301	0.0328	-	7470A	PPB	21 Nov 2017 11:26:44	U	SMPL
I6525-01DUP - 1	MW-21DUP	-204	0.0597	-	7470A	PPB	21 Nov 2017 11:28:52	U	SMPL
I6525-01MS - 1	MW-21MS	13798	3.9498	-	7470A	PPB	21 Nov 2017 11:31:00	U	SMPL
I6525-01MSD - 1	MW-21MSD	14249	4.0751	-	7470A	PPB	21 Nov 2017 11:33:09	U	SMPL
I6525-02 - 1	MW-24	-275	0.0400	-	7470A	PPB	21 Nov 2017 11:35:15	U	SMPL
CCV56 - 1	CCV56	17018	4.8444	-	7470A	PPB	21 Nov 2017 11:37:22	U	SMPL
CCB56 - 1	CCB56	-312	0.0297	-	7470A	PPB	21 Nov 2017 11:39:29	U	SMPL
I6525-03 - 1	T71-MW7	-203	0.0600	-	7470A	PPB	21 Nov 2017 11:41:36	U	SMPL
I6525-04 - 1	MW-12	-190	0.0636	-	7470A	PPB	21 Nov 2017 11:43:43	U	SMPL
I6525-05 - 1	MW-20	-159	0.0722	-	7470A	PPB	21 Nov 2017 11:45:50	U	SMPL
I6525-06 - 1	MW-11	-153	0.0739	-	7470A	PPB	21 Nov 2017 11:47:57	U	SMPL
I6525-07 - 1	FB111517	-190	0.0636	-	7470A	PPB	21 Nov 2017 11:50:04	U	SMPL
I6525-09 - 1	T70-DW1	-1	0.1161	-	7470A	PPB	21 Nov 2017 11:52:12	U	SMPL
I6525-10 - 1	DUP111617	-33	0.1072	-	7470A	PPB	21 Nov 2017 11:54:20	U	SMPL
I6525-11 - 1	T71-MW5	-175	0.0678	-	7470A	PPB	21 Nov 2017 11:56:28	U	SMPL
I6525-12 - 1	MW-18	-188	0.0642	-	7470A	PPB	21 Nov 2017 11:58:37	U	SMPL
I6525-13 - 1	T71-MW4	-85	0.0928	-	7470A	PPB	21 Nov 2017 12:00:44	U	SMPL
CCV57 - 1	CCV57	16914	4.8155	-	7470A	PPB	21 Nov 2017 12:02:51	U	SMPL
CCB57 - 1	CCB57	-310	0.0303	-	7470A	PPB	21 Nov 2017 12:04:58	U	SMPL
I6525-14 - 1	FB111617	-212	0.0575	-	7470A	PPB	21 Nov 2017 12:07:06	U	SMPL
6535-04 - 1	RT-4252-RT-3025-RT-2342	28	0.1242	-	7470A	PPB	21 Nov 2017 12:09:13	U	SMPL
6536-02 - 1	MW-17	-109	0.0861	-	7470A	PPB	21 Nov 2017 12:11:20	U	SMPL
6536-03 - 1	FB111717	-192	0.0631	-	7470A	PPB	21 Nov 2017 12:13:27	U	SMPL
6536-04 - 1	MW-10	-177	0.0672	-	7470A	PPB	21 Nov 2017 12:15:35	U	SMPL
6536-05 - 1	MW-16	-84	0.0931	-	7470A	PPB	21 Nov 2017 12:17:42	U	SMPL
6537-01 - 1	SV22482L	-114	0.0847	-	7470A	PPB	21 Nov 2017 12:19:50	U	SMPL
PB104423BL - 1	PBW	-174	0.0681	-	7470A	PPB	21 Nov 2017 12:21:59	U	SMPL
PB104423BS - 1	LCSW	14029	4.0140	-	7470A	PPB	21 Nov 2017 12:24:07	U	SMPL
PB104423TB - 1	PB104382TB	-276	0.0397	-	7470A	PPB	21 Nov 2017 12:26:14	U	SMPL
CCV58 - 1	CCV58	16809	4.7863	-	7470A	PPB	21 Nov 2017 12:28:21	U	SMPL
CCB58 - 1	CCB58	-293	0.0350	-	7470A	PPB	21 Nov 2017 12:30:29	U	SMPL
6300-06 - 1	OR-01-111717	-191	0.0633	-	7470A	PPB	21 Nov 2017 12:32:37	U	SMPL
6306-02 - 1	NB-08-111717	-180	0.0664	-	7470A	PPB	21 Nov 2017 12:34:44	U	SMPL
6311-06 - 1	TR-04-111717	-184	0.0653	-	7470A	PPB	21 Nov 2017 12:36:51	U	SMPL
6535-03 - 1	RT-4252-RT-3025-RT-2342	-178	0.0670	-	7470A	PPB	21 Nov 2017 12:38:58	U	SMPL
6535-05 - 1	RT-4252-RT-3025-RT-2342	-165	0.0706	-	7470A	PPB	21 Nov 2017 12:41:06	U	SMPL
6537-02 - 1	SV22482L	-196	0.0620	-	7470A	PPB	21 Nov 2017 12:43:13	U	SMPL
6538-02 - 1	OILY-STONES	-165	0.0706	-	7470A	PPB	21 Nov 2017 12:45:21	U	SMPL
6538-02DUP - 1	OILY-STONESDUP	-187	0.0645	-	7470A	PPB	21 Nov 2017 12:47:30	U	SMPL
6538-02MS - 1	OILY-STONESMS	13297	3.8106	-	7470A	PPB	21 Nov 2017 12:49:38	U	SMPL
6538-02MSD - 1	OILY-STONESMSD	12942	3.7120	-	7470A	PPB	21 Nov 2017 12:51:45	U	SMPL
CCV59 - 1	CCV59	17143	4.8791	-	7470A	PPB	21 Nov 2017 13:03:28	U	SMPL
CCB59 - 1	CCB59	-292	0.0353	-	7470A	PPB	21 Nov 2017 13:05:35	U	SMPL
6538-03 - 1	OILY-DEBRIS	-156	0.0731	-	7470A	PPB	21 Nov 2017 13:07:43	U	SMPL
6525-01LX5 - 1		-145	0.0761	-	7470A	PPB	21 Nov 2017 13:09:50	U	SMPL
6525-01A - 1		14188	4.0581	-	7470A	PPB	21 Nov 2017 13:11:57	U	SMPL
6538-02LX5 - 1		-267	0.0422	-	7470A	PPB	21 Nov 2017 13:14:04	U	SMPL
6538-02A - 1		13178	3.7775	-	7470A	PPB	21 Nov 2017 13:16:12	U	SMPL
CCV60 - 1	CCV60	17085	4.8630	-	7470A	PPB	21 Nov 2017 13:18:20	U	SMPL
CCB60 - 1	CCB60	-298	0.0336	-	7470A	PPB	21 Nov 2017 13:20:29	U	SMPL