

245.1


LB91717

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34432



A= 0.0000e+000

B= 2.8959e-004

C= 3.4642e-002

Rho= 0.9999613

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.014	-0.014	-167	0.000	-167				
0.2	0.200	0.183	-0.017	513	0.0 %	513				
2.5	2.500	2.523	0.023	8593	0.0 %	8593				
5.0	5.000	5.052	0.052	17327	0.0 %	17327				
7.5	7.500	7.449	-0.051	25605	0.0 %	25605				
10.0	10.000	10.006	0.006	34432	0.0 %	34432				

LB91717 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 21 Nov 2017 13:29:40

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-167	-	0.0000	245.1	PPB	21 Nov 2017 13:31:55	S	Std
0.2 - 1	50.2	513	-	0.2000	245.1	PPB	21 Nov 2017 13:34:03	S	Std
2.5 - 1	52.5	8593	-	2.5000	245.1	PPB	21 Nov 2017 13:36:12	S	Std
5.0 - 1	55	17327	-	5.0000	245.1	PPB	21 Nov 2017 13:38:20	S	Std
7.5 - 1	57.5	25605	-	7.5000	245.1	PPB	21 Nov 2017 13:40:28	S	Std
10.0 - 1	57.8	34432	-	10.0000	245.1	PPB	21 Nov 2017 13:42:35	S	Std
ICV74 - 1	ICV74	13176	3.8502	-245.1		PPB	21 Nov 2017 13:47:29	U	SMPL
ICB74 - 1	ICB74	-254	-0.0389	-245.1		PPB	21 Nov 2017 13:49:35	U	SMPL
CCV61 - 1	CCV61	17247	5.0291	-245.1		PPB	21 Nov 2017 13:51:42	U	SMPL
CCB61 - 1	CCB61	-251	-0.0380	-245.1		PPB	21 Nov 2017 13:53:48	U	SMPL
CRA - 1	CRA	481	0.1739	-245.1		PPB	21 Nov 2017 13:55:56	U	SMPL
HighStd - 1	HighStd	33359	9.6949	-245.1		PPB	21 Nov 2017 13:58:02	U	SMPL
ChkStd - 1	ChkStd	22899	6.6659	-245.1		PPB	21 Nov 2017 14:00:09	U	SMPL
PB104395BL - 1	PBW	-343	-0.0647	-245.1		PPB	21 Nov 2017 14:02:18	U	SMPL
PB104395BS - 1	LCSW	13254	3.8728	-245.1		PPB	21 Nov 2017 14:04:28	U	SMPL
I6539-01 - 1	111517-TIDO-F-D-B	-311	-0.0554	-245.1		PPB	21 Nov 2017 14:06:35	U	SMPL
I6539-02 - 1	111517-TIDO-F-D-S	-225	-0.0305	-245.1		PPB	21 Nov 2017 14:08:44	U	SMPL
I6539-03 - 1	111517-TIDO-F-U-B	-203	-0.0241	-245.1		PPB	21 Nov 2017 14:10:52	U	SMPL
I6539-04 - 1	111517-TIDO-F-U-S	-171	-0.0149	-245.1		PPB	21 Nov 2017 14:13:00	U	SMPL
I6539-05 - 1	111517-JETT-E-D-B	-179	-0.0172	-245.1		PPB	21 Nov 2017 14:15:07	U	SMPL
CCV62 - 1	CCV62	16902	4.9292	-245.1		PPB	21 Nov 2017 14:17:14	U	SMPL
CCB62 - 1	CCB62	-269	-0.0433	-245.1		PPB	21 Nov 2017 14:19:20	U	SMPL
I6539-06 - 1	111517-JETT-E-D-S	-191	-0.0207	-245.1		PPB	21 Nov 2017 14:21:28	U	SMPL
I6539-07 - 1	111517-JETT-E-U-B	-198	-0.0227	-245.1		PPB	21 Nov 2017 14:23:35	U	SMPL
I6539-08 - 1	111517-JETT-E-U-S	-211	-0.0265	-245.1		PPB	21 Nov 2017 14:25:41	U	SMPL
I6540-01 - 1	111617-TIDO-F-D-B	-172	-0.0152	-245.1		PPB	21 Nov 2017 14:27:48	U	SMPL
I6540-02 - 1	111617-TIDO-F-D-S	-160	-0.0117	-245.1		PPB	21 Nov 2017 14:29:56	U	SMPL
I6540-03 - 1	111617-TIDO-F-U-B	-202	-0.0239	-245.1		PPB	21 Nov 2017 14:32:03	U	SMPL
I6540-04 - 1	111617-TIDO-F-U-S	-200	-0.0233	-245.1		PPB	21 Nov 2017 14:34:11	U	SMPL
I6540-04DUP - 1	111617-TIDO-F-U-SDUP	-195	-0.0218	-245.1		PPB	21 Nov 2017 14:36:20	U	SMPL
I6540-04MS - 1	111617-TIDO-F-U-SMS	15164	4.4259	-245.1		PPB	21 Nov 2017 14:38:28	U	SMPL
I6540-04MSD - 1	111617-TIDO-F-U-SMSD	14884	4.3448	-245.1		PPB	21 Nov 2017 14:40:35	U	SMPL
CCV63 - 1	CCV63	17246	5.0288	-245.1		PPB	21 Nov 2017 14:42:42	U	SMPL
CCB63 - 1	CCB63	-291	-0.0496	-245.1		PPB	21 Nov 2017 14:44:48	U	SMPL
PB104396BL - 1	PBW	-217	-0.0282	-245.1		PPB	21 Nov 2017 14:46:56	U	SMPL
PB104396BS - 1	LCSW	13018	3.8045	-245.1		PPB	21 Nov 2017 14:49:03	U	SMPL
6541-01 - 1	111717-JETT-F-D-B	-281	-0.0467	-245.1		PPB	21 Nov 2017 14:51:10	U	SMPL
6541-02 - 1	111717-JETT-F-D-S	-201	-0.0236	-245.1		PPB	21 Nov 2017 14:53:18	U	SMPL
6541-03 - 1	111717-JETT-F-U-B	-202	-0.0239	-245.1		PPB	21 Nov 2017 14:55:25	U	SMPL
6541-04 - 1	111717-JETT-F-U-S	-187	-0.0195	-245.1		PPB	21 Nov 2017 14:57:32	U	SMPL
6541-05 - 1	111717-SIDI-F-D-B	-183	-0.0184	-245.1		PPB	21 Nov 2017 14:59:40	U	SMPL
6541-06 - 1	111717-SIDI-F-D-M	-185	-0.0189	-245.1		PPB	21 Nov 2017 15:01:48	U	SMPL
6541-07 - 1	111717-SIDI-F-D-S	-171	-0.0149	-245.1		PPB	21 Nov 2017 15:03:57	U	SMPL
6541-08 - 1	111717-SIDI-F-U-B	-194	-0.0215	-245.1		PPB	21 Nov 2017 15:06:04	U	SMPL
CCV64 - 1	CCV64	17277	5.0378	-245.1		PPB	21 Nov 2017 15:08:11	U	SMPL
CCB64 - 1	CCB64	-278	-0.0459	-245.1		PPB	21 Nov 2017 15:10:19	U	SMPL
6541-09 - 1	111717-SIDI-F-U-M	-198	-0.0227	-245.1		PPB	21 Nov 2017 15:12:27	U	SMPL
6541-10 - 1	111717-SIDI-F-U-S	-206	-0.0250	-245.1		PPB	21 Nov 2017 15:14:34	U	SMPL
6542-01 - 1	111717-TIDO-F-D-B	-190	-0.0204	-245.1		PPB	21 Nov 2017 15:16:41	U	SMPL
6542-02 - 1	111717-TIDO-F-D-S	-191	-0.0207	-245.1		PPB	21 Nov 2017 15:18:48	U	SMPL
6542-03 - 1	111717-TIDO-F-U-B	-162	-0.0123	-245.1		PPB	21 Nov 2017 15:20:56	U	SMPL
6542-04 - 1	111717-TIDO-F-U-S	-152	-0.0094	-245.1		PPB	21 Nov 2017 15:23:03	U	SMPL
6542-04DUP - 1	111717-TIDO-F-U-SDUP	-186	-0.0192	-245.1		PPB	21 Nov 2017 15:25:11	U	SMPL
6542-04MS - 1	111717-TIDO-F-U-SMS	16113	4.7007	-245.1		PPB	21 Nov 2017 15:27:20	U	SMPL
6542-04MSD - 1	111717-TIDO-F-U-SMSD	15152	4.4225	-245.1		PPB	21 Nov 2017 15:29:29	U	SMPL
CCV65 - 1	CCV65	16900	4.9286	-245.1		PPB	21 Nov 2017 15:31:36	U	SMPL
CCB65 - 1	CCB65	-288	-0.0488	-245.1		PPB	21 Nov 2017 15:33:43	U	SMPL
6540-04LX5 - 1		-158	-0.0111	-245.1		PPB	21 Nov 2017 15:35:51	U	SMPL
6540-04A - 1		15188	4.4329	-245.1		PPB	21 Nov 2017 15:37:58	U	SMPL
6542-04LX5 - 1		-237	-0.0340	-245.1		PPB	21 Nov 2017 15:40:06	U	SMPL
6542-04A - 1		15116	4.4120	-245.1		PPB	21 Nov 2017 15:42:13	U	SMPL
CCV66 - 1	CCV66	16593	4.8397	-245.1		PPB	21 Nov 2017 15:44:20	U	SMPL
CCB66 - 1	CCB66	-301	-0.0525	-245.1		PPB	21 Nov 2017 15:46:28	U	SMPL