


ISM02.4

LB 91723
INSTRUMENT ID: CV1

Linear

 μ Abs.:

4059E



A= 0.0000e+000

B= 2.4816e-004

C= 1.2939e-002

Rho= 0.9998299

Accept=Accepted

slope
+ 1/1000000

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.012	-0.012	-101	0.000	-101					-
0.2	0.200	0.214	0.014	810	0.0 %	810					7
2.5	2.500	2.557	0.057	10251	0.0 %	10251					20
5.0	5.000	4.981	-0.019	20019	0.0 %	20019					-2
7.5	7.500	7.373	-0.127	29658	0.0 %	29658					1
10.0	10.000	10.088	0.088	40598	0.0 %	40598					

NB
11/21/18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	-101	-		0 ISM02.4	PPB	11/21/2017 15:57	Std
	0.2 S0.2	810	-		0.2 ISM02.4	PPB	11/21/2017 15:59	Std
	2.5 S2.5	10251	-		2.5 ISM02.4	PPB	11/21/2017 16:02	Std
	5 S5	20019	-		5 ISM02.4	PPB	11/21/2017 16:04	Std
	7.5 S7.5	29658	-		7.5 ISM02.4	PPB	11/21/2017 16:06	Std
	10 S10	40598	-		10 ISM02.4	PPB	11/21/2017 16:08	Std
ICV	ICV	16109	4.0105 -		ISM02.4	PPB	11/21/2017 16:12	SMPL
ICB	ICB	-242	-0.0471 -		ISM02.4	PPB	11/21/2017 16:14	SMPL
CCV067	CCV067	20290	5.0481 -		ISM02.4	PPB	11/21/2017 16:16	SMPL
CCB067	CCB067	-267	-0.0533 -		ISM02.4	PPB	11/21/2017 16:18	SMPL
PB104434BL	PBW003	-208	-0.0387 -		ISM02.4	PPB	11/21/2017 16:21	SMPL
I6516-16	MC0B93	22	0.0184 -		ISM02.4	PPB	11/21/2017 16:23	SMPL
I6516-17	MC0B93D	10	0.0154 -		ISM02.4	PPB	11/21/2017 16:25	SMPL
I6516-18	MC0B93S	4401	1.1051 -		ISM02.4	PPB	11/21/2017 16:27	SMPL
I6516-19	MC0B94	-12	0.01 -		ISM02.4	PPB	11/21/2017 16:30	SMPL
I6516-20	MC0B95	-209	-0.0389 -		ISM02.4	PPB	11/21/2017 16:32	SMPL
CCV068	CCV068	20360	5.0655 -		ISM02.4	PPB	11/21/2017 16:34	SMPL
CCB068	CCB068	-244	-0.0476 -		ISM02.4	PPB	11/21/2017 16:37	SMPL