


ISM02.4

LB 91732
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44613



A= 0.0000e+000

B= 2.2266e-004

C= 3.1686e-002

Rho= 0.9999707

Accept=Accepted

slope
y=mx+cept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/p
0.00	0.000	0.001	0.001	-138	0.000	-138					0
0.2	0.200	0.200	0.000	758	0.0 %	758					-1
2.5	2.500	2.465	-0.035	10928	0.0 %	10928					1
5.0	5.000	5.033	0.033	22462	0.0 %	22462					0
7.5	7.500	7.535	0.035	33700	0.0 %	33700					0
10.0	10.000	9.965	-0.035	44613	0.0 %	44613					0

NB
11/21/18

LB91732
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	-138	-		0 ISM02.4	PPB	11/21/2017 16:50	Std
	0.2 S0.2	758	-		0.2 ISM02.4	PPB	11/21/2017 16:52	Std
	2.5 S2.5	10928	-		2.5 ISM02.4	PPB	11/21/2017 16:55	Std
	5 S5	22462	-		5 ISM02.4	PPB	11/21/2017 16:57	Std
	7.5 S7.5	33700	-		7.5 ISM02.4	PPB	11/21/2017 16:59	Std
	10 S10	44613	-		10 ISM02.4	PPB	11/21/2017 17:02	Std
ICV	ICV	17259	3.8746 -		ISM02.4	PPB	11/21/2017 17:04	SMPL
ICB	ICB	-302	-0.0356 -		ISM02.4	PPB	11/21/2017 17:07	SMPL
CCV069	CCV069	22386	5.0162 -		ISM02.4	PPB	11/21/2017 17:09	SMPL
CCB069	CCB069	-309	-0.0371 -		ISM02.4	PPB	11/21/2017 17:11	SMPL
PB104440BL	PBS 004	-190	-0.0106 -		ISM02.4	PPB	11/21/2017 17:13	SMPL
I6515-01	MC0AA0	373	0.1147 -		ISM02.4	PPB	11/21/2017 17:21	SMPL
I6515-02	MC0AA0D	376	0.1154 -		ISM02.4	PPB	11/21/2017 17:23	SMPL
I6515-03	MC0AA0S	11674	2.631 -		ISM02.4	PPB	11/21/2017 17:25	SMPL
I6515-04	MC0AA1	-32	0.0246 -		ISM02.4	PPB	11/21/2017 17:27	SMPL
I6515-05	MC0AA2	76	0.0486 -		ISM02.4	PPB	11/21/2017 17:30	SMPL
I6515-06	MC0AA3	563	0.157 -		ISM02.4	PPB	11/21/2017 17:32	SMPL
I6515-07	MC0AA4	3710	0.8578 -		ISM02.4	PPB	11/21/2017 17:34	SMPL
I6515-08	MC0AA6	30	0.0384 -		ISM02.4	PPB	11/21/2017 17:36	SMPL
I6515-09	MC0AA7	426	0.1265 -		ISM02.4	PPB	11/21/2017 17:39	SMPL
I6515-10	MC0AA8	1705	0.4113 -		ISM02.4	PPB	11/21/2017 17:41	SMPL
I6515-11	MC0AA9	161	0.0675 -		ISM02.4	PPB	11/21/2017 17:43	SMPL
I6515-12	MC0AB0	391	0.1187 -		ISM02.4	PPB	11/21/2017 17:46	SMPL
I6515-13	MC0AB1	722	0.1924 -		ISM02.4	PPB	11/21/2017 17:48	SMPL
I6515-14	MC0AB2	543	0.1526 -		ISM02.4	PPB	11/21/2017 17:50	SMPL
I6515-15	MC0AB3	670	0.1809 -		ISM02.4	PPB	11/21/2017 17:52	SMPL
I6515-16	MC0AB4	530	0.1497 -		ISM02.4	PPB	11/21/2017 17:55	SMPL
I6515-17	MC0AB5	1961	0.4683 -		ISM02.4	PPB	11/21/2017 17:57	SMPL
I6515-18	MC0AB6	681	0.1833 -		ISM02.4	PPB	11/21/2017 17:59	SMPL
I6515-19	MC0AB7	616	0.1688 -		ISM02.4	PPB	11/21/2017 18:01	SMPL
CCV070	CCV070	21837	4.8939 -		ISM02.4	PPB	11/21/2017 18:04	SMPL
CCB070	CCB070	-290	-0.0329 -		ISM02.4	PPB	11/21/2017 18:06	SMPL
I6515-20	MC0AB8	448	0.1314 -		ISM02.4	PPB	11/21/2017 18:08	SMPL
PB104444BL	PBS004	-162	-0.0044 -		ISM02.4	PPB	11/21/2017 18:11	SMPL
I6516-04	MC0AC0	3	0.0324 -		ISM02.4	PPB	11/21/2017 18:13	SMPL
I6516-05	MC0AC0D	-18	0.0277 -		ISM02.4	PPB	11/21/2017 18:15	SMPL
I6516-06	MC0AC0S	11118	2.5072 -		ISM02.4	PPB	11/21/2017 18:17	SMPL
I6516-07	MC0AC1	264	0.0905 -		ISM02.4	PPB	11/21/2017 18:20	SMPL
I6516-08	MC0AC2	89	0.0515 -		ISM02.4	PPB	11/21/2017 18:22	SMPL
I6516-09	MC0AC3	124	0.0593 -		ISM02.4	PPB	11/21/2017 18:24	SMPL
I6516-10	MC0AC4	330	0.1052 -		ISM02.4	PPB	11/21/2017 18:27	SMPL
I6516-11	MC0AC6	11	0.0341 -		ISM02.4	PPB	11/21/2017 18:29	SMPL
I6516-12	MC0AC7	254	0.0882 -		ISM02.4	PPB	11/21/2017 18:31	SMPL
I6516-13	MC0AC8	383	0.117 -		ISM02.4	PPB	11/21/2017 18:33	SMPL
I6516-14	MC0AC9	323	0.1036 -		ISM02.4	PPB	11/21/2017 18:36	SMPL
I6516-15	MC0AD0	306	0.0998 -		ISM02.4	PPB	11/21/2017 18:38	SMPL

LB91732

INSTRUMENT ID : CV1

CCV071	CCV071	22116	4.956 -	ISM02.4	PPB	11/21/2017 18:40 SMPL
CCB071	CCB071	-292	-0.0333 -	ISM02.4	PPB	11/21/2017 18:43 SMPL