


7470A

LB91740
INSTRUMENT ID: CVJ

Linear

 μ Abs.:

34509



A= 0.0000e+000

B= 2.9213e-004

C= 5.9454e-002

Rho= 0.9994970

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.029	0.029	-105	0.000	-105				
0.2	0.200	0.228	0.028	578	0.0 %	578				
2.5	2.500	2.467	-0.033	8243	0.0 %	8243				
5.0	5.000	5.069	0.069	17150	0.0 %	17150				
7.5	7.500	7.266	-0.234	24668	0.0 %	24668				
10.0	10.000	10.140	0.140	34509	0.0 %	34509				

LB91740 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Nov 2017 10:21:35

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-105	-	0.00007470A	PPB		22 Nov 2017 10:26:59	S	Std
0.2 - 1	50.2	578	-	0.20007470A	PPB		22 Nov 2017 10:29:08	S	Std
2.5 - 1	52.5	8243	-	2.50007470A	PPB		22 Nov 2017 10:31:16	S	Std
5.0 - 1	55	17150	-	5.00007470A	PPB		22 Nov 2017 10:33:24	S	Std
7.5 - 1	57.5	24668	-	7.50007470A	PPB		22 Nov 2017 10:35:33	S	Std
10.0 - 1	58	34509	-	10.00007470A	PPB		22 Nov 2017 10:37:41	S	Std
ICV76 - 1	ICV76	14019	4.1548	-7470A	PPB		22 Nov 2017 10:40:17	U	SMPL
ICB76 - 1	ICB76	-240	-0.0107	-7470A	PPB		22 Nov 2017 10:42:23	U	SMPL
CCV76 - 1	CCV76	17149	5.0691	-7470A	PPB		22 Nov 2017 10:44:30	U	SMPL
CCB76 - 1	CCB76	-227	-0.0069	-7470A	PPB		22 Nov 2017 10:46:37	U	SMPL
CRA - 1	CRA	571	0.2263	-7470A	PPB		22 Nov 2017 10:48:44	U	SMPL
HighStd - 1	HighStd	34890	10.2517	-7470A	PPB		22 Nov 2017 10:50:51	U	SMPL
ChkStd - 1	ChkStd	24204	7.1301	-7470A	PPB		22 Nov 2017 10:52:58	U	SMPL
PB104468BL - 1	PBW	-291	-0.0256	-7470A	PPB		22 Nov 2017 10:55:07	U	SMPL
PB104468BS - 1	LCSW	13418	3.9792	-7470A	PPB		22 Nov 2017 10:57:16	U	SMPL
PB104468TB - 1	PB104442TB	-221	-0.0051	-7470A	PPB		22 Nov 2017 10:59:24	U	SMPL
I6305-12 - 1	PL-02-112017-A	-132	0.0209	-7470A	PPB		22 Nov 2017 11:01:32	U	SMPL
I6305-14 - 1	PL-02-112017-B	-121	0.0241	-7470A	PPB		22 Nov 2017 11:03:41	U	SMPL
I6305-16 - 1	PL-02-112017-C	-28	0.0513	-7470A	PPB		22 Nov 2017 11:05:49	U	SMPL
I6310-06 - 1	AU-05-112017-A	-40	0.0478	-7470A	PPB		22 Nov 2017 11:07:56	U	SMPL
CCV77 - 1	CCV77	17373	5.1346	-7470A	PPB		22 Nov 2017 11:10:03	U	SMPL
CCB77 - 1	CCB77	-219	-0.0045	-7470A	PPB		22 Nov 2017 11:12:09	U	SMPL
I6310-08 - 1	AU-05-112017-B	-55	0.0434	-7470A	PPB		22 Nov 2017 11:14:17	U	SMPL
I6557-02 - 1	RT-3407-RT-1867-2077	-56	0.0431	-7470A	PPB		22 Nov 2017 11:16:23	U	SMPL
I6558-02 - 1	A1243	-50	0.0448	-7470A	PPB		22 Nov 2017 11:18:31	U	SMPL
I6558-04 - 1	SUMP-WELL-1	-130	0.0215	-7470A	PPB		22 Nov 2017 11:20:38	U	SMPL
I6558-04DUP - 1	SUMP-WELL-1DUP	-135	0.0200	-7470A	PPB		22 Nov 2017 11:22:45	U	SMPL
I6558-04MS - 1	SUMP-WELL-1MS	13681	4.0560	-7470A	PPB		22 Nov 2017 11:30:15	U	SMPL
I6558-04MSD - 1	SUMP-WELL-1MSD	13483	3.9982	-7470A	PPB		22 Nov 2017 11:32:22	U	SMPL
I6565-04 - 1	RT-3874	-169	0.0101	-7470A	PPB		22 Nov 2017 11:34:31	U	SMPL
PB104469BL - 1	PBW	-69	0.0393	-7470A	PPB		22 Nov 2017 11:36:40	U	SMPL
PB104469BS - 1	LCSW	13980	4.1434	-7470A	PPB		22 Nov 2017 11:41:12	U	SMPL
CCV78 - 1	CCV78	17582	5.1956	-7470A	PPB		22 Nov 2017 11:43:19	U	SMPL
CCB78 - 1	CCB78	-223	-0.0057	-7470A	PPB		22 Nov 2017 11:45:26	U	SMPL
I6558-01 - 1	A1243	4405	1.3463	-7470A	PPB		22 Nov 2017 11:53:57	U	SMPL
I6558-01DUP - 1	A1243DUP	4215	1.2908	-7470A	PPB		22 Nov 2017 11:56:04	U	SMPL
I6558-01MS - 1	A1243MS	15527	4.5953	-7470A	PPB		22 Nov 2017 11:58:11	U	SMPL
I6558-01MSD - 1	A1243MSD	15750	4.6604	-7470A	PPB		22 Nov 2017 12:00:19	U	SMPL
I6558-03 - 1	SUMP-WELL-1	123	0.0954	-7470A	PPB		22 Nov 2017 12:02:27	U	SMPL
I6558-04LX5 - 1		-131	0.0212	-7470A	PPB		22 Nov 2017 12:04:36	U	SMPL
I6558-04A - 1		13571	4.0239	-7470A	PPB		22 Nov 2017 12:06:43	U	SMPL
I6558-01LX5 - 1		825	0.3005	-7470A	PPB		22 Nov 2017 12:12:28	U	SMPL
I6558-01A - 1		19682	5.8091	-7470A	PPB		22 Nov 2017 12:14:36	U	SMPL
CCV79 - 1	CCV79	17325	5.1205	-7470A	PPB		22 Nov 2017 12:16:44	U	SMPL
CCB79 - 1	CCB79	-250	-0.0136	-7470A	PPB		22 Nov 2017 12:18:51	U	SMPL