

LB 91744


7471A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41147



A= 0.0000e+000

B= 2.4203e-004

C= 1.8105e-002

Rho= 0.9999683

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.009	-0.009	-113	0.000	-113				
0.2	0.200	0.196	-0.004	737	0.0 %	737				
2.5	2.500	2.477	-0.023	10159	0.0 %	10159				
5.0	5.000	5.063	0.063	20845	0.0 %	20845				
7.5	7.500	7.496	-0.004	30896	0.0 %	30896	0			
10.0	10.000	9.977	-0.023	41147	0.0 %	41147				

LB91744 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 22 Nov 2017 12:41:35

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-113	-	0.00007471A	PPB		22 Nov 2017 12:45:00	S	Std
0.2 - 1	50.2	737	-	0.20007471A	PPB		22 Nov 2017 12:47:17	S	Std
2.5 - 1	52.5	10159	-	2.50007471A	PPB		22 Nov 2017 12:49:35	S	Std
5.0 - 1	55	20845	-	5.00007471A	PPB		22 Nov 2017 12:51:52	S	Std
7.5 - 1	57.5	30896	-	7.50007471A	PPB		22 Nov 2017 12:54:10	S	Std
10.0 - 1	510	41147	-	10.00007471A	PPB		22 Nov 2017 12:59:08	S	Std
ICV77 - 1	ICV77	16413	3.9905	-7471A	PPB		22 Nov 2017 13:01:52	U	SMPL
ICB77 - 1	ICB77	-265	-0.0460	-7471A	PPB		22 Nov 2017 13:04:08	U	SMPL
CCV80 - 1	CCV80	21032	5.1085	-7471A	PPB		22 Nov 2017 13:06:23	U	SMPL
CCB80 - 1	CCB80	-286	-0.0511	-7471A	PPB		22 Nov 2017 13:08:39	U	SMPL
CRA - 1	CRA	695	0.1863	-7471A	PPB		22 Nov 2017 13:10:56	U	SMPL
HighStd - 1	HighStd	41353	10.0267	-7471A	PPB		22 Nov 2017 13:13:12	U	SMPL
ChkStd - 1	ChkStd	28519	6.9205	-7471A	PPB		22 Nov 2017 13:15:28	U	SMPL
PB104472BL - 1	PBS	-117	-0.0102	-7471A	PPB		22 Nov 2017 13:17:46	U	SMPL
PB104472BS - 1	LCSS	14708	3.5779	-7471A	PPB		22 Nov 2017 13:20:04	U	SMPL
I6043-15 - 1	N-T0A-BAJA-S001-0001-01	890	0.2335	-7471A	PPB		22 Nov 2017 13:22:21	U	SMPL
I6043-15DUP - 1	N-T0A-BAJA-S001-0001-01DUP	1075	0.2783	-7471A	PPB		22 Nov 2017 13:24:38	U	SMPL
I6043-16 - 1	I6043-15MS	17096	4.1558	-7471A	PPB		22 Nov 2017 13:26:56	U	SMPL
I6043-17 - 1	I6043-15MSD	16476	4.0058	-7471A	PPB		22 Nov 2017 13:29:13	U	SMPL
I6043-18 - 1	N-T0A-BAJA-S001-0001-02	1854	0.4668	-7471A	PPB		22 Nov 2017 13:31:29	U	SMPL
CCV81 - 1	CCV81	20909	5.0787	-7471A	PPB		22 Nov 2017 13:33:46	U	SMPL
CCB81 - 1	CCB81	-173	-0.0238	-7471A	PPB		22 Nov 2017 13:36:01	U	SMPL
I6043-19 - 1	N-T0A-BAJA-S001-0001-03	1121	0.2894	-7471A	PPB		22 Nov 2017 13:38:18	U	SMPL
I6289-07 - 1	SU-04-112117	1327	0.3393	-7471A	PPB		22 Nov 2017 13:40:34	U	SMPL
I6570-01 - 1	P003-S001-0001-01	7665	1.8733	-7471A	PPB		22 Nov 2017 13:42:50	U	SMPL
I6570-02 - 1	I6570-01MS	20311	4.9340	-7471A	PPB		22 Nov 2017 13:45:06	U	SMPL
I6570-03 - 1	I6570-01MSD	21144	5.1348	-7471A	PPB		22 Nov 2017 13:47:23	U	SMPL
I6570-04 - 1	P003-S001-0001-02	3317	0.8209	-7471A	PPB		22 Nov 2017 13:49:40	U	SMPL
I6570-05 - 1	P003-S002-0001-01	3797	0.9371	-7471A	PPB		22 Nov 2017 13:51:59	U	SMPL
I6570-06 - 1	P003-S003-0001-01	3173	0.7861	-7471A	PPB		22 Nov 2017 13:54:16	U	SMPL
I6574-01 - 1	EPE-109-5-(0-5)	9172	2.2380	-7471A	PPB		22 Nov 2017 13:56:34	U	SMPL
I6574-02 - 1	EPE-109-5-(5-10)	1820	0.4586	-7471A	PPB		22 Nov 2017 13:58:50	U	SMPL
CCV82 - 1	CCV82	20822	5.0576	-7471A	PPB		22 Nov 2017 14:01:07	U	SMPL
CCB82 - 1	CCB82	-160	-0.0206	-7471A	PPB		22 Nov 2017 14:03:24	U	SMPL
I6574-03 - 1	EPE-120-4-(0-5)	9982	2.4340	-7471A	PPB		22 Nov 2017 14:05:41	U	SMPL
I6574-04 - 1	EPE-120-4-(5-10)	5695	1.3965	-7471A	PPB		22 Nov 2017 14:07:57	U	SMPL
I6043-15LX5 - 1		107	0.0440	-7471A	PPB		22 Nov 2017 14:10:14	U	SMPL
I6043-15A - 1		17317	4.2093	-7471A	PPB		22 Nov 2017 14:12:31	U	SMPL
CCV83 - 1	CCV83	20609	5.0061	-7471A	PPB		22 Nov 2017 14:14:47	U	SMPL
CCB83 - 1	CCB83	-180	-0.0255	-7471A	PPB		22 Nov 2017 14:17:04	U	SMPL