



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91757

Temp.: 17.2 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/22/2017	16:00
2	CAL2	1	Water			7.00	11/22/2017	16:01
3	CAL3	1	Water			10.00	11/22/2017	16:02
4	ICV	1	Water			6.99	11/22/2017	16:03
5	CCV1	1	Water			2.01	11/22/2017	16:05
6	I6583-02	1	Water			6.14	11/22/2017	16:06
7	I6583-02DUP	1	Water			6.15	11/22/2017	16:07
8	CCV2	1	Water			12.01	11/22/2017	16:10

2301757

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6583 WorkList ID : 105816 Date : 11/22/2017 2:50:23 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/22/2017	Water	I6583-02-17	Corrosivity	Cool 4 deg C	PSEG01	M52	OILY-WATER-1	11/22/2017	9040C

Date/Time 11-22-17 2:50 PM
Received by: JP
Relinquished by:

Date/Time 11-22-17 4:35 PM
Received by: CP
Relinquished by: JP