



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91758

Temp.: 24.2 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/22/2017	16:00
2	CAL2	1	Water			7.00	11/22/2017	16:01
3	CAL3	1	Water			10.00	11/22/2017	16:02
4	ICV	1	Water			6.99	11/22/2017	16:05
5	CCV1	1	Water			2.01	11/22/2017	16:30
6	I6289-08	1	Solid	20.02	20	7.66	11/22/2017	16:31
7	I6289-08DUP	1	Solid	20.03	20	7.67	11/22/2017	16:32
8	I6299-08	1	Solid	20.04	20	9.25	11/22/2017	16:33
9	I6582-02	1	Solid	20.03	20	8.30	11/22/2017	16:34
10	I6594-02	1	Solid	20.05	20	11.52	11/22/2017	16:35
11	CCV2	1	Water			12.02	11/22/2017	16:37

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6594

WorkList ID : 105815

Date : 11/22/2017 2:49:53 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/21/2017	Solid	I6289-08	Corrosivity	Cool 4 deg C	PSEG05	M52	SU-04-112117	11/21/2017	9045C
11/22/2017	Solid	I6299-08	Corrosivity	Cool 4 deg C	PSEG05	M52	OK-03-112217	11/22/2017	9045C
11/22/2017	Solid	I6582-02	Corrosivity	Cool 4 deg C	PSEG02	M52	OILY-DEBRIS	11/22/2017	9045C
11/22/2017	Solid	I6594-02	Corrosivity	Cool 4 deg C	PSEG01	M52	KUSER-CONCRETE	11/22/2017	9045C

Date/Time

Received by:

Relinquished by:

11/22/2017 2:50 PM
JP
CP

Date/Time

Received by:

Relinquished by:

11-22-17 5:35 PM
CP
JP