



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB91771
Analysis Date: 11/27/2017
BalanceID: WC SC-4
OvenID: EX Oven-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 11/27/2017
Extraction IN Time: 08:00
Extraction OUT Time: 08:40

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB91771BL	LB91771BL	WATER	2	1000	100	2.733	2.733	3.02	2.733	2.733	0	0
2	LB91771BS	LB91771BS	WATER	2	1000	100	2.8411	2.8411	3.04	2.8596	2.8596	0.0185	18.5
3	LB91771BSD	LB91771BSD	WATER	2	1000	100	2.7436	2.7436	3.01	2.7622	2.7622	0.0186	18.6
4	I6583-01	OILY-WATER-1	WATER	2	1000	100	2.9897	2.9897	3.05	3.4344	3.4344	0.4447	444.7
5	I6597-01	STORM-DRAIN-1	WATER	2	1000	100	3.005	3.005	3.04	3.0077	3.0077	0.0027	2.7
6	I6597-01DUP	STORM-DRAIN-1DUP	WATER	2	1000	100	3.1997	3.1997	3.01	3.2028	3.2028	0.0031	3.1
7	I6597-02	STORM-DRAIN-2	WATER	2	1000	100	2.9572	2.9572	3.02	2.9582	2.9582	0.001	1
8	I6597-02MS	STORM-DRAIN-2	WATER	2	1000	100	2.8103	2.8103	3.05	2.8314	2.8314	0.0211	21.1



QC Batch# LB91771 Test: TPH Analysis Date: 11/27/2017

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2288
pH Paper 0-14	W2287
Sodium Sulfate	EP1795
1:1 HCL	WP59249
Silica Gel	W2199
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP58006
LCSWD	10.00 ML	WP58005
MS/MSD	10.00 ML	WP58006

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0018 In OVEN TEMP : 70 Dessicator Time In : 11:00
1.0000 gram Balance: 1.0005 In Time: 09:20
Bal Check Time: 09:20

After Analysis

0.0020 gram Balance: 0.0021 Out OVEN TEMP: 71 Dessicator Time Out: 11:00
1.0000 gram Balance: 1.0003 Out Time: 10:20
Bal Check Time: 12:00

LB9177

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH I6594

WorkList ID : 105821

Date : 11/27/2017 7:43:38 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/26/2017	Water	I6583-01-H	TPH	Conc H2SO4 tr	PSEG01	M52	OILY-WATER-1	11/22/2017	1664A
11/29/2017	Water	I6597-01-E, F	TPH	Conc H2SO4 tr	KWAR01	M33	STORM-DRAIN-1	11/22/2017	1664A
11/29/2017	Water	I6597-02 E, F	TPH	Conc H2SO4 tr	KWAR01	M33	STORM-DRAIN-2	11/22/2017	1664A

Date/Time 11-27-17 7:42 PM
Received by: JP
Relinquished by: JP

Date/Time 11-27-17 10:00 AM
Received by: JP
Relinquished by: JP