


7471A

LB91777
INSTRUMENT ID: CVL

Linear

 μ Abs.:

42816



A= 0.0000e+000

B= 2.3308e-004

C= 4.1234e-002

Rho= 0.9999380

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.023	0.023	-78	0.000	-78				
0.2	0.200	0.217	0.017	752	0.0 %	752				
2.5	2.500	2.498	-0.002	10542	0.0 %	10542				
5.0	5.000	4.911	-0.089	20892	0.0 %	20892				
7.5	7.500	7.530	0.030	32131	0.0 %	32131	0			
10.0	10.000	10.021	0.021	42816	0.0 %	42816				

LB91777 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 27 Nov 2017 10:25:37

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	-78	-	0.00007471A	PPB	27 Nov 2017 11:16:47	S	Std
0.2 - 1	50.2	752	-	0.20007471A	PPB	27 Nov 2017 11:19:05	S	Std
2.5 - 1	52.5	10542	-	2.50007471A	PPB	27 Nov 2017 11:21:23	S	Std
5.0 - 1	55	20892	-	5.00007471A	PPB	27 Nov 2017 11:23:40	S	Std
7.5 - 1	57.5	32131	-	7.50007471A	PPB	27 Nov 2017 11:25:57	S	Std
10.0 - 1	510	42816	-	10.00007471A	PPB	27 Nov 2017 11:28:15	S	Std
ICV78 - 1	ICV78	16146	3.8046	-7471A	PPB	27 Nov 2017 11:30:58	U	SMPL
ICB78 - 1	ICB78	-214	-0.0086	-7471A	PPB	27 Nov 2017 11:33:14	U	SMPL
CCV88 - 1	CCV88	19373	4.5567	-7471A	PPB	27 Nov 2017 11:35:29	U	SMPL
CCB88 - 1	CCB88	-224	-0.0110	-7471A	PPB	27 Nov 2017 11:37:45	U	SMPL
CRA - 1	CRA	734	0.2123	-7471A	PPB	27 Nov 2017 11:40:02	U	SMPL
HighStd - 1	HighStd	42676	9.9882	-7471A	PPB	27 Nov 2017 11:42:18	U	SMPL
ChkStd - 1	ChkStd	29016	6.8043	-7471A	PPB	27 Nov 2017 11:44:33	U	SMPL
PB104509BL - 1	PBS	-283	-0.0247	-7471A	PPB	27 Nov 2017 11:46:52	U	SMPL
PB104509BS - 1	LCSS	15122	3.5659	-7471A	PPB	27 Nov 2017 11:49:10	U	SMPL
I6299-07 - 1	OK-03-112217	906	0.2524	-7471A	PPB	27 Nov 2017 11:51:27	U	SMPL
I6578-04 - 1	BBR-2-E(0-1)	3718	0.9078	-7471A	PPB	27 Nov 2017 11:53:44	U	SMPL
I6578-05 - 1	BBR-2-E(4-5)	131466	30.6836	-7471A	PPB	27 Nov 2017 11:56:02	U	SMPL
I6578-06 - 1	BBR-2-E(7-8)	56022	13.0990	-7471A	PPB	27 Nov 2017 12:12:22	U	SMPL
I6578-06DUP - 1	BBR-2-E(7-8)DUP	55016	12.8645	-7471A	PPB	27 Nov 2017 12:14:40	U	SMPL
CCV89 - 1	CCV89	20031	4.7101	-7471A	PPB	27 Nov 2017 12:16:58	U	SMPL
CCB89 - 1	CCB89	-273	-0.0224	-7471A	PPB	27 Nov 2017 12:19:16	U	SMPL
I6578-06MS - 1	BBR-2-E(7-8)MS	72488	16.9369	-7471A	PPB	27 Nov 2017 12:21:33	U	SMPL
I6578-06MSD - 1	BBR-2-E(7-8)MSD	73325	17.1320	-7471A	PPB	27 Nov 2017 12:23:49	U	SMPL
I6579-01 - 1	HAR-4-E(0-1)	18001	4.2369	-7471A	PPB	27 Nov 2017 12:26:08	U	SMPL
I6579-02 - 1	HAR-4-E(4-5)	16653	3.9227	-7471A	PPB	27 Nov 2017 12:28:27	U	SMPL
I6579-03 - 1	HAR-4-E(7-8)	1522	0.3960	-7471A	PPB	27 Nov 2017 12:30:44	U	SMPL
I6579-04 - 1	HAR-5-E(0-1)	1054	0.2869	-7471A	PPB	27 Nov 2017 12:33:01	U	SMPL
I6579-05 - 1	HAR-5-E(4-5)	2251	0.5659	-7471A	PPB	27 Nov 2017 12:35:18	U	SMPL
I6579-06 - 1	HAR-5-E(7-8)	5526	1.3292	-7471A	PPB	27 Nov 2017 12:37:36	U	SMPL
I6579-07 - 1	HAR-3-E(0-1)	10295	2.4408	-7471A	PPB	27 Nov 2017 12:39:54	U	SMPL
I6579-08 - 1	HAR-3-E(7-8)	5440	1.3092	-7471A	PPB	27 Nov 2017 12:42:10	U	SMPL
CCV90 - 1	CCV90	21123	4.9646	-7471A	PPB	27 Nov 2017 12:44:26	U	SMPL
CCB90 - 1	CCB90	-216	-0.0091	-7471A	PPB	27 Nov 2017 12:46:43	U	SMPL
I6579-09 - 1	HAR-2-E(0-1)	18750	4.4115	-7471A	PPB	27 Nov 2017 12:49:00	U	SMPL
I6579-10 - 1	HAR-2-E(4-5)	25297	5.9375	-7471A	PPB	27 Nov 2017 12:51:16	U	SMPL
I6579-11 - 1	HAR-2-E(7-8)	24933	5.8527	-7471A	PPB	27 Nov 2017 12:53:34	U	SMPL
PB104510BL - 1	PBS	-249	-0.0168	-7471A	PPB	27 Nov 2017 12:55:52	U	SMPL
PB104510BS - 1	LCSS	13788	3.2550	-7471A	PPB	27 Nov 2017 12:58:09	U	SMPL
6570-01 - 1	P003-S001-0001-01	5080	1.2253	-7471A	PPB	27 Nov 2017 13:00:26	U	SMPL
6570-01DUP - 1	P003-S001-0001-01DUP	4480	1.0854	-7471A	PPB	27 Nov 2017 13:02:43	U	SMPL
6570-02 - 1	I6570-01MS	19464	4.5779	-7471A	PPB	27 Nov 2017 13:05:01	U	SMPL
6570-03 - 1	I6570-01MSD	19438	4.5719	-7471A	PPB	27 Nov 2017 13:07:19	U	SMPL
6590-01 - 1	BSC-SB08-FD	365	0.1263	-7471A	PPB	27 Nov 2017 13:09:35	U	SMPL
CCV91 - 1	CCV91	21497	5.0518	-7471A	PPB	27 Nov 2017 13:11:52	U	SMPL
CCB91 - 1	CCB91	-193	-0.0038	-7471A	PPB	27 Nov 2017 13:14:09	U	SMPL
6590-02 - 1	BSC-SB08	474	0.1517	-7471A	PPB	27 Nov 2017 13:16:26	U	SMPL
6590-03 - 1	BSC-SB-07	1610	0.4165	-7471A	PPB	27 Nov 2017 13:18:43	U	SMPL
6590-04 - 1	BSC-SB-05	292	0.1093	-7471A	PPB	27 Nov 2017 13:20:59	U	SMPL
6592-01 - 1	EPE-109-5-(44-46)	605	0.1822	-7471A	PPB	27 Nov 2017 13:23:15	U	SMPL
6594-01 - 1	KUSER-CONCRETE	10	0.0436	-7471A	PPB	27 Nov 2017 13:25:32	U	SMPL
6578-06LX5 - 1		10623	2.5173	-7471A	PPB	27 Nov 2017 13:27:49	U	SMPL
6578-06A - 1		70443	16.4602	-7471A	PPB	27 Nov 2017 13:30:06	U	SMPL
6570-01LX5 - 1		725	0.2102	-7471A	PPB	27 Nov 2017 13:35:24	U	SMPL
6570-01A - 1		21428	5.0357	-7471A	PPB	27 Nov 2017 13:37:41	U	SMPL
CCV92 - 1	CCV92	20444	4.8064	-7471A	PPB	27 Nov 2017 13:39:58	U	SMPL
CCB92 - 1	CCB92	-211	-0.0079	-7471A	PPB	27 Nov 2017 13:42:15	U	SMPL
6578-06DLX5 - 1		10732	2.5427	-7471A	PPB	27 Nov 2017 13:44:32	U	SMPL
6578-06DUPDLX5 - 1		10635	2.5201	-7471A	PPB	27 Nov 2017 13:46:48	U	SMPL
6578-06MSDLX5 - 1		14638	3.4531	-7471A	PPB	27 Nov 2017 13:49:05	U	SMPL
6578-06MSDDLX5 - 1		14754	3.4801	-7471A	PPB	27 Nov 2017 13:51:21	U	SMPL
6578-06DLX25 - 1		2065	0.5225	-7471A	PPB	27 Nov 2017 13:53:38	U	SMPL
6578-06ADLX5 - 1		14477	3.4156	-7471A	PPB	27 Nov 2017 13:55:54	U	SMPL
CCV93 - 1	CCV93	20384	4.7924	-7471A	PPB	27 Nov 2017 13:58:11	U	SMPL
CCB93 - 1	CCB93	-162	0.0035	-7471A	PPB	27 Nov 2017 14:00:28	U	SMPL