



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91796

Temp.: 25.2 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/27/2017	17:30
2	CAL2	1	Water			7.00	11/27/2017	17:31
3	CAL3	1	Water			10.02	11/27/2017	17:33
4	ICV	1	Water			6.99	11/27/2017	17:34
5	CCV1	1	Water			2.01	11/27/2017	17:35
6	I6602-03	1	Solid	20.02	20	9.74	11/27/2017	17:36
7	I6602-03DUP	1	Solid	20.04	20	9.75	11/27/2017	17:38
8	CCV2	1	Water			12.02	11/27/2017	17:39

2B91796

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I6602

WorkList ID : 105882

Date : 11/27/2017 4:58:46 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/27/2017	Solid	I6602-03	Corrosivity	Cool 4 deg C	PSEG01	M52	ETGI-294	11/27/2017	9045C

Date/Time 11-27-17 6:58 PM
Received by: JD
Relinquished by: JD

Date/Time 11-27-17 6:10 PM
Received by: JD
Relinquished by: JD