


7470A

LB91807
INSTRUMENT ID: CVZ

Linear

 μ Abs.:

34388



A= 0.0000e+000

B= 2.9211e-004

C= 1.6186e-002

Rho= 0.9999076

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.000	-0.000	-57	0.000	-57				
0.2	0.200	0.206	0.006	649	0.0 %	649				
2.5	2.500	2.515	0.015	8555	0.0 %	8555				
5.0	5.000	5.021	0.021	17135	0.0 %	17135				
7.5	7.500	7.397	-0.103	25267	0.0 %	25267				
10.0	10.000	10.061	0.061	34388	0.0 %	34388				

LB91807 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Nov 2017 10:20:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-57	-	0.0000	7470A	PPB	28 Nov 2017 10:23:50	S	Std
0.2 - 1	50.2	649	-	0.2000	7470A	PPB	28 Nov 2017 10:25:59	S	Std
2.5 - 1	52.5	8555	-	2.5000	7470A	PPB	28 Nov 2017 10:28:07	S	Std
5.0 - 1	55	17135	-	5.0000	7470A	PPB	28 Nov 2017 10:30:16	S	Std
7.5 - 1	57.5	25267	-	7.5000	7470A	PPB	28 Nov 2017 10:32:23	S	Std
10.0 - 1	510	34388	-	10.0000	7470A	PPB	28 Nov 2017 10:34:32	S	Std
ICV79 - 1	ICV79	13102	3.8434	-	7470A	PPB	28 Nov 2017 10:37:13	U	SMPL
ICB79 - 1	ICB79	-187	-0.0384	-	7470A	PPB	28 Nov 2017 10:39:19	U	SMPL
CCV94 - 1	CCV94	17506	5.1298	-	7470A	PPB	28 Nov 2017 10:41:25	U	SMPL
CCB94 - 1	CCB94	-177	-0.0355	-	7470A	PPB	28 Nov 2017 10:43:32	U	SMPL
CRA - 1	CRA	654	0.2072	-	7470A	PPB	28 Nov 2017 10:45:39	U	SMPL
HighStd - 1	HighStd	35133	10.2788	-	7470A	PPB	28 Nov 2017 10:47:46	U	SMPL
ChkStd - 1	ChkStd	23614	6.9140	-	7470A	PPB	28 Nov 2017 10:49:53	U	SMPL
PB104547BL - 1	PBW	-256	-0.0586	-	7470A	PPB	28 Nov 2017 10:52:02	U	SMPL
PB104547BS - 1	LCSW	13148	3.8568	-	7470A	PPB	28 Nov 2017 10:54:11	U	SMPL
PB104547TB - 1	PB104466TB	-161	-0.0308	-	7470A	PPB	28 Nov 2017 10:56:19	U	SMPL
I6289-08 - 1	SU-04-112117	-82	-0.0078	-	7470A	PPB	28 Nov 2017 10:58:27	U	SMPL
I6299-08 - 1	OK-03-112217	-71	-0.0046	-	7470A	PPB	28 Nov 2017 11:00:35	U	SMPL
I6299-08DUP - 1	OK-03-112217DUP	-54	0.0004	-	7470A	PPB	28 Nov 2017 11:02:44	U	SMPL
I6299-08MS - 1	OK-03-112217MS	12009	3.5241	-	7470A	PPB	28 Nov 2017 11:04:50	U	SMPL
CCV95 - 1	CCV95	17886	5.2408	-	7470A	PPB	28 Nov 2017 11:06:57	U	SMPL
CCB95 - 1	CCB95	-186	-0.0381	-	7470A	PPB	28 Nov 2017 11:09:03	U	SMPL
I6299-08MSD - 1	OK-03-112217MSD	12356	3.6255	-	7470A	PPB	28 Nov 2017 11:11:11	U	SMPL
I6582-02 - 1	OILY-DEBRIS	-141	-0.0250	-	7470A	PPB	28 Nov 2017 11:13:17	U	SMPL
I6583-02 - 1	OILY-WATER-1	-74	-0.0054	-	7470A	PPB	28 Nov 2017 11:15:25	U	SMPL
I6594-02 - 1	KUSER-CONCRETE	-50	0.0016	-	7470A	PPB	28 Nov 2017 11:17:32	U	SMPL
PB104548BL - 1	PBW	-45	0.0030	-	7470A	PPB	28 Nov 2017 11:19:40	U	SMPL
PB104548BS - 1	LCSW	13176	3.8650	-	7470A	PPB	28 Nov 2017 11:21:48	U	SMPL
I6583-01 - 1	OILY-WATER-1	-133	-0.0227	-	7470A	PPB	28 Nov 2017 11:23:56	U	SMPL
I6591-01 - 1	MW-8R	-62	-0.0019	-	7470A	PPB	28 Nov 2017 11:26:04	U	SMPL
I6591-01DUP - 1	MW-8RDUP	-65	-0.0028	-	7470A	PPB	28 Nov 2017 11:28:13	U	SMPL
I6591-01MS - 1	MW-8RMS	13217	3.8770	-	7470A	PPB	28 Nov 2017 11:30:20	U	SMPL
CCV96 - 1	CCV96	17584	5.1526	-	7470A	PPB	28 Nov 2017 11:32:27	U	SMPL
CCB96 - 1	CCB96	-178	-0.0358	-	7470A	PPB	28 Nov 2017 11:34:35	U	SMPL
I6591-01MSD - 1	MW-8RMSD	13468	3.9503	-	7470A	PPB	28 Nov 2017 11:36:43	U	SMPL
I6591-02 - 1	MW-9R	-131	-0.0221	-	7470A	PPB	28 Nov 2017 11:38:50	U	SMPL
I6591-03 - 1	SW-01	-88	-0.0095	-	7470A	PPB	28 Nov 2017 11:40:57	U	SMPL
I6591-04 - 1	SW-02	-72	-0.0048	-	7470A	PPB	28 Nov 2017 11:43:04	U	SMPL
I6591-05 - 1	SW-03	-68	-0.0037	-	7470A	PPB	28 Nov 2017 11:45:11	U	SMPL
I6299-08LX5 - 1		-53	0.0007	-	7470A	PPB	28 Nov 2017 11:47:19	U	SMPL
I6299-08A - 1		13575	3.9816	-	7470A	PPB	28 Nov 2017 11:49:27	U	SMPL
I6591-01LX5 - 1		-152	-0.0282	-	7470A	PPB	28 Nov 2017 11:51:35	U	SMPL
I6591-01A - 1		12912	3.7879	-	7470A	PPB	28 Nov 2017 11:53:44	U	SMPL
CCV97 - 1	CCV97	17783	5.2107	-	7470A	PPB	28 Nov 2017 11:55:51	U	SMPL
CCB97 - 1	CCB97	-180	-0.0364	-	7470A	PPB	28 Nov 2017 11:57:59	U	SMPL