



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB91808

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	11/28/2017	09:15
2	I6583-02		1	>100.0	>212.0	11/28/2017	09:50
3	I6583-02DUP		1	>100.0	>212.0	11/28/2017	10:15

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP112817

WorkList ID : 105893

Date : 11/28/2017 8:30:04 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/26/2017	Water	I6583-02	C Flash Point	Cool 4 deg C	PSEG01	M52	OILY-WATER-1	11/22/2017	1010A

Date/Time 11/28/17 8:35 AM
 Received by: KS
 Relinquished by: CP

Date/Time 11/28/17 10:20 AM
 Received by: CP
 Relinquished by: KS

LB91808