

7471A


LB91811

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41975



A= 0.0000e+000

B= 2.3761e-004

C= 5.8325e-002

Rho= 0.9999125

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.041	0.041	-72	0.000	-72				
0.2	0.200	0.228	0.028	715	0.0 %	715				
2.5	2.500	2.462	-0.038	10116	0.0 %	10116				
5.0	5.000	4.908	-0.092	20409	0.0 %	20409				
7.5	7.500	7.529	0.029	31439	0.0 %	31439	0			
10.0	10.000	10.032	0.032	41975	0.0 %	41975				

LB91811 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 28 Nov 2017 12:44:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-72	-	0.0000	7471A	PPB	28 Nov 2017 12:45:59	S	Std
0.2 - 1	20.2	715	-	0.2000	7471A	PPB	28 Nov 2017 12:48:17	S	Std
2.5 - 1	52.5	10116	-	2.5000	7471A	PPB	28 Nov 2017 12:50:34	S	Std
5.0 - 1	55	20409	-	5.0000	7471A	PPB	28 Nov 2017 12:52:52	S	Std
7.5 - 1	57.5	31439	-	7.5000	7471A	PPB	28 Nov 2017 12:55:09	S	Std
10.0 - 1	510	41975	-	10.0000	7471A	PPB	28 Nov 2017 12:57:27	S	Std
ICV80 - 1	ICV80	15900	3.8364	-	7471A	PPB	28 Nov 2017 13:00:08	U	SMPL
ICB80 - 1	ICB80	-192	0.0127	-	7471A	PPB	28 Nov 2017 13:02:24	U	SMPL
CCV98 - 1	CCV98	20378	4.9004	-	7471A	PPB	28 Nov 2017 13:04:40	U	SMPL
CCB98 - 1	CCB98	-201	0.0106	-	7471A	PPB	28 Nov 2017 13:06:56	U	SMPL
CRA - 1	CRA	703	0.2254	-	7471A	PPB	28 Nov 2017 13:09:12	U	SMPL
HighStd - 1	HighStd	42338	10.1184	-	7471A	PPB	28 Nov 2017 13:11:27	U	SMPL
ChkStd - 1	ChkStd	29321	7.0254	-	7471A	PPB	28 Nov 2017 13:13:44	U	SMPL
PB104560BL - 1	PBS	-256	-0.0025	-	7471A	PPB	28 Nov 2017 13:16:02	U	SMPL
PB104560BS - 1	LCSS	15803	3.8133	-	7471A	PPB	28 Nov 2017 13:20:53	U	SMPL
I6602-01 - 1	ETGI-294	778	0.2432	-	7471A	PPB	28 Nov 2017 13:34:27	U	SMPL
I6602-01DUP - 1	ETGI-294DUP	774	0.2422	-	7471A	PPB	28 Nov 2017 13:36:43	U	SMPL
I6602-01MS - 1	ETGI-294MS	14275	3.4503	-	7471A	PPB	28 Nov 2017 13:39:01	U	SMPL
I6602-01MSD - 1	ETGI-294MSD	14876	3.5931	-	7471A	PPB	28 Nov 2017 13:41:19	U	SMPL
I6578-05DLX5 - 1		25067	6.0146	-	7471A	PPB	28 Nov 2017 13:43:35	U	SMPL
CCV99 - 1	CCV99	20194	4.8567	-	7471A	PPB	28 Nov 2017 13:45:51	U	SMPL
CCB99 - 1	CCB99	-214	0.0075	-	7471A	PPB	28 Nov 2017 13:48:08	U	SMPL
I6602-01LX5 - 1		85	0.0785	-	7471A	PPB	28 Nov 2017 13:50:24	U	SMPL
I6602-01A - 1		15738	3.7979	-	7471A	PPB	28 Nov 2017 13:52:40	U	SMPL
CCV100 - 1	CCV100	20330	4.8890	-	7471A	PPB	28 Nov 2017 13:54:56	U	SMPL
CCB100 - 1	CCB100	-202	0.0103	-	7471A	PPB	28 Nov 2017 13:57:12	U	SMPL