


245.1

LB91815
INSTRUMENT 10: CV 1

Linear

 μ Abs.:

33986



A= 0.0000e+000

B= 2.9464e-004

C= 3.5066e-002

Rho= 0.9999273

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.018	0.018	-58	0.000	-58				
0.2	0.200	0.221	0.021	630	0.0 %	630				
2.5	2.500	2.515	0.015	8418	0.0 %	8418				
5.0	5.000	4.908	-0.092	16540	0.0 %	16540				
7.5	7.500	7.489	-0.011	25297	0.0 %	25297				
10.0	10.000	10.049	0.049	33986	0.0 %	33986				

LB91815 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 28 Nov 2017 14:18:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-58	-	0.0000245.1		PPB	28 Nov 2017 14:22:39	S	Std
0.2 - 1	50.2	630	-	0.2000245.1		PPB	28 Nov 2017 14:24:47	S	Std
2.5 - 1	52.5	8418	-	2.5000245.1		PPB	28 Nov 2017 14:26:56	S	Std
5.0 - 1	55	16540	-	5.0000245.1		PPB	28 Nov 2017 14:29:04	S	Std
7.5 - 1	57.5	25297	-	7.5000245.1		PPB	28 Nov 2017 14:31:12	S	Std
10.0 - 1	510	33986	-	10.0000245.1		PPB	28 Nov 2017 14:33:20	S	Std
ICV81 - 1	ICV81	13335	3.9641	-245.1		PPB	28 Nov 2017 14:35:59	U	SMPL
ICB81 - 1	ICB81	-186	-0.0197	-245.1		PPB	28 Nov 2017 14:38:05	U	SMPL
CCV01 - 1	CCV01	16880	5.0086	-245.1		PPB	28 Nov 2017 14:40:11	U	SMPL
CCB01 - 1	CCB01	-176	-0.0168	-245.1		PPB	28 Nov 2017 14:42:18	U	SMPL
CRA - 1	CRA	619	0.2175	-245.1		PPB	28 Nov 2017 14:44:25	U	SMPL
HighStd - 1	HighStd	33782	9.9887	-245.1		PPB	28 Nov 2017 14:46:32	U	SMPL
ChkStd - 1	ChkStd	22936	6.7930	-245.1		PPB	28 Nov 2017 14:48:38	U	SMPL
PB104545BL - 1	PBW	-239	-0.0354	-245.1		PPB	28 Nov 2017 14:50:48	U	SMPL
PB104545BS - 1	LCSW	13663	4.0608	-245.1		PPB	28 Nov 2017 14:52:56	U	SMPL
I6567-01 - 1	112017-TIDO-F-D-B	-138	0.0056	-245.1		PPB	28 Nov 2017 14:55:04	U	SMPL
I6567-02 - 1	112017-TIDO-F-D-S	-68	0.0150	-245.1		PPB	28 Nov 2017 14:57:11	U	SMPL
I6567-03 - 1	112017-TIDO-F-U-B	-48	0.0209	-245.1		PPB	28 Nov 2017 14:59:19	U	SMPL
I6567-04 - 1	112017-TIDO-F-U-S	-52	0.0197	-245.1		PPB	28 Nov 2017 15:01:28	U	SMPL
I6567-05 - 1	112017-JETT-F-D-B	-50	0.0203	-245.1		PPB	28 Nov 2017 15:03:35	U	SMPL
CCV02 - 1	CCV02	16692	4.9533	-245.1		PPB	28 Nov 2017 15:05:42	U	SMPL
CCB02 - 1	CCB02	-180	-0.0180	-245.1		PPB	28 Nov 2017 15:07:48	U	SMPL
I6567-06 - 1	112017-JETT-F-D-S	-65	0.0159	-245.1		PPB	28 Nov 2017 15:09:55	U	SMPL
I6567-07 - 1	112017-JETT-F-U-B	-48	0.0209	-245.1		PPB	28 Nov 2017 15:12:01	U	SMPL
I6567-08 - 1	112017-JETT-F-U-S	-41	0.0230	-245.1		PPB	28 Nov 2017 15:14:08	U	SMPL
I6584-01 - 1	112117-JETT-F-D-B	-52	0.0197	-245.1		PPB	28 Nov 2017 15:16:15	U	SMPL
I6584-02 - 1	112117-JETT-F-D-S	-44	0.0221	-245.1		PPB	28 Nov 2017 15:18:22	U	SMPL
I6584-03 - 1	112117-JETT-F-U-B	-50	0.0203	-245.1		PPB	28 Nov 2017 15:20:30	U	SMPL
I6584-04 - 1	112117-JETT-F-U-S	-42	0.0227	-245.1		PPB	28 Nov 2017 15:22:38	U	SMPL
I6584-05 - 1	112117-TIDO-F-D-B	-53	0.0195	-245.1		PPB	28 Nov 2017 15:24:47	U	SMPL
I6584-06 - 1	112117-TIDO-F-D-S	-54	0.0192	-245.1		PPB	28 Nov 2017 15:26:55	U	SMPL
I6584-07 - 1	112117-TIDO-F-U-B	-42	0.0227	-245.1		PPB	28 Nov 2017 15:29:02	U	SMPL
CCV03 - 1	CCV03	16870	5.0057	-245.1		PPB	28 Nov 2017 15:31:10	U	SMPL
CCB03 - 1	CCB03	-157	-0.0112	-245.1		PPB	28 Nov 2017 15:33:16	U	SMPL
I6584-08 - 1	112117-TIDO-F-U-S	-46	0.0215	-245.1		PPB	28 Nov 2017 15:35:24	U	SMPL
I6584-08DUP - 1	112117-TIDO-F-U-SDUP	-48	0.0209	-245.1		PPB	28 Nov 2017 15:37:31	U	SMPL
I6584-08MS - 1	112117-TIDO-F-U-SMS	16233	4.8180	-245.1		PPB	28 Nov 2017 15:39:38	U	SMPL
I6584-08MSD - 1	112117-TIDO-F-U-SMSD	15183	4.5086	-245.1		PPB	28 Nov 2017 15:41:45	U	SMPL
PB104546BL - 1	PBW	-161	-0.0124	-245.1		PPB	28 Nov 2017 15:43:53	U	SMPL
PB104546BS - 1	LCSW	13121	3.9011	-245.1		PPB	28 Nov 2017 15:46:00	U	SMPL
6603-01 - 1	112217-TIDO-F-D-B	-114	0.0015	-245.1		PPB	28 Nov 2017 15:48:09	U	SMPL
6603-02 - 1	112217-TIDO-F-D-S	-87	0.0094	-245.1		PPB	28 Nov 2017 15:50:17	U	SMPL
6603-03 - 1	112217-TIDO-F-U-B	-55	0.0189	-245.1		PPB	28 Nov 2017 15:52:25	U	SMPL
6603-04 - 1	112217-TIDO-F-U-S	-67	0.0153	-245.1		PPB	28 Nov 2017 15:54:32	U	SMPL
CCV04 - 1	CCV04	16703	4.9565	-245.1		PPB	28 Nov 2017 15:56:39	U	SMPL
CCB04 - 1	CCB04	-170	-0.0150	-245.1		PPB	28 Nov 2017 15:58:47	U	SMPL
6603-05 - 1	112217-JETT-F-D-B	-65	0.0159	-245.1		PPB	28 Nov 2017 16:00:54	U	SMPL
6603-06 - 1	112217-JETT-F-D-S	-67	0.0153	-245.1		PPB	28 Nov 2017 16:03:02	U	SMPL
6603-07 - 1	112217-JETT-F-U-B	-51	0.0200	-245.1		PPB	28 Nov 2017 16:05:09	U	SMPL
6603-08 - 1	112217-JETT-F-U-S	-28	0.0268	-245.1		PPB	28 Nov 2017 16:07:16	U	SMPL
6604-01 - 1	112417-TIDO-F-D-B	-62	0.0168	-245.1		PPB	28 Nov 2017 16:09:24	U	SMPL
6604-02 - 1	112417-TIDO-F-D-S	-36	0.0245	-245.1		PPB	28 Nov 2017 16:11:32	U	SMPL
6604-03 - 1	112417-TIDO-F-U-B	-51	0.0200	-245.1		PPB	28 Nov 2017 16:13:40	U	SMPL
6604-04 - 1	112417-TIDO-F-U-S	-57	0.0183	-245.1		PPB	28 Nov 2017 16:15:48	U	SMPL
6604-04DUP - 1	112417-TIDO-F-U-SDUP	-44	0.0221	-245.1		PPB	28 Nov 2017 16:17:57	U	SMPL
6604-04MS - 1	112417-TIDO-F-U-SMS	14232	4.2284	-245.1		PPB	28 Nov 2017 16:20:04	U	SMPL
CCV05 - 1	CCV05	16583	4.9211	-245.1		PPB	28 Nov 2017 16:22:11	U	SMPL
CCB05 - 1	CCB05	-184	-0.0191	-245.1		PPB	28 Nov 2017 16:24:18	U	SMPL
6604-04MSD - 1	112417-TIDO-F-U-SMSD	14874	4.4176	-245.1		PPB	28 Nov 2017 16:26:26	U	SMPL
6584-08LX5 - 1		-158	-0.0115	-245.1		PPB	28 Nov 2017 16:28:33	U	SMPL
6584-08A - 1		15349	4.5575	-245.1		PPB	28 Nov 2017 16:30:40	U	SMPL
6604-04LX5 - 1		-149	-0.0088	-245.1		PPB	28 Nov 2017 16:32:47	U	SMPL
6604-04A - 1		15415	4.5770	-245.1		PPB	28 Nov 2017 16:34:54	U	SMPL
CCV06 - 1	CCV06	16556	4.9132	-245.1		PPB	28 Nov 2017 16:37:02	U	SMPL
CCB06 - 1	CCB06	-176	-0.0168	-245.1		PPB	28 Nov 2017 16:39:10	U	SMPL