


ISM02.4

LB91824
INSTRUMENT ID : CV2

Linear

 μ Abs.:

41207



A= 0.0000e+000

B= 2.4321e-004

C= -1.0964e-002

Rho= 0.9999530

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.017	-0.017	-26	0.000	-26					-
0.2	0.200	0.190	-0.010	826	0.0 %	826					-5
2.5	2.500	2.519	0.019	10403	0.0 %	10403					1
5.0	5.000	5.057	0.057	20838	0.0 %	20838					1
7.5	7.500	7.440	-0.060	30638	0.0 %	30638					-1
10.0	10.000	10.011	0.011	41207	0.0 %	41207					0

NB
11/28/18

LB91824
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	-26	-		0 ISM02.4	PPB	11/28/2017 16:51	Std
	0.2 S0.2	826	-		0.2 ISM02.4	PPB	11/28/2017 16:54	Std
	2.5 S2.5	10403	-		2.5 ISM02.4	PPB	11/28/2017 16:56	Std
	5 S5	20838	-		5 ISM02.4	PPB	11/28/2017 16:58	Std
	7.5 S7.5	30638	-		7.5 ISM02.4	PPB	11/28/2017 17:00	Std
	10 S10	41207	-		10 ISM02.4	PPB	11/28/2017 17:03	Std
ICV	ICV	15906	3.8575 -		ISM02.4	PPB	11/28/2017 17:06	SMPL
ICB	ICB	-172	-0.0528 -		ISM02.4	PPB	11/28/2017 17:08	SMPL
CCV007	CCV007	20717	5.0276 -		ISM02.4	PPB	11/28/2017 17:10	SMPL
CCB007	CCB007	-189	-0.0569 -		ISM02.4	PPB	11/28/2017 17:12	SMPL
PB104556BL	PBW003	-58	-0.0251 -		ISM02.4	PPB	11/28/2017 17:15	SMPL
I6510-01	MBE2W5	378	0.081 -		ISM02.4	PPB	11/28/2017 17:17	SMPL
I6510-02	MBE2Z1	-73	-0.0287 -		ISM02.4	PPB	11/28/2017 17:19	SMPL
I6510-03	MBE2Z2	44	-0.0003 -		ISM02.4	PPB	11/28/2017 17:21	SMPL
I6510-04	MBE2Z3	1145	0.2675 -		ISM02.4	PPB	11/28/2017 17:24	SMPL
I6510-05	MBE2Z4	-85	-0.0316 -		ISM02.4	PPB	11/28/2017 17:26	SMPL
I6510-06	MBE2Z5	-75	-0.0292 -		ISM02.4	PPB	11/28/2017 17:28	SMPL
I6510-07	MBE2Z6	-66	-0.027 -		ISM02.4	PPB	11/28/2017 17:31	SMPL
I6510-08	MBE2Z8	-49	-0.0229 -		ISM02.4	PPB	11/28/2017 17:33	SMPL
I6510-09	MBE2Z8D	-58	-0.0251 -		ISM02.4	PPB	11/28/2017 17:35	SMPL
I6510-10	MBE2Z8S	4098	0.9857 -		ISM02.4	PPB	11/28/2017 17:37	SMPL
I6510-11	MBE2Z9	-88	-0.0324 -		ISM02.4	PPB	11/28/2017 17:40	SMPL
I6510-12	MBE300	-44	-0.0217 -		ISM02.4	PPB	11/28/2017 17:42	SMPL
I6510-13	MBE301	-63	-0.0263 -		ISM02.4	PPB	11/28/2017 17:44	SMPL
I6510-14	MBE303	-25	-0.017 -		ISM02.4	PPB	11/28/2017 17:47	SMPL
I6510-15	MBE304	-44	-0.0217 -		ISM02.4	PPB	11/28/2017 17:49	SMPL
I6510-16	MBE305	-42	-0.0212 -		ISM02.4	PPB	11/28/2017 17:51	SMPL
I6552-02	MA41R2	6838	1.6521 -		ISM02.4	PPB	11/28/2017 17:53	SMPL
CCV008	CCV008	20370	4.9432 -		ISM02.4	PPB	11/28/2017 17:56	SMPL
CCB008	CCB008	-188	-0.0567 -		ISM02.4	PPB	11/28/2017 17:58	SMPL