



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91825

Temp.: 24.3 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/28/2017	17:00
2	CAL2	1	Water			7.01	11/28/2017	17:01
3	CAL3	1	Water			10.03	11/28/2017	17:02
4	ICV	1	Water			6.99	11/28/2017	17:05
5	CCV1	1	Water			2.01	11/28/2017	17:06
6	I6302-08	1	Solid	20.02	20	7.72	11/28/2017	17:08
7	I6302-10	1	Solid	20.04	20	8.96	11/28/2017	17:10
8	I6302-12	1	Solid	20.01	20	8.45	11/28/2017	17:11
9	I6304-23	1	Solid	20.02	20	8.10	11/28/2017	17:12
10	I6304-25	1	Solid	20.05	20	8.15	11/28/2017	17:14
11	I6304-27	1	Solid	20.03	20	8.82	11/28/2017	17:15
12	I6304-29	1	Solid	20.02	20	8.59	11/28/2017	17:16
13	I6304-31	1	Solid	20.03	20	8.15	11/28/2017	17:20
14	I6306-04	1	Solid	20.04	20	6.63	11/28/2017	17:21
15	I6306-06	1	Solid	20.02	20	6.51	11/28/2017	17:22
16	CCV2	1	Water			12.01	11/28/2017	17:25
17	I6612-03	1	Solid	20.04	20	6.74	11/28/2017	17:26
18	I6612-03DUP	1	Solid	20.05	20	6.75	11/28/2017	17:29
19	CCV3	1	Water			12.02	11/28/2017	17:30

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6612

WorkList ID : 105913

Date : 11/28/2017 2:13:26 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/28/2017	Solid	I6302-08	Corrosivity	Cool 4 deg C	PSEG05	M52	HR-06-112817-A	11/28/2017	9045C
11/28/2017	Solid	I6302-10	Corrosivity	Cool 4 deg C	PSEG05	M52	HR-06-112817-B	11/28/2017	9045C
11/28/2017	Solid	I6302-12	Corrosivity	Cool 4 deg C	PSEG05	M52	HR-06-112817-C	11/28/2017	9045C
11/28/2017	Solid	I6304-23	Corrosivity	Cool 4 deg C	PSEG05	M52	CL-02-112817-A	11/28/2017	9045C
11/28/2017	Solid	I6304-25	Corrosivity	Cool 4 deg C	PSEG05	M52	CL-02-112817-B	11/28/2017	9045C
11/28/2017	Solid	I6304-27	Corrosivity	Cool 4 deg C	PSEG05	M52	CL-02-112817-C	11/28/2017	9045C
11/28/2017	Solid	I6304-29	Corrosivity	Cool 4 deg C	PSEG05	M52	CL-02-112817-D	11/28/2017	9045C
11/28/2017	Solid	I6304-31	Corrosivity	Cool 4 deg C	PSEG05	M52	CL-02-112817-E	11/28/2017	9045C
11/28/2017	Solid	I6306-04	Corrosivity	Cool 4 deg C	PSEG05	M52	NB-08-112817-A	11/28/2017	9045C
11/28/2017	Solid	I6306-06	Corrosivity	Cool 4 deg C	PSEG05	M52	NB-08-112817-B	11/28/2017	9045C
11/28/2017	Solid	I6612-03	Corrosivity	Cool 4 deg C	PSEG01	M11	RT-3460	11/28/2017	9045C

Date/Time

Received by:

Relinquished by:

Date/Time

Received by:

Relinquished by: