

ISM02.4


LB 91826

INSTRUMENT ID: CV1

Linear

 μ Abs.:

45024



A= 0.0000e+000

B= 2.2210e-004

C= -2.1999e-002

Rho= 0.9999488

Accept=Accepted

Handwritten notes:
 slope
 y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/011
0.00	0.000	-0.026	-0.026	-18	0.000	-18					-
0.2	0.200	0.192	-0.008	963	0.0 %	963					-4
2.5	2.500	2.563	0.063	11639	0.0 %	11639					3
5.0	5.000	4.957	-0.043	22417	0.0 %	22417					-1
7.5	7.500	7.536	0.036	34030	0.0 %	34030					0
10.0	10.000	9.978	-0.022	45024	0.0 %	45024					0

Handwritten notes:
 MSB
 11/28/18

LB91826
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	-18	-	0	ISM02.4	PPB	11/28/2017 18:23	Std
	0.2 S0.2	963	-	0.2	ISM02.4	PPB	11/28/2017 18:26	Std
	2.5 S2.5	11639	-	2.5	ISM02.4	PPB	11/28/2017 18:28	Std
	5 S5	22417	-	5	ISM02.4	PPB	11/28/2017 18:30	Std
	7.5 S7.5	34030	-	7.5	ISM02.4	PPB	11/28/2017 18:32	Std
	10 S10	45024	-	10	ISM02.4	PPB	11/28/2017 18:35	Std
ICV	ICV	17692	3.9075 -		ISM02.4	PPB	11/28/2017 18:37	SMPL
ICB	ICB	-185	-0.0631 -		ISM02.4	PPB	11/28/2017 18:40	SMPL
CCV009	CCV009	22438	4.9616 -		ISM02.4	PPB	11/28/2017 18:42	SMPL
CCB009	CCB009	-176	-0.0611 -		ISM02.4	PPB	11/28/2017 18:44	SMPL
PB104557BL	PBS004	-27	-0.028 -		ISM02.4	PPB	11/28/2017 18:46	SMPL
I6552-01	MA41R1	19311	4.267 -		ISM02.4	PPB	11/28/2017 18:49	SMPL
CCV010	CCV010	22222	4.9136 -		ISM02.4	PPB	11/28/2017 18:51	SMPL
CCB010	CCB010	-185	-0.0631 -		ISM02.4	PPB	11/28/2017 18:53	SMPL