

LB91832

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

11/28/2017 16:24

Test: ~~Total~~ ^{Reactive} CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	98.967	0.0	0.096	
ICB1	-0.484	0.0	0.004	
CCV1	251.115	0.0	0.238	
CCB1	-0.199	0.0	0.004	
PB104535BL	-0.527	0.0	0.004	
I6299-08	-0.851	0.0	0.003	
I6299-08DUP	-0.891	0.0	0.003	
I6582-02	-1.109	0.0	0.003	
I6594-02	-1.048	0.0	0.003	
I6602-03	-1.025	0.0	0.003	
PB104536BL	-0.719	0.0	0.004	
I6583-02	-1.033	0.0	0.003	
I6583-02DUP	-0.820	0.0	0.003	
CCV2	254.516	0.0	0.241	
CCB2	-0.480	0.0	0.004	

N	15
Mean	39.694
SD	90.2397
CV%	227.34

AS

Aquakem v. 7.2AQ1

Results from time period:

Tue Nov 28 16:15:19 2017

Tue Nov 28 16:21:37 2017

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.8885	µg/l	11/28/2017 13:59:09	
5.OPPBCN	A	Reactive CN	P	4.0375	µg/l	11/28/2017 13:59:10	
10PPBCN	A	Reactive CN	P	8.9447	µg/l	11/28/2017 13:59:11	
50PPBCN	A	Reactive CN	P	50.1105	µg/l	11/28/2017 13:59:12	
100PPBCN	A	Reactive CN	P	101.4508	µg/l	11/28/2017 13:59:13	
250PPBCN	A	Reactive CN	P	253.2309	µg/l	11/28/2017 13:59:14	
500PPBCN	A	Reactive CN	P	498.1141	µg/l	11/28/2017 13:59:15	
ICV1	S	Reactive CN	P	98.9669	µg/l	11/28/2017 16:15:20	
ICB1	S	Reactive CN	P	-0.4839	µg/l	11/28/2017 16:15:21	
CCV1	S	Reactive CN	P	251.1146	µg/l	11/28/2017 16:15:24	
CCB1	S	Reactive CN	P	-0.1987	µg/l	11/28/2017 16:15:25	
PB104535BL	S	Reactive CN	P	-0.5265	µg/l	11/28/2017 16:15:27	
I6299-08	S	Reactive CN	P	-0.8512	µg/l	11/28/2017 16:15:28	
I6299-08DUP	S	Reactive CN	P	-0.8912	µg/l	11/28/2017 16:15:29	
I6582-02	S	Reactive CN	P	-1.1085	µg/l	11/28/2017 16:21:30	
I6594-02	S	Reactive CN	P	-1.0476	µg/l	11/28/2017 16:21:31	
I6602-03	S	Reactive CN	P	-1.0246	µg/l	11/28/2017 16:21:32	
PB104536BL	S	Reactive CN	P	-0.7192	µg/l	11/28/2017 16:21:33	
I6583-02	S	Reactive CN	P	-1.0335	µg/l	11/28/2017 16:21:34	
I6583-02DUP	S	Reactive CN	P	-0.8205	µg/l	11/28/2017 16:21:35	
CCV2	S	Reactive CN	P	254.516	µg/l	11/28/2017 16:21:36	
CCB2	S	Reactive CN	P	-0.4797	µg/l	11/28/2017 16:21:37	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

11/28/2017 14:03


Test Total CN

Accepted 11/28/2017 14:03

Factor 1076
Bias 0.004

Coeff. of det. 0.999908

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.8885	0.0000	
2	5.0PPBCN	0.008	4.0375	5.0000	
3	10PPBCN	0.013	8.9447	10.0000	
4	50PPBCN	0.051	50.1105	50.0000	
5	100PPBCN	0.098	101.4508	100.0000	
6	250PPBCN	0.239	253.2309	250.0000	
7	500PPBCN	0.467	498.1141	500.0000	