


7471A

LB91835
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45739



A= 0.0000e+000

B= 2.1986e-004

C= -1.3423e-002

Rho= 0.9999080

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.018	-0.018	-19	0.000	-19				
0.2	0.200	0.184	-0.016	899	0.0 %	899				
2.5	2.500	2.582	0.082	11803	0.0 %	11803				
5.0	5.000	4.985	-0.015	22732	0.0 %	22732				
7.5	7.500	7.424	-0.076	33829	0.0 %	33829	0			
10.0	10.000	10.043	0.043	45739	0.0 %	45739				

LB91835 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 29 Nov 2017 11:50:39

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-19	-	0.00007471A	PPB	29 Nov 2017 11:52:42	S	Std
0.2 - 1	50.2	899	-	0.20007471A	PPB	29 Nov 2017 11:55:00	S	Std
2.5 - 1	52.5	11803	-	2.50007471A	PPB	29 Nov 2017 11:57:17	S	Std
5.0 - 1	55	22732	-	5.00007471A	PPB	29 Nov 2017 11:59:35	S	Std
7.5 - 1	57.5	33829	-	7.50007471A	PPB	29 Nov 2017 12:01:52	S	Std
10.0 - 1	511	45739	-	10.00007471A	PPB	29 Nov 2017 12:06:41	S	Std
ICV83 - 1	ICV83	17197	3.7676	-7471A	PPB	29 Nov 2017 12:09:18	U	SMPL
ICB83 - 1	ICB83	-161	-0.0488	-7471A	PPB	29 Nov 2017 12:11:35	U	SMPL
CCV14 - 1	CCV14	21970	4.8170	-7471A	PPB	29 Nov 2017 12:13:51	U	SMPL
CCB14 - 1	CCB14	-160	-0.0486	-7471A	PPB	29 Nov 2017 12:16:07	U	SMPL
CRA - 1	CRA	904	0.1853	-7471A	PPB	29 Nov 2017 12:18:23	U	SMPL
HighStd - 1	HighStd	46112	10.1249	-7471A	PPB	29 Nov 2017 12:20:39	U	SMPL
ChkStd - 1	ChkStd	30549	6.7032	-7471A	PPB	29 Nov 2017 12:22:55	U	SMPL
PB104585BL - 1	PBS	-264	-0.0715	-7471A	PPB	29 Nov 2017 12:25:13	U	SMPL
PB104585BS - 1	LCSS	15034	3.2920	-7471A	PPB	29 Nov 2017 12:27:31	U	SMPL
I6610-02 - 1	SB-3(0-2)-20171127	7641	1.6666	-7471A	PPB	29 Nov 2017 12:29:49	U	SMPL
I6610-03 - 1	SB-3(10-12)-20171127	7189	1.5672	-7471A	PPB	29 Nov 2017 12:32:06	U	SMPL
I6610-04 - 1	SB-5(0-2)-20171127	8677	1.8943	-7471A	PPB	29 Nov 2017 12:34:24	U	SMPL
I6610-05 - 1	SB-5(10-12)-20171127	383	0.0708	-7471A	PPB	29 Nov 2017 12:36:41	U	SMPL
I6610-06 - 1	SB-2(0-2)-20171127	269	0.0457	-7471A	PPB	29 Nov 2017 12:38:57	U	SMPL
CCV15 - 1	CCV15	21777	4.7745	-7471A	PPB	29 Nov 2017 12:41:13	U	SMPL
CCB15 - 1	CCB15	-152	-0.0468	-7471A	PPB	29 Nov 2017 12:43:29	U	SMPL
I6610-07 - 1	SB-2(10-12)-20171127	512	0.0991	-7471A	PPB	29 Nov 2017 12:45:45	U	SMPL
I6610-08 - 1	SB-8(0-2)-20171127	12009	2.6269	-7471A	PPB	29 Nov 2017 12:48:01	U	SMPL
I6610-09 - 1	SB-8(10-12)-20171127	9	-0.0114	-7471A	PPB	29 Nov 2017 12:50:17	U	SMPL
I6610-10 - 1	SB-7(0-2)-20171127	5302	1.1523	-7471A	PPB	29 Nov 2017 12:52:33	U	SMPL
I6610-11 - 1	SB-7(4-6)-20171127	1472	0.3102	-7471A	PPB	29 Nov 2017 12:54:50	U	SMPL
I6610-12 - 1	SB-9(0-2)-20171127	32798	7.1977	-7471A	PPB	29 Nov 2017 12:57:07	U	SMPL
I6610-13 - 1	SB-9(8-10)-20171127	763	0.1543	-7471A	PPB	29 Nov 2017 12:59:25	U	SMPL
I6610-14 - 1	SB-10(0-2)-20171127	42911	9.4211	-7471A	PPB	29 Nov 2017 13:01:43	U	SMPL
I6610-15 - 1	SB-X(0-2)-20171127	22324	4.8948	-7471A	PPB	29 Nov 2017 13:04:01	U	SMPL
I6610-16 - 1	SB-10(8-10)-20171127	1041	0.2155	-7471A	PPB	29 Nov 2017 13:06:19	U	SMPL
CCV16 - 1	CCV16	21807	4.7811	-7471A	PPB	29 Nov 2017 13:08:36	U	SMPL
CCB16 - 1	CCB16	-168	-0.0504	-7471A	PPB	29 Nov 2017 13:10:53	U	SMPL
I6610-17 - 1	SB-12(0-2)-20171127	25614	5.6182	-7471A	PPB	29 Nov 2017 13:18:03	U	SMPL
I6610-17DUP - 1	SB-12(0-2)-20171127DUP	25167	5.5199	-7471A	PPB	29 Nov 2017 13:20:19	U	SMPL
I6610-18 - 1	I6610-17MS	41135	9.0307	-7471A	PPB	29 Nov 2017 13:28:53	U	SMPL
I6610-19 - 1	I6610-17MSD	41651	9.1441	-7471A	PPB	29 Nov 2017 13:31:09	U	SMPL
I6610-20 - 1	SB-12(10-12)-20171127	10374	2.2674	-7471A	PPB	29 Nov 2017 13:33:28	U	SMPL
I6612-01 - 1	RT-3460	1495	0.3153	-7471A	PPB	29 Nov 2017 13:35:46	U	SMPL
PB104586BL - 1	PBS	-49	-0.0242	-7471A	PPB	29 Nov 2017 13:38:04	U	SMPL
PB104586BS - 1	LCSS	15171	3.3221	-7471A	PPB	29 Nov 2017 13:40:21	U	SMPL
6302-07 - 1	HR-06-112817-A	5765	1.2541	-7471A	PPB	29 Nov 2017 13:42:40	U	SMPL
6302-09 - 1	HR-06-112817-B	40843	8.9665	-7471A	PPB	29 Nov 2017 13:44:56	U	SMPL
CCV17 - 1	CCV17	21810	4.7818	-7471A	PPB	29 Nov 2017 13:47:12	U	SMPL
CCB17 - 1	CCB17	-179	-0.0528	-7471A	PPB	29 Nov 2017 13:49:30	U	SMPL
6302-11 - 1	HR-06-112817-C	3879	0.8394	-7471A	PPB	29 Nov 2017 13:51:47	U	SMPL
6302-11DUP - 1	HR-06-112817-CDUP	3513	0.7590	-7471A	PPB	29 Nov 2017 13:54:04	U	SMPL
6302-11MS - 1	HR-06-112817-CMS	19404	4.2528	-7471A	PPB	29 Nov 2017 13:56:20	U	SMPL
6302-11MSD - 1	HR-06-112817-CMSD	19637	4.3040	-7471A	PPB	29 Nov 2017 13:58:36	U	SMPL
6304-22 - 1	CL-02-112817-A	4771	1.0355	-7471A	PPB	29 Nov 2017 14:00:53	U	SMPL
6304-24 - 1	CL-02-112817-B	7222	1.5744	-7471A	PPB	29 Nov 2017 14:03:10	U	SMPL
6304-26 - 1	CL-02-112817-C	2689	0.5778	-7471A	PPB	29 Nov 2017 14:05:28	U	SMPL
6304-28 - 1	CL-02-112817-D	2606	0.5595	-7471A	PPB	29 Nov 2017 14:07:45	U	SMPL
6304-30 - 1	CL-02-112817-E	3505	0.7572	-7471A	PPB	29 Nov 2017 14:10:03	U	SMPL
6306-03 - 1	NB-08-112817-A	460	0.0877	-7471A	PPB	29 Nov 2017 14:12:19	U	SMPL
CCV18 - 1	CCV18	22064	4.8376	-7471A	PPB	29 Nov 2017 14:14:36	U	SMPL
CCB18 - 1	CCB18	-147	-0.0457	-7471A	PPB	29 Nov 2017 14:16:52	U	SMPL
6306-05 - 1	NB-08-112817-B	1255	0.2625	-7471A	PPB	29 Nov 2017 14:19:10	U	SMPL
6616-01 - 1	P7-LOT-SP1-C	3587	0.7752	-7471A	PPB	29 Nov 2017 14:21:26	U	SMPL
6610-17LX5 - 1		4566	0.9905	-7471A	PPB	29 Nov 2017 14:23:42	U	SMPL
6610-17A - 1		40671	8.9286	-7471A	PPB	29 Nov 2017 14:25:58	U	SMPL
6302-11LX5 - 1		441	0.0835	-7471A	PPB	29 Nov 2017 14:28:15	U	SMPL
6302-11A - 1		19669	4.3111	-7471A	PPB	29 Nov 2017 14:30:34	U	SMPL
CCV19 - 1	CCV19	22626	4.9612	-7471A	PPB	29 Nov 2017 14:32:51	U	SMPL
CCB19 - 1	CCB19	-199	-0.0572	-7471A	PPB	29 Nov 2017 14:35:08	U	SMPL