



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91847

Temp.: 24.2 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/29/2017	18:00
2	CAL2	1	Water			7.01	11/29/2017	18:01
3	CAL3	1	Water			10.02	11/29/2017	18:02
4	ICV	1	Water			6.99	11/29/2017	18:05
5	CCV1	1	Water			2.01	11/29/2017	18:06
6	I6289-10	1	Solid	20.02	20	7.53	11/29/2017	18:07
7	I6300-08	1	Solid	20.03	20	7.44	11/29/2017	18:08
8	I6615-03	1	Solid	20.01	20	9.55	11/29/2017	18:10
9	I6616-03	1	Solid	20.04	20	8.30	11/29/2017	18:11
10	I6639-02	1	Solid	20.01	20	8.57	11/29/2017	18:12
11	I6641-02	1	Solid	20.03	20	8.30	11/29/2017	18:15
12	I6641-04	1	Solid	20.05	20	7.61	11/29/2017	18:18
13	I6641-06	1	Solid	20.03	20	8.30	11/29/2017	18:20
14	I6641-06DUP	1	Solid	20.02	20	8.31	11/29/2017	18:22
15	CCV2	1	Water			12.01	11/29/2017	18:23

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6616

WorkList ID : 105946

Date : 11/29/2017 5:01:10 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/29/2017	Solid	I6289-10	Corrosivity	Cool 4 deg C	PSEG05		SU-03-112917	11/29/2017	9045C
11/29/2017	Solid	I6300-08	Corrosivity	Cool 4 deg C	PSEG05		OR-02-112917	11/29/2017	9045C
11/28/2017	Solid	I6615-03	Corrosivity	Cool 4 deg C	WSP001	N32	PB-2-OW(0-10)	11/28/2017	9045C
11/28/2017	Solid	I6616-03	Corrosivity	Cool 4 deg C	SKAN06	N33	P7-LOT-SP1-C	11/28/2017	9045C
11/29/2017	Solid	I6639-02	Corrosivity	Cool 4 deg C	PSEG01	N42	MH-SP	11/29/2017	9045C
11/29/2017	Solid	I6641-02	Corrosivity	Cool 4 deg C	PSEG01	N41	TP-1	11/29/2017	9045C
11/29/2017	Solid	I6641-04	Corrosivity	Cool 4 deg C	PSEG01	N41	TP-2	11/29/2017	9045C
11/29/2017	Solid	I6641-06	Corrosivity	Cool 4 deg C	PSEG01	N41	TP-3	11/29/2017	9045C

Date/Time 11-29-17 5:45 PM

Received by: J

Relinquished by: _____

Date/Time 11-29-17 7:00 PM

Received by: CP

Relinquished by: Jg