

7470A


LB91851

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35080



A= 0.0000e+000

B= 2.8650e-004

C= -7.5739e-003

Rho= 0.9999494

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.008	-0.008	0	0.000	0				
0.2	0.200	0.199	-0.001	721	0.0 %	721				
2.5	2.500	2.549	0.049	8925	0.0 %	8925				
5.0	5.000	4.973	-0.027	17384	0.0 %	17384				
7.5	7.500	7.443	-0.057	26007	0.0 %	26007				
10.0	10.000	10.043	0.043	35080	0.0 %	35080				

LB91851 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Nov 2017 09:41:32

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	0	-	0.00007470A	PPB	30 Nov 2017 10:41:54	S	Std
0.2 - 1	50.2	721	-	0.20007470A	PPB	30 Nov 2017 10:44:03	S	Std
2.5 - 1	52.5	8925	-	2.50007470A	PPB	30 Nov 2017 10:46:11	S	Std
5.0 - 1	55	17384	-	5.00007470A	PPB	30 Nov 2017 10:48:19	S	Std
7.5 - 1	57.5	26007	-	7.50007470A	PPB	30 Nov 2017 10:50:27	S	Std
10.0 - 1	510	35080	-	10.00007470A	PPB	30 Nov 2017 10:52:35	S	Std
ICV84 - 1	ICV84	13671	3.9092	-7470A	PPB	30 Nov 2017 10:55:06	U	SMPL
ICB84 - 1	ICB84	-137	-0.0468	-7470A	PPB	30 Nov 2017 10:57:12	U	SMPL
CCV20 - 1	CCV20	17877	5.1142	-7470A	PPB	30 Nov 2017 10:59:20	U	SMPL
CCB20 - 1	CCB20	-136	-0.0465	-7470A	PPB	30 Nov 2017 11:01:26	U	SMPL
CRA - 1	CRA	745	0.2059	-7470A	PPB	30 Nov 2017 11:03:34	U	SMPL
HighStd - 1	HighStd	34808	9.9649	-7470A	PPB	30 Nov 2017 11:05:41	U	SMPL
ChkStd - 1	ChkStd	24472	7.0036	-7470A	PPB	30 Nov 2017 11:07:48	U	SMPL
PB104603BL - 1	PBW	-192	-0.0626	-7470A	PPB	30 Nov 2017 11:09:58	U	SMPL
PB104603BS - 1	LCSW	14012	4.0069	-7470A	PPB	30 Nov 2017 11:12:07	U	SMPL
PB104603TB - 1	PB104579TB	-96	-0.0351	-7470A	PPB	30 Nov 2017 11:14:14	U	SMPL
I6302-08 - 1	HR-06-112817-A	-17	-0.0124	-7470A	PPB	30 Nov 2017 11:16:22	U	SMPL
I6302-10 - 1	HR-06-112817-B	-11	-0.0107	-7470A	PPB	30 Nov 2017 11:18:31	U	SMPL
I6302-12 - 1	HR-06-112817-C	-1	-0.0079	-7470A	PPB	30 Nov 2017 11:20:39	U	SMPL
I6304-23 - 1	CL-02-112817-A	-15	-0.0119	-7470A	PPB	30 Nov 2017 11:22:46	U	SMPL
CCV21 - 1	CCV21	17752	5.0784	-7470A	PPB	30 Nov 2017 11:24:53	U	SMPL
CCB21 - 1	CCB21	-143	-0.0485	-7470A	PPB	30 Nov 2017 11:26:59	U	SMPL
I6304-25 - 1	CL-02-112817-B	-40	-0.0190	-7470A	PPB	30 Nov 2017 11:29:07	U	SMPL
I6304-27 - 1	CL-02-112817-C	-26	-0.0150	-7470A	PPB	30 Nov 2017 11:31:13	U	SMPL
I6304-29 - 1	CL-02-112817-D	-28	-0.0156	-7470A	PPB	30 Nov 2017 11:33:20	U	SMPL
I6304-31 - 1	CL-02-112817-E	-13	-0.0113	-7470A	PPB	30 Nov 2017 11:35:27	U	SMPL
I6306-04 - 1	NB-08-112817-A	-40	-0.0190	-7470A	PPB	30 Nov 2017 11:37:35	U	SMPL
I6306-06 - 1	NB-08-112817-B	18	-0.0024	-7470A	PPB	30 Nov 2017 11:39:42	U	SMPL
I6612-03 - 1	RT-3460	9	-0.0050	-7470A	PPB	30 Nov 2017 11:41:50	U	SMPL
I6612-03DUP - 1	RT-3460DUP	11	-0.0044	-7470A	PPB	30 Nov 2017 11:43:58	U	SMPL
I6612-03MS - 1	RT-3460MS	12916	3.6928	-7470A	PPB	30 Nov 2017 11:46:07	U	SMPL
I6612-03MSD - 1	RT-3460MSD	12896	3.6871	-7470A	PPB	30 Nov 2017 11:48:14	U	SMPL
CCV22 - 1	CCV22	17715	5.0678	-7470A	PPB	30 Nov 2017 11:50:21	U	SMPL
CCB22 - 1	CCB22	-153	-0.0514	-7470A	PPB	30 Nov 2017 11:52:28	U	SMPL
I6612-03LX5 - 1		-12	-0.0110	-7470A	PPB	30 Nov 2017 11:54:37	U	SMPL
I6612-03A - 1		12387	3.5413	-7470A	PPB	30 Nov 2017 11:56:43	U	SMPL
CCV23 - 1	CCV23	17589	5.0317	-7470A	PPB	30 Nov 2017 11:58:51	U	SMPL
CCB23 - 1	CCB23	-149	-0.0503	-7470A	PPB	30 Nov 2017 12:00:58	U	SMPL