

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB91854

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP56414
calibration std. phosphate 1 ppm	WP59986
calibration std. phosphate 0.5 ppm	WP59985
calibration std. phosphate 0.3 ppm	WP59984
calibration std. 0 ppm	WP59981
calibration std. phosphate 0.1 ppm	WP59983
calibration std. phosphate 0.05 ppm	WP59982
Combined reagent	WP59990
11 N sulfuric acid	WP57045
Phenolphthalein indicator	WP57038
Sodium hydroxide, 1N	WP57036
phosphate ICV std.	WP59988
phosphate CCV std.	WP59987

Intercept: -0.0009Slope: 0.6504Regression: 0.999829

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	11/30/2017	11:03
2	CAL2	0.05	1	50	50	0.033	11/30/2017	11:03
3	CAL3	0.10	1	50	50	0.065	11/30/2017	11:04
4	CAL4	0.30	1	50	50	0.186	11/30/2017	11:04
5	CAL5	0.50	1	50	50	0.330	11/30/2017	11:06
6	CAL6	1.00	1	50	50	0.649	11/30/2017	11:06



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB91854

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.321	0.495	11/30/2017	11:07
2	ICB		1	50	50	0.000	0.001	11/30/2017	11:07
3	CCV1	0.50	1	50	50	0.329	0.507	11/30/2017	11:08
4	CCB1		1	50	50	0.000	0.001	11/30/2017	11:08
5	PB104628BL		1	50	50	0.000	0.001	11/30/2017	11:08
6	PB104628BS	0.50	1	50	50	0.339	0.523	11/30/2017	11:10
7	I6597-01		1	50	50	0.024	0.038	11/30/2017	11:10
8	I6597-01DUP		1	50	50	0.024	0.038	11/30/2017	11:11
9	I6597-01MS	0.50	1	50	50	0.339	0.523	11/30/2017	11:11
10	I6597-01MSD	0.50	1	50	50	0.376	0.579	11/30/2017	11:12
11	I6597-02		1	50	50	0.016	0.026	11/30/2017	11:12
12	CCV2	0.50	1	50	50	0.327	0.504	11/30/2017	11:13
13	CCB2		1	50	50	0.000	0.001	11/30/2017	11:13