

7471A


LB 91859

INSTRUMENT ID: CV4

Linear

 μ Abs.:

43659



A= 0.0000e+000

B= 2.3047e-004

C= -1.8248e-002

Rho= 0.9999277

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.004	-0.004	60	0.000	60				
0.2	0.200	0.182	-0.018	869	0.0 %	869				
2.5	2.500	2.565	0.065	11210	0.0 %	11210				
5.0	5.000	4.984	-0.016	21703	0.0 %	21703				
7.5	7.500	7.429	-0.071	32314	0.0 %	32314	0			
10.0	10.000	10.044	0.044	43659	0.0 %	43659				

LB91859 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 30 Nov 2017 14:08:37

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	60	-	0.00007471A	-7471A	PPB	30 Nov 2017 14:11:39	S	Std
0.2 - 1	50.2	869	-	0.20007471A	-7471A	PPB	30 Nov 2017 14:13:57	S	Std
2.5 - 1	52.5	11210	-	2.50007471A	-7471A	PPB	30 Nov 2017 14:16:14	S	Std
5.0 - 1	55	21703	-	5.00007471A	-7471A	PPB	30 Nov 2017 14:21:09	S	Std
7.5 - 1	57.5	32314	-	7.50007471A	-7471A	PPB	30 Nov 2017 14:23:26	S	Std
10.0 - 1	510	43659	-	10.00007471A	-7471A	PPB	30 Nov 2017 14:25:43	S	Std
ICV85 - 1	ICV85	17752	4.0731	-7471A	-7471A	PPB	30 Nov 2017 14:28:29	U	SMPL
ICB85 - 1	ICB85	-83	-0.0374	-7471A	-7471A	PPB	30 Nov 2017 14:30:45	U	SMPL
CCV24 - 1	CCV24	20504	4.7074	-7471A	-7471A	PPB	30 Nov 2017 14:33:01	U	SMPL
CCB24 - 1	CCB24	-79	-0.0365	-7471A	-7471A	PPB	30 Nov 2017 14:35:16	U	SMPL
CRA - 1	CRA	870	0.1823	-7471A	-7471A	PPB	30 Nov 2017 14:37:33	U	SMPL
HighStd - 1	HighStd	43617	10.0343	-7471A	-7471A	PPB	30 Nov 2017 14:39:49	U	SMPL
ChkStd - 1	ChkStd	30850	7.0919	-7471A	-7471A	PPB	30 Nov 2017 14:42:05	U	SMPL
PB104609BL - 1	PBS	-38	-0.0270	-7471A	-7471A	PPB	30 Nov 2017 14:44:23	U	SMPL
PB104609BS - 1	LCSS	15276	3.5025	-7471A	-7471A	PPB	30 Nov 2017 14:46:41	U	SMPL
I6615-01 - 1	PB-2-OW(0-10)	4114	0.9299	-7471A	-7471A	PPB	30 Nov 2017 14:48:59	U	SMPL
I6618-01 - 1	ENV-106-4(0-5)	3783	0.8536	-7471A	-7471A	PPB	30 Nov 2017 14:51:16	U	SMPL
I6618-02 - 1	ENV-106-4(5-10)	485	0.0935	-7471A	-7471A	PPB	30 Nov 2017 14:53:33	U	SMPL
I6618-03 - 1	ENV-106-4(10-15)	415	0.0774	-7471A	-7471A	PPB	30 Nov 2017 14:55:52	U	SMPL
I6618-04 - 1	ENV-106-6(0-5)	11567	2.6476	-7471A	-7471A	PPB	30 Nov 2017 14:58:08	U	SMPL
CCV25 - 1	CCV25	21120	4.8494	-7471A	-7471A	PPB	30 Nov 2017 15:00:24	U	SMPL
CCB25 - 1	CCB25	-80	-0.0367	-7471A	-7471A	PPB	30 Nov 2017 15:02:40	U	SMPL
I6618-05 - 1	ENV-106-1(0-5)	19273	4.4237	-7471A	-7471A	PPB	30 Nov 2017 15:04:56	U	SMPL
I6619-01 - 1	EPE-120-4(10-12)	38	-0.0095	-7471A	-7471A	PPB	30 Nov 2017 15:07:12	U	SMPL
I6620-01 - 1	SB-1(0-2)-20171128	7734	1.7642	-7471A	-7471A	PPB	30 Nov 2017 15:09:29	U	SMPL
I6620-02 - 1	SB-1(6-8)-20171128	20911	4.8012	-7471A	-7471A	PPB	30 Nov 2017 15:11:45	U	SMPL
I6620-03 - 1	SB-4(0-2)-20171128	9616	2.1980	-7471A	-7471A	PPB	30 Nov 2017 15:14:01	U	SMPL
I6620-04 - 1	SB-4(10-12)-20171128	148	0.0159	-7471A	-7471A	PPB	30 Nov 2017 15:16:19	U	SMPL
I6620-05 - 1	SB-6(0-2)-20171128	9070	2.0722	-7471A	-7471A	PPB	30 Nov 2017 15:18:36	U	SMPL
I6620-06 - 1	SB-6(5-7)-20171128	120139	27.6707	-7471A	-7471A	PPB	30 Nov 2017 15:20:53	U	SMPL
I6630-01 - 1	P004-S001-0001-01	3287	0.7393	-7471A	-7471A	PPB	30 Nov 2017 15:23:11	U	SMPL
CCV26 - 1	CCV26	21191	4.8657	-7471A	-7471A	PPB	30 Nov 2017 15:27:47	U	SMPL
CCB26 - 1	CCB26	-86	-0.0381	-7471A	-7471A	PPB	30 Nov 2017 15:30:03	U	SMPL
I6630-02 - 1	I6630-01MS	19758	4.5355	-7471A	-7471A	PPB	30 Nov 2017 15:32:20	U	SMPL
I6630-03 - 1	I6630-01MSD	19381	4.4486	-7471A	-7471A	PPB	30 Nov 2017 15:34:37	U	SMPL
I6630-04 - 1	P004-S001-0001-02	3466	0.7806	-7471A	-7471A	PPB	30 Nov 2017 15:36:54	U	SMPL
I6630-05 - 1	P004-S002-0001-01	1650	0.3620	-7471A	-7471A	PPB	30 Nov 2017 15:39:10	U	SMPL
PB104610BL - 1	PBS	71	-0.0019	-7471A	-7471A	PPB	30 Nov 2017 15:41:27	U	SMPL
PB104610BS - 1	LCSS	15163	3.4764	-7471A	-7471A	PPB	30 Nov 2017 15:43:44	U	SMPL
6289-09 - 1	SU-03-112917	1658	0.3639	-7471A	-7471A	PPB	30 Nov 2017 15:46:01	U	SMPL
6300-07 - 1	OR-02-112917	3644	0.8216	-7471A	-7471A	PPB	30 Nov 2017 15:48:19	U	SMPL
6622-07 - 1	DBEW1A-2-4	1471	0.3208	-7471A	-7471A	PPB	30 Nov 2017 15:50:37	U	SMPL
6622-08 - 1	DBEW1B-4-9	1276	0.2758	-7471A	-7471A	PPB	30 Nov 2017 15:52:53	U	SMPL
CCV27 - 1	CCV27	21062	4.8360	-7471A	-7471A	PPB	30 Nov 2017 15:55:09	U	SMPL
CCB27 - 1	CCB27	-57	-0.0314	-7471A	-7471A	PPB	30 Nov 2017 15:57:26	U	SMPL
6622-09 - 1	DBNW1B-4-9	1508	0.3293	-7471A	-7471A	PPB	30 Nov 2017 15:59:43	U	SMPL
6622-10 - 1	DBNW1A-2-4	5041	1.1436	-7471A	-7471A	PPB	30 Nov 2017 16:01:59	U	SMPL
6622-11 - 1	DBWW1A-2-4	985	0.2088	-7471A	-7471A	PPB	30 Nov 2017 16:04:15	U	SMPL
6622-12 - 1	DBWW1B-4-9	1552	0.3394	-7471A	-7471A	PPB	30 Nov 2017 16:06:32	U	SMPL
6622-12DUP - 1	DBWW1B-4-9DUP	1391	0.3023	-7471A	-7471A	PPB	30 Nov 2017 16:08:49	U	SMPL
6622-12MS - 1	DBWW1B-4-9MS	17340	3.9782	-7471A	-7471A	PPB	30 Nov 2017 16:11:06	U	SMPL
6622-12MSD - 1	DBWW1B-4-9MSD	17316	3.9726	-7471A	-7471A	PPB	30 Nov 2017 16:13:23	U	SMPL
6639-03 - 1	342-VNJ216	2859	0.6407	-7471A	-7471A	PPB	30 Nov 2017 16:15:40	U	SMPL
6640-01 - 1	SB-11(0-2)-20171129	5087	1.1542	-7471A	-7471A	PPB	30 Nov 2017 16:17:58	U	SMPL
6640-02 - 1	SB-11(6-8)-20171129	7386	1.6840	-7471A	-7471A	PPB	30 Nov 2017 16:20:14	U	SMPL
CCV28 - 1	CCV28	21106	4.8461	-7471A	-7471A	PPB	30 Nov 2017 16:22:30	U	SMPL
CCB28 - 1	CCB28	-59	-0.0318	-7471A	-7471A	PPB	30 Nov 2017 16:24:46	U	SMPL
6641-01 - 1	TP-1	1047	0.2231	-7471A	-7471A	PPB	30 Nov 2017 16:27:03	U	SMPL
6641-03 - 1	TP-2	1671	0.3669	-7471A	-7471A	PPB	30 Nov 2017 16:29:19	U	SMPL
6641-05 - 1	TP-3	604	0.1210	-7471A	-7471A	PPB	30 Nov 2017 16:31:36	U	SMPL
6630-01LX5 - 1		841	0.1756	-7471A	-7471A	PPB	30 Nov 2017 16:33:52	U	SMPL
6630-01A - 1		18834	4.3225	-7471A	-7471A	PPB	30 Nov 2017 16:36:09	U	SMPL
6622-12LX5 - 1		259	0.0414	-7471A	-7471A	PPB	30 Nov 2017 16:38:26	U	SMPL
6622-12A - 1		16889	3.8742	-7471A	-7471A	PPB	30 Nov 2017 16:40:43	U	SMPL
6630-01DUP - 1	P004-S001-0001-01DUP	3849	0.8688	-7471A	-7471A	PPB	30 Nov 2017 16:43:00	U	SMPL
6620-06DLX5 - 1		23562	5.4122	-7471A	-7471A	PPB	30 Nov 2017 16:45:18	U	SMPL
CCV29 - 1	CCV29	21202	4.8683	-7471A	-7471A	PPB	30 Nov 2017 16:47:35	U	SMPL
CCB29 - 1	CCB29	-56	-0.0312	-7471A	-7471A	PPB	30 Nov 2017 16:49:51	U	SMPL