



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB91875

Temp.: 25.4 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/01/2017	09:00
2	CAL2	1	Water			7.00	12/01/2017	09:01
3	CAL3	1	Water			10.01	12/01/2017	09:02
4	ICV	1	Water			7.01	12/01/2017	09:04
5	CCV1	1	Water			2.02	12/01/2017	09:05
6	I6664-01	1	Solid	20.02	20	7.42	12/01/2017	09:09
7	I6664-01DUP	1	Solid	20.03	20	7.43	12/01/2017	09:11
8	I6664-02	1	Solid	20.05	20	4.95	12/01/2017	09:12
9	CCV2	1	Water			12.01	12/01/2017	09:13

2391875

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I6664

WorkList ID : 105999

Date : 12/1/2017 8:39:36 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/29/2017	Solid	I6664-01-13	pH	Cool 4 deg C	TULL02	M11	EME-S	11/29/2017	9045C
11/29/2017	Solid	I6664-02-13	pH	Cool 4 deg C	TULL02	M11	EME-C	11/29/2017	9045C

Date/Time 12-01-17 8:40 AM
Received by: JD
Relinquished by: CR

Date/Time 12-01-17 10:00 AM
Received by: CR
Relinquished by: JD