


7470A

LB91877  
INSTRUMENT ID: CVJ

Linear

$\mu$  Abs.:

35187



A= 0.0000e+000

B= 2.8484e-004

C= -2.0936e-002

Rho= 0.9999774

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.021	-0.021	1	0.000	1				
0.2	0.200	0.187	-0.013	730	0.0 %	730				
2.5	2.500	2.540	0.040	8991	0.0 %	8991				
5.0	5.000	5.023	0.023	17708	0.0 %	17708				
7.5	7.500	7.469	-0.031	26295	0.0 %	26295				
10.0	10.000	10.002	0.002	35187	0.0 %	35187				

# LB91877 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Dec 2017 10:27:10

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	1	-	0.0000	7470A	PPB	01 Dec 2017 10:35:31	S	Std
0.2 - 1	50.2	730	-	0.2000	7470A	PPB	01 Dec 2017 10:37:40	S	Std
2.5 - 1	52.5	8991	-	2.5000	7470A	PPB	01 Dec 2017 10:39:48	S	Std
5.0 - 1	55	17708	-	5.0000	7470A	PPB	01 Dec 2017 10:46:12	S	Std
7.5 - 1	57.5	26295	-	7.5000	7470A	PPB	01 Dec 2017 10:48:20	S	Std
10.0 - 1	610	35187	-	10.0000	7470A	PPB	01 Dec 2017 10:55:16	S	Std
ICV86 - 1	ICV86	13755	3.8970	-	7470A	PPB	01 Dec 2017 10:57:46	U	SMPL
ICB86 - 1	ICB86	-134	-0.0591	-	7470A	PPB	01 Dec 2017 10:59:53	U	SMPL
CCV30 - 1	CCV30	17861	5.0666	-	7470A	PPB	01 Dec 2017 11:01:59	U	SMPL
CCB30 - 1	CCB30	-125	-0.0565	-	7470A	PPB	01 Dec 2017 11:04:06	U	SMPL
CRA - 1	CRA	728	0.1864	-	7470A	PPB	01 Dec 2017 11:06:13	U	SMPL
HighStd - 1	HighStd	35487	10.0871	-	7470A	PPB	01 Dec 2017 11:08:20	U	SMPL
ChkStd - 1	ChkStd	25157	7.1448	-	7470A	PPB	01 Dec 2017 11:10:27	U	SMPL
PB104635BL - 1	PBW	-213	-0.0816	-	7470A	PPB	01 Dec 2017 11:12:36	U	SMPL
PB104635BS - 1	LCSW	13418	3.8010	-	7470A	PPB	01 Dec 2017 11:14:45	U	SMPL
PB104635TB - 1	PB104611TB	-129	-0.0577	-	7470A	PPB	01 Dec 2017 11:16:53	U	SMPL
I6289-10 - 1	SU-03-112917	-42	-0.0329	-	7470A	PPB	01 Dec 2017 11:19:01	U	SMPL
I6300-08 - 1	OR-02-112917	-30	-0.0295	-	7470A	PPB	01 Dec 2017 11:21:09	U	SMPL
I6615-03 - 1	PB-2-OW(0-10)	-24	-0.0278	-	7470A	PPB	01 Dec 2017 11:23:18	U	SMPL
I6616-03 - 1	P7-LOT-SP1-C	-9	-0.0235	-	7470A	PPB	01 Dec 2017 11:25:25	U	SMPL
CCV31 - 1	CCV31	17974	5.0988	-	7470A	PPB	01 Dec 2017 11:27:32	U	SMPL
CCB31 - 1	CCB31	-124	-0.0563	-	7470A	PPB	01 Dec 2017 11:29:39	U	SMPL
I6641-02 - 1	TP-1	-20	-0.0266	-	7470A	PPB	01 Dec 2017 11:31:47	U	SMPL
I6641-04 - 1	TP-2	-19	-0.0263	-	7470A	PPB	01 Dec 2017 11:33:54	U	SMPL
I6641-06 - 1	TP-3	-19	-0.0263	-	7470A	PPB	01 Dec 2017 11:36:01	U	SMPL
I6641-06DUP - 1	TP-3DUP	-13	-0.0246	-	7470A	PPB	01 Dec 2017 11:38:08	U	SMPL
I6641-06MS - 1	TP-3MS	12854	3.6404	-	7470A	PPB	01 Dec 2017 11:40:15	U	SMPL
I6641-06MSD - 1	TP-3MSD	12616	3.5726	-	7470A	PPB	01 Dec 2017 11:42:23	U	SMPL
I6641-06LX5 - 1		-112	-0.0528	-	7470A	PPB	01 Dec 2017 11:44:31	U	SMPL
I6641-06A - 1		12354	3.4980	-	7470A	PPB	01 Dec 2017 11:46:39	U	SMPL
CCV32 - 1	CCV32	18075	5.1275	-	7470A	PPB	01 Dec 2017 11:48:48	U	SMPL
CCB32 - 1	CCB32	-121	-0.0554	-	7470A	PPB	01 Dec 2017 11:50:55	U	SMPL