

245.1


LB 91888

INSTRUMENT ID: CV1

Linear

 μ Abs.:

32811



A= 0.0000e+000

B= 3.0416e-004

C= -1.1203e-002

Rho= 0.9999772

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.013	-0.013	-7	0.000	-7				
0.2	0.200	0.203	0.003	704	0.0 %	704				
2.5	2.500	2.508	0.008	8281	0.0 %	8281				
5.0	5.000	4.986	-0.014	16431	0.0 %	16431				
7.5	7.500	7.548	0.048	24853	0.0 %	24853				
10.0	10.000	9.968	-0.032	32811	0.0 %	32811				

LB91888 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 01 Dec 2017 14:33:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	-7	-	0.0000	245.1	PPB	01 Dec 2017 14:37:45	S	Std
0.2 - 1	S0.2	704	-	0.2000	245.1	PPB	01 Dec 2017 14:39:54	S	Std
2.5 - 1	S2.5	8281	-	2.5000	245.1	PPB	01 Dec 2017 14:42:02	S	Std
5.0 - 1	S5	16431	-	5.0000	245.1	PPB	01 Dec 2017 14:44:10	S	Std
7.5 - 1	S7.5	24853	-	7.5000	245.1	PPB	01 Dec 2017 14:46:18	S	Std
10.0 - 1	S10	32811	-	10.0000	245.1	PPB	01 Dec 2017 14:48:26	S	Std
ICV87 - 1	ICV87	13301	4.0344	-245.1		PPB	01 Dec 2017 14:51:26	U	SMPL
ICB87 - 1	ICB87	-107	-0.0437	-245.1		PPB	01 Dec 2017 14:53:33	U	SMPL
CCV33 - 1	CCV33	16793	5.0965	-245.1		PPB	01 Dec 2017 14:57:54	U	SMPL
CCB33 - 1	CCB33	-119	-0.0474	-245.1		PPB	01 Dec 2017 15:00:00	U	SMPL
CRA - 1	CRA	720	0.2078	-245.1		PPB	01 Dec 2017 15:02:07	U	SMPL
HighStd - 1	HighStd	32783	9.9600	-245.1		PPB	01 Dec 2017 15:04:14	U	SMPL
ChkStd - 1	ChkStd	23580	7.1608	-245.1		PPB	01 Dec 2017 15:06:21	U	SMPL
PB104626BL - 1	PBW	-174	-0.0641	-245.1		PPB	01 Dec 2017 15:08:30	U	SMPL
PB104626BS - 1	LCSW	13455	4.0812	-245.1		PPB	01 Dec 2017 15:10:39	U	SMPL
I6634-01 - 1	112717-JETT-E-D-B	-115	-0.0462	-245.1		PPB	01 Dec 2017 15:12:47	U	SMPL
I6634-02 - 1	112717-JETT-E-D-S	2	-0.0106	-245.1		PPB	01 Dec 2017 15:14:55	U	SMPL
I6634-03 - 1	112717-JETT-E-U-B	-12	-0.0149	-245.1		PPB	01 Dec 2017 15:17:03	U	SMPL
I6634-04 - 1	112717-JETT-E-U-S	-7	-0.0133	-245.1		PPB	01 Dec 2017 15:19:12	U	SMPL
I6634-05 - 1	112717-TIDO-E-D-B	6	-0.0094	-245.1		PPB	01 Dec 2017 15:21:19	U	SMPL
CCV34 - 1	CCV34	16629	5.0466	-245.1		PPB	01 Dec 2017 15:23:25	U	SMPL
CCB34 - 1	CCB34	-123	-0.0486	-245.1		PPB	01 Dec 2017 15:25:32	U	SMPL
I6634-06 - 1	112717-TIDO-E-D-S	-20	-0.0173	-245.1		PPB	01 Dec 2017 15:27:40	U	SMPL
I6634-07 - 1	112717-TIDO-E-U-B	-11	-0.0145	-245.1		PPB	01 Dec 2017 15:29:47	U	SMPL
I6634-08 - 1	112717-TIDO-E-U-S	-11	-0.0145	-245.1		PPB	01 Dec 2017 15:31:54	U	SMPL
I6635-01 - 1	112817-JETT-E-D-B	-12	-0.0149	-245.1		PPB	01 Dec 2017 15:34:01	U	SMPL
I6635-02 - 1	112817-JETT-E-D-S	9	-0.0085	-245.1		PPB	01 Dec 2017 15:36:08	U	SMPL
I6635-03 - 1	112817-JETT-E-U-B	-2	-0.0118	-245.1		PPB	01 Dec 2017 15:38:15	U	SMPL
I6635-04 - 1	112817-JETT-E-U-S	-16	-0.0161	-245.1		PPB	01 Dec 2017 15:40:23	U	SMPL
I6635-05 - 1	112817-TIDO-E-D-B	-18	-0.0167	-245.1		PPB	01 Dec 2017 15:42:32	U	SMPL
I6635-06 - 1	112817-TIDO-E-D-S	-12	-0.0149	-245.1		PPB	01 Dec 2017 15:44:40	U	SMPL
I6635-07 - 1	112817-TIDO-E-U-B	-5	-0.0127	-245.1		PPB	01 Dec 2017 15:46:47	U	SMPL
CCV35 - 1	CCV35	16598	5.0372	-245.1		PPB	01 Dec 2017 15:48:55	U	SMPL
CCB35 - 1	CCB35	-118	-0.0471	-245.1		PPB	01 Dec 2017 15:51:02	U	SMPL
I6635-08 - 1	112817-TIDO-E-U-S	-18	-0.0167	-245.1		PPB	01 Dec 2017 15:53:10	U	SMPL
I6636-01 - 1	112917-TIDO-E-D-B	17	-0.0060	-245.1		PPB	01 Dec 2017 15:55:17	U	SMPL
I6636-02 - 1	112917-TIDO-E-D-S	23	-0.0042	-245.1		PPB	01 Dec 2017 15:57:24	U	SMPL
I6636-03 - 1	112917-TIDO-E-U-B	53	0.0049	-245.1		PPB	01 Dec 2017 15:59:31	U	SMPL
I6636-04 - 1	112917-TIDO-E-U-S	54	0.0052	-245.1		PPB	01 Dec 2017 16:01:39	U	SMPL
I6636-04DUP - 1	112917-TIDO-E-U-SDUP	33	-0.0012	-245.1		PPB	01 Dec 2017 16:03:47	U	SMPL
I6636-04MS - 1	112917-TIDO-E-U-SMS	15648	4.7482	-245.1		PPB	01 Dec 2017 16:05:55	U	SMPL
I6636-04MSD - 1	112917-TIDO-E-U-SMSD	16029	4.8641	-245.1		PPB	01 Dec 2017 16:08:03	U	SMPL
I6636-04LX5 - 1		-122	-0.0483	-245.1		PPB	01 Dec 2017 16:10:12	U	SMPL
I6636-04A - 1		14156	4.2944	-245.1		PPB	01 Dec 2017 16:12:19	U	SMPL
CCV36 - 1	CCV36	16758	5.0859	-245.1		PPB	01 Dec 2017 16:14:26	U	SMPL
CCB36 - 1	CCB36	-141	-0.0541	-245.1		PPB	01 Dec 2017 16:16:33	U	SMPL