



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB91894

Temp.: 22.1 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/01/2017	17:35
2	CAL2	1	Water			6.99	12/01/2017	17:36
3	CAL3	1	Water			10.01	12/01/2017	17:37
4	ICV	1	Water			7.00	12/01/2017	17:39
5	CCV1	1	Water			2.01	12/01/2017	17:41
6	I6690-01	1	Water			6.58	12/01/2017	17:42
7	I6690-01DUP	1	Water			6.59	12/01/2017	17:43
8	I6701-01	1	Water			7.59	12/01/2017	17:45
9	CCV2	1	Water			12.01	12/01/2017	17:46

2391894

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I6701

WorkList ID : 106029

Date : 12/1/2017 5:28:20 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/01/2017	Water	I6690-01-F	pH	Cool 4 deg C	PSEG01	M52	BAYWAY-PHASE-3	12/01/2017	SM 4500-PH B
12/01/2017	Water	I6701-01-F	pH	Cool 4 deg C	PSEG01	M52	MOBILE-TRANSFORMER-CC	12/01/2017	SM 4500-PH B

Date/Time 12-01-17 5:30 PM
Received by: JP
Relinquished by: CP

Date/Time 12-01-17 6:25 PM
Received by: CP
Relinquished by: JP