



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91895

Temp.: 24.4 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 2.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/01/2017	17:35
2	CAL2	1	Water			7.00	12/01/2017	17:36
3	CAL3	1	Water			10.03	12/01/2017	17:37
4	ICV	1	Water			6.99	12/01/2017	17:38
5	CCV1	1	Water			2.01	12/01/2017	17:40
6	I6308-02	1	Solid	20.02	20	7.57	12/01/2017	17:41
7	I6308-02DUP	1	Solid	20.03	20	7.58	12/01/2017	17:42
8	I6659-04	1	Solid	20.02	20	6.40	12/01/2017	17:44
9	I6699-04	1	Solid	20.03	20	8.08	12/01/2017	17:45
10	I6700-01	1	Solid	20.05	20	5.16	12/01/2017	17:46
11	I6700-02	1	Solid	20.02	20	6.06	12/01/2017	17:47
12	I6700-03	1	Solid	20.01	20	5.65	12/01/2017	17:48
13	I6700-04	1	Solid	20.04	20	5.95	12/01/2017	17:49
14	I6700-05	1	Solid	20.03	20	6.05	12/01/2017	17:50
15	I6700-06	1	Solid	20.02	20	6.52	12/01/2017	17:51
16	CCV2	1	Water			2.01	12/01/2017	18:00
17	I6700-07	1	Solid	20.04	20	7.84	12/01/2017	18:01
18	I6700-08	1	Solid	20.02	20	6.73	12/01/2017	18:02
19	I6700-09	1	Solid	20.04	20	6.40	12/01/2017	18:03
20	I6700-10	1	Solid	20.02	20	9.41	12/01/2017	18:05

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	I6700-11	1	Solid	20.03	20	7.80	12/01/2017	18:06
22	I6700-12	1	Solid	20.02	20	9.02	12/01/2017	18:07
23	I6700-13	1	Solid	20.03	20	8.43	12/01/2017	18:09
24	I6700-14	1	Solid	20.02	20	8.93	12/01/2017	18:10
25	I6700-15	1	Solid	20.04	20	8.96	12/01/2017	18:11
26	CCV3	1	Water			12.02	12/01/2017	18:13

1391895

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6699

WorkList ID : 106027

Date : 12/1/2017 5:26:49 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/01/2017	Solid	I6308-02	Corrosivity	Cool 4 deg C	PSEG05	M11	SA-06-120117	12/01/2017	9045C
11/30/2017	Solid	I6659-04	Corrosivity	Cool 4 deg C	PSEG01	M52	TRANSFORMER-1-2	11/30/2017	9045C
12/01/2017	Solid	I6699-04	Corrosivity	Cool 4 deg C	PSEG01	M11	RT-3874	12/01/2017	9045C
12/01/2017	Solid	I6700-01	Corrosivity	Cool 4 deg C	PSEG01	M52	DRUM-1	12/01/2017	9045C
12/01/2017	Solid	I6700-02	Corrosivity	Cool 4 deg C	PSEG01	M52	DRUM-2	12/01/2017	9045C
12/01/2017	Solid	I6700-03	Corrosivity	Cool 4 deg C	PSEG01	M52	DRUM-3	12/01/2017	9045C
12/01/2017	Solid	I6700-04	Corrosivity	Cool 4 deg C	PSEG01	M52	DRUM-4	12/01/2017	9045C
12/01/2017	Solid	I6700-05	Corrosivity	Cool 4 deg C	PSEG01	M52	DRUM-6	12/01/2017	9045C
12/01/2017	Solid	I6700-06	Corrosivity	Cool 4 deg C	PSEG01	M52	DRUM-7	12/01/2017	9045C
12/01/2017	Solid	I6700-07	Corrosivity	Cool 4 deg C	PSEG01	M52	DRUM-8	12/01/2017	9045C
12/01/2017	Solid	I6700-08	Corrosivity	Cool 4 deg C	PSEG01	M52	DRUM-9	12/01/2017	9045C
12/01/2017	Solid	I6700-09	Corrosivity	Cool 4 deg C	PSEG01	M52	DRUM-10	12/01/2017	9045C
12/01/2017	Solid	I6700-10	Corrosivity	Cool 4 deg C	PSEG01	M52	DRUM-12	12/01/2017	9045C
12/01/2017	Solid	I6700-11	Corrosivity	Cool 4 deg C	PSEG01	M52	DRUM-13	12/01/2017	9045C
12/01/2017	Solid	I6700-12	Corrosivity	Cool 4 deg C	PSEG01	M52	DRUM-15	12/01/2017	9045C
12/01/2017	Solid	I6700-13	Corrosivity	Cool 4 deg C	PSEG01	M52	DRUM-16	12/01/2017	9045C
12/01/2017	Solid	I6700-14	Corrosivity	Cool 4 deg C	PSEG01	M52	DRUM-17	12/01/2017	9045C
12/01/2017	Solid	I6700-15	Corrosivity	Cool 4 deg C	PSEG01	M52	DRUM-18	12/01/2017	9045C

Date/Time 12-01-17 5:30 PM
 Received by: Jo
 Relinquished by: CR

Date/Time 12-01-17 7:00 PM
 Received by: CD
 Relinquished by: Jo