


7471A

LB91901
INSTRUMENT ID: CV 1

Linear

 μ Abs.:

4400E



A= 0.0000e+000

B= 2.2988e-004

C= 4.9347e-003

Rho= 0.9996639

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.024	0.024	83	0.000	83				
0.2	0.200	0.198	-0.002	838	0.0 %	838				
2.5	2.500	2.525	0.025	10964	0.0 %	10964				
5.0	5.000	5.027	0.027	21848	0.0 %	21848				
7.5	7.500	7.304	-0.196	31753	0.0 %	31753	0			
10.0	10.000	10.121	0.121	44006	0.0 %	44006				

LB91901 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 04 Dec 2017 11:19:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	83	-	0.0000	7471A	PPB	04 Dec 2017 11:26:34	S	Std
0.2 - 1	SO-2	838	-	0.2000	7471A	PPB	04 Dec 2017 11:28:51	S	Std
2.5 - 1	SO-5	10964	-	2.5000	7471A	PPB	04 Dec 2017 11:31:09	S	Std
5.0 - 1	SO-5	21848	-	5.0000	7471A	PPB	04 Dec 2017 11:33:27	S	Std
7.5 - 1	SO-5	31753	-	7.5000	7471A	PPB	04 Dec 2017 11:35:44	S	Std
10.0 - 1	SO-5	44006	-	10.0000	7471A	PPB	04 Dec 2017 11:38:01	S	Std
ICV88 - 1	ICV88	16922	3.8950	-	7471A	PPB	04 Dec 2017 11:40:44	U	SMPL
ICB88 - 1	ICB88	-69	-0.0109	-	7471A	PPB	04 Dec 2017 11:43:00	U	SMPL
CCV37 - 1	CCV37	21224	4.8840	-	7471A	PPB	04 Dec 2017 11:45:16	U	SMPL
CCB37 - 1	CCB37	-61	-0.0091	-	7471A	PPB	04 Dec 2017 11:47:32	U	SMPL
CRA - 1	CRA	831	0.1960	-	7471A	PPB	04 Dec 2017 11:49:49	U	SMPL
HighStd - 1	HighStd	44345	10.1991	-	7471A	PPB	04 Dec 2017 11:52:05	U	SMPL
ChkStd - 1	ChkStd	30790	7.0830	-	7471A	PPB	04 Dec 2017 11:54:21	U	SMPL
PB104662BL - 1	PBS	-217	-0.0450	-	7471A	PPB	04 Dec 2017 11:56:40	U	SMPL
PB104662BS - 1	LCSS	16343	3.7619	-	7471A	PPB	04 Dec 2017 11:58:58	U	SMPL
I6642-01 - 1	PENN-1-E(0-1)	113954	26.2011	-	7471A	PPB	04 Dec 2017 12:06:08	U	SMPL
I6642-01DUP - 1	PENN-1-E(0-1)DUP	114839	26.4045	-	7471A	PPB	04 Dec 2017 12:11:02	U	SMPL
I6642-01MS - 1	PENN-1-E(0-1)MS	129039	29.6689	-	7471A	PPB	04 Dec 2017 12:13:20	U	SMPL
I6642-01MSD - 1	PENN-1-E(0-1)MSD	129308	29.7307	-	7471A	PPB	04 Dec 2017 12:15:41	U	SMPL
I6642-02 - 1	PENN-1-E(4-5)	1184	0.2771	-	7471A	PPB	04 Dec 2017 12:18:04	U	SMPL
CCV38 - 1	CCV38	21584	4.9667	-	7471A	PPB	04 Dec 2017 12:20:27	U	SMPL
CCB38 - 1	CCB38	-100	-0.0181	-	7471A	PPB	04 Dec 2017 12:22:43	U	SMPL
I6642-03 - 1	PENN-1-E(7-8)	358	0.0872	-	7471A	PPB	04 Dec 2017 12:25:00	U	SMPL
I6645-01 - 1	BSC-SB-04	618	0.1470	-	7471A	PPB	04 Dec 2017 12:27:15	U	SMPL
I6645-02 - 1	BSC-SB-02	391	0.0948	-	7471A	PPB	04 Dec 2017 12:29:31	U	SMPL
I6645-03 - 1	BSC-SB-01	388	0.0941	-	7471A	PPB	04 Dec 2017 12:31:48	U	SMPL
I6656-01 - 1	BART-4-E(0-1)	2516	0.5833	-	7471A	PPB	04 Dec 2017 12:34:04	U	SMPL
I6656-02 - 1	BART-4-E(4-5)	16417	3.7789	-	7471A	PPB	04 Dec 2017 12:36:21	U	SMPL
I6656-03 - 1	BART-4-E(7-8)	16470	3.7911	-	7471A	PPB	04 Dec 2017 12:38:38	U	SMPL
I6656-04 - 1	BART-3-E(0-1)	14912	3.4330	-	7471A	PPB	04 Dec 2017 12:40:56	U	SMPL
I6656-05 - 1	BART-3-E(7-8)	9906	2.2822	-	7471A	PPB	04 Dec 2017 12:43:14	U	SMPL
I6656-06 - 1	BART-2-E(0-1)	20071	4.6189	-	7471A	PPB	04 Dec 2017 12:45:31	U	SMPL
CCV39 - 1	CCV39	21589	4.9679	-	7471A	PPB	04 Dec 2017 12:47:47	U	SMPL
CCB39 - 1	CCB39	-67	-0.0105	-	7471A	PPB	04 Dec 2017 12:50:03	U	SMPL
I6656-07 - 1	BART-2-E(7-8)	8700	2.0049	-	7471A	PPB	04 Dec 2017 12:52:20	U	SMPL
I6656-08 - 1	BART-1-E(0-1)	1572	0.3663	-	7471A	PPB	04 Dec 2017 12:54:37	U	SMPL
I6656-09 - 1	BART-1-E(4-5)	1684	0.3921	-	7471A	PPB	04 Dec 2017 12:56:54	U	SMPL
I6659-01 - 1	TRANSFORMER-1-2	1170	0.2739	-	7471A	PPB	04 Dec 2017 12:59:10	U	SMPL
I6664-01 - 1	EME-S	943	0.2217	-	7471A	PPB	04 Dec 2017 13:01:27	U	SMPL
I6664-02 - 1	EME-C	6002	1.3847	-	7471A	PPB	04 Dec 2017 13:03:44	U	SMPL
PB104663BL - 1	PBS	-44	-0.0052	-	7471A	PPB	04 Dec 2017 13:06:02	U	SMPL
PB104663BS - 1	LCSS	16405	3.7762	-	7471A	PPB	04 Dec 2017 13:08:19	U	SMPL
I6671-01 - 1	ENV-106-1(5-10)	9174	2.1139	-	7471A	PPB	04 Dec 2017 13:10:37	U	SMPL
I6671-02 - 1	ENV-106-1(10-15)	1566	0.3649	-	7471A	PPB	04 Dec 2017 13:12:53	U	SMPL
CCV40 - 1	CCV40	21794	5.0150	-	7471A	PPB	04 Dec 2017 13:15:10	U	SMPL
CCB40 - 1	CCB40	-60	-0.0089	-	7471A	PPB	04 Dec 2017 13:17:26	U	SMPL
I6671-03 - 1	ENV-106-1(15-20)	47	0.0157	-	7471A	PPB	04 Dec 2017 13:19:43	U	SMPL
I6671-04 - 1	ENV-106-1(20-30)	58	0.0183	-	7471A	PPB	04 Dec 2017 13:21:59	U	SMPL
I6671-05 - 1	ENV-106-1(30-40)	69	0.0208	-	7471A	PPB	04 Dec 2017 13:24:16	U	SMPL
I6671-06 - 1	ENV-106-1(40-50)	154	0.0403	-	7471A	PPB	04 Dec 2017 13:26:32	U	SMPL
I6671-06DUP - 1	ENV-106-1(40-50)DUP	154	0.0403	-	7471A	PPB	04 Dec 2017 13:28:49	U	SMPL
I6671-06MS - 1	ENV-106-1(40-50)MS	16491	3.7959	-	7471A	PPB	04 Dec 2017 13:31:06	U	SMPL
I6671-06MSD - 1	ENV-106-1(40-50)MSD	16229	3.7357	-	7471A	PPB	04 Dec 2017 13:33:23	U	SMPL
I6671-07 - 1	ENV-106-1(50-60)	37	0.0134	-	7471A	PPB	04 Dec 2017 13:35:41	U	SMPL
I6671-08 - 1	ENV-105-3(0-5)	4376	1.0109	-	7471A	PPB	04 Dec 2017 13:37:59	U	SMPL
I6671-09 - 1	ENV-105-3(5-10)	249	0.0622	-	7471A	PPB	04 Dec 2017 13:40:15	U	SMPL
CCV41 - 1	CCV41	22154	5.0978	-	7471A	PPB	04 Dec 2017 13:42:32	U	SMPL
CCB41 - 1	CCB41	-54	-0.0075	-	7471A	PPB	04 Dec 2017 13:44:48	U	SMPL
I6671-10 - 1	ENV-105-3(10-15)	536	0.1282	-	7471A	PPB	04 Dec 2017 13:47:04	U	SMPL
I6642-01LX5 - 1		23076	5.3097	-	7471A	PPB	04 Dec 2017 13:49:21	U	SMPL
I6642-01A - 1		129585	29.7944	-	7471A	PPB	04 Dec 2017 13:51:37	U	SMPL
I6671-06LX5 - 1		-94	-0.0167	-	7471A	PPB	04 Dec 2017 13:56:54	U	SMPL
I6671-06A - 1		17297	3.9812	-	7471A	PPB	04 Dec 2017 13:59:11	U	SMPL
CCV42 - 1	CCV42	21801	5.0166	-	7471A	PPB	04 Dec 2017 14:01:27	U	SMPL
CCB42 - 1	CCB42	-56	-0.0079	-	7471A	PPB	04 Dec 2017 14:03:45	U	SMPL
I6642-01DLX5 - 1		23348	5.3723	-	7471A	PPB	04 Dec 2017 14:06:03	U	SMPL
I6642-01DUPDLX5 - 1		22955	5.2819	-	7471A	PPB	04 Dec 2017 14:08:21	U	SMPL
I6642-01MSDLX5 - 1		27641	6.3591	-	7471A	PPB	04 Dec 2017 14:10:38	U	SMPL
I6642-01MSDDLX5 - 1		24876	5.7235	-	7471A	PPB	04 Dec 2017 14:12:56	U	SMPL
I6642-01LDLX25 - 1		4493	1.0378	-	7471A	PPB	04 Dec 2017 14:15:13	U	SMPL
I6642-01ADLX5 - 1		26302	6.0513	-	7471A	PPB	04 Dec 2017 14:17:31	U	SMPL
CCV43 - 1	CCV43	21661	4.9844	-	7471A	PPB	04 Dec 2017 14:19:48	U	SMPL
CCB43 - 1	CCB43	-74	-0.0121	-	7471A	PPB	04 Dec 2017 14:22:05	U	SMPL