



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: HIRAL

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB91924

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	12/04/2017	09:15
2	I6690-02		1	>100.0	>212.0	12/04/2017	09:40
3	I6690-02DUP		1	>100.0	>212.0	12/04/2017	10:02
4	I6701-02		1	>100.0	>212.0	12/04/2017	10:32

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP12417

WorkList ID : 106044

Date : 12/4/2017 8:18:36 AM

LB 91924

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/06/2017	Water	I6690-02	Flash Point	Cool 4 deg C	PSEG01	M52	BAYWAY-PHASE-3	12/01/2017	1010A
12/06/2017	Water	I6701-02	Flash Point	Cool 4 deg C	PSEG01	M52	MOBILE-TRANSFORMER-CC	12/01/2017	1010A

Date/Time 12/4/17 8:25 AM
 Received by: KS
 Relinquished by: CP

Date/Time 12/4/17 10:40 AM
 Received by: CP
 Relinquished by: KS