



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB91927

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/04/2017	17:30
2	CAL2	1	Water			7.01	12/04/2017	17:31
3	CAL3	1	Water			10.03	12/04/2017	17:32
4	ICV	1	Water			6.99	12/04/2017	17:33
5	CCV1	1	Water			2.01	12/04/2017	17:34
6	I6728-01	1	Solid	20.02	20	7.10	12/04/2017	17:35
7	I6728-01DUP	1	Solid	20.03	20	7.11	12/04/2017	17:36
8	I6728-02	1	Solid	20.04	20	7.08	12/04/2017	17:37
9	I6728-03	1	Solid	20.02	20	7.56	12/04/2017	17:39
10	CCV2	1	Water			12.01	12/04/2017	17:40

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I6728

WorkList ID : 106085

Date : 12/4/2017 5:12:03 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/17/2017	Solid	I6728-01 ✓	pH	Cool 4 deg C	ECHE01	M62	LOCATION-1	10/17/2017	9045C
10/17/2017	Solid	I6728-02 ✓	pH	Cool 4 deg C	ECHE01	M62	LOCATION-2	10/17/2017	9045C
10/17/2017	Solid	I6728-03 ✓	pH	Cool 4 deg C	ECHE01	M62	LOCATION-3	10/17/2017	9045C

Date/Time 12-04-17 5:12 PM

Received by: JP

Relinquished by: JP

Date/Time 12-04-17 6:30 PM

Received by: CP

Relinquished by: JP