



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91928

Temp.: 19.8 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/04/2017	18:34
2	CAL2	1	Water			6.99	12/04/2017	18:35
3	CAL3	1	Water			10.03	12/04/2017	18:36
4	ICV	1	Water			7.00	12/04/2017	18:37
5	CCV1	1	Water			2.02	12/04/2017	18:38
6	I6726-02	1	Water			7.96	12/04/2017	18:39
7	I6726-02DUP	1	Water			7.97	12/04/2017	18:40
8	CCV2	1	Water			12.01	12/04/2017	18:41

LB 91928

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I6726

WorkList ID : 106089

Date : 12/4/2017 6:32:45 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/04/2017	Water	I6726-02	Corrosivity -G	Cool 4 deg C	PSEG01	M52	256434	12/04/2017	9040C

Date/Time 12-04-17 6:32PM

Received by: JH

Relinquished by: JH

Date/Time 12-04-17 7:30PM

Received by: JP

Relinquished by: JP