



## Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB91929

Temp.: 19.8 °C

SLOPE.: 98.1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2270         |
| buffer solution pH 7 yellow    | W2159         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2154         |
| buffer solution pH 7 yellow    | W2219         |
| pH 2.00 Buffer                 | W2289         |
| Buffer Solution, PH12 (500ml)  | W2267         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  |             |             | 4.01        | 12/04/2017 | 18:34     |
| 2   | CAL2        | 1  | Water  |             |             | 6.99        | 12/04/2017 | 18:35     |
| 3   | CAL3        | 1  | Water  |             |             | 10.03       | 12/04/2017 | 18:36     |
| 4   | ICV         | 1  | Water  |             |             | 7.00        | 12/04/2017 | 18:37     |
| 5   | CCV1        | 1  | Water  |             |             | 2.02        | 12/04/2017 | 18:38     |
| 6   | I6726-01    | 1  | Water  |             |             | 7.96        | 12/04/2017 | 18:39     |
| 7   | I6726-01DUP | 1  | Water  |             |             | 7.97        | 12/04/2017 | 18:40     |
| 8   | CCV2        | 1  | Water  |             |             | 12.01       | 12/04/2017 | 18:41     |

287 91929

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I6726

WorkList ID : 106090

Date : 12/4/2017 6:33:22 PM

| Due Date   | Matrix | Sample     | Test | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|------------|------|--------------|----------|------------------|-----------------|--------------|--------|
| 12/04/2017 | Water  | I6726-01-G | pH   | Cool 4 deg C | PSEG01   | M52              | 256434          | 12/04/2017   | 9045C  |

Date/Time 12.04.17 6:33:22 PM  
Received by: JV  
Relinquished by: CP

Date/Time 12.04.17 7:30:23  
Received by: CP  
Relinquished by: JV