

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 12/05/2017

DATE & TIME IN: 12/05/2017 10:15

OVEN TEMP: 103 °C

Run Number: LB91934

DATE & TIME OUT: 12/05/2017 14:15

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB91934BL	LB91934BL	1.4162	1.4162	100	1.4162	1.4162	1.4162	0	0
2	LB91934BS	LB91934BS	1.414	1.414	100	1.4687	1.4687	1.4687	0.0547	547
3	I6726-01	256434	1.4139	1.4139	20	1.477	1.477	1.477	0.0631	3155
4	I6731-01	113017-TIDO-E-D-B	1.4079	1.4079	300	1.4255	1.4255	1.4255	0.0176	58.7
5	I6731-02	113017-TIDO-E-D-S	1.4027	1.4027	400	1.4138	1.4138	1.4138	0.0111	27.8
6	I6731-02DUP	113017-TIDO-E-D-SDUP	1.4185	1.4185	400	1.4292	1.4292	1.4292	0.0107	26.8
7	I6731-03	113017-TIDO-E-U-B	1.4025	1.4025	400	1.4161	1.4161	1.4161	0.0136	34
8	I6731-04	113017-TIDO-E-U-S	1.4152	1.4152	500	1.4259	1.4259	1.4259	0.0107	21.4
9	I6732-01	120117-TIDO-F-D-B	1.4175	1.4175	400	1.4322	1.4322	1.4322	0.0147	36.8
10	I6732-02	120117-TIDO-F-D-S	1.412	1.412	400	1.4224	1.4224	1.4224	0.0104	26
11	I6732-03	120117-TIDO-F-U-B	1.4078	1.4078	300	1.424	1.424	1.424	0.0162	54
12	I6732-04	120117-TIDO-F-U-S	1.418	1.418	400	1.428	1.428	1.428	0.01	25
13	I6732-05	120117-FOPI-E-D-M	1.4084	1.4084	300	1.4308	1.4308	1.4308	0.0224	74.7
14	I6732-06	120117-FOPI-E-U-M	1.4188	1.4188	300	1.4356	1.4356	1.4356	0.0168	56

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS120517

WorkList ID : 106114

Date : 12/5/2017 8:40:22 AM

12/5/17 10:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/07/2017	Water	I6726-01	TSS	Cool 4 deg C	PSEG01	M52	256434	12/04/2017	SM2540 D
12/07/2017	Water	I6731-01	TSS	Cool 4 deg C	HDR103	O11	113017-TIDO-E-D-B	11/30/2017	SM2540 D
12/07/2017	Water	I6731-02	TSS	Cool 4 deg C	HDR103	O11	113017-TIDO-E-D-S	11/30/2017	SM2540 D
12/07/2017	Water	I6731-03	TSS	Cool 4 deg C	HDR103	O11	113017-TIDO-E-U-B	11/30/2017	SM2540 D
12/07/2017	Water	I6731-04	TSS	Cool 4 deg C	HDR103	O11	113017-TIDO-E-U-S	11/30/2017	SM2540 D
12/08/2017	Water	I6732-01	TSS	Cool 4 deg C	HDR103	O31	120117-TIDO-F-D-B	12/01/2017	SM2540 D
12/08/2017	Water	I6732-02	TSS	Cool 4 deg C	HDR103	O31	120117-TIDO-F-D-S	12/01/2017	SM2540 D
12/08/2017	Water	I6732-03	TSS	Cool 4 deg C	HDR103	O31	120117-TIDO-F-U-B	12/01/2017	SM2540 D
12/08/2017	Water	I6732-04	TSS	Cool 4 deg C	HDR103	O31	120117-TIDO-F-U-S	12/01/2017	SM2540 D
12/08/2017	Water	I6732-05	TSS	Cool 4 deg C	HDR103	O31	120117-FOPI-E-D-M	12/01/2017	SM2540 D
12/08/2017	Water	I6732-06	TSS	Cool 4 deg C	HDR103	O31	120117-FOPI-E-U-M	12/01/2017	SM2540 D

Date/Time 12/5/17, 9:20
 Received by: CP AS
 Relinquished by: CP AS

Date/Time 12/5/17 10:30
 Received by: CP AS
 Relinquished by: AS