



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91940

Temp.: 24.6 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 2.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/04/2017	18:40
2	CAL2	1	Water			7.01	12/04/2017	18:41
3	CAL3	1	Water			10.02	12/04/2017	18:42
4	ICV	1	Water			7.00	12/04/2017	18:43
5	CCV1	1	Water			2.02	12/04/2017	18:45
6	I6676-02	1	Solid	20.02	20	8.94	12/04/2017	18:46
7	I6676-02DUP	1	Solid	20.03	20	8.95	12/04/2017	18:47
8	I6705-03	1	Solid	20.01	20	6.85	12/04/2017	18:48
9	I6717-05	1	Solid	20.03	20	8.99	12/04/2017	18:50
10	I6717-06	1	Solid	20.02	20	9.28	12/04/2017	18:51
11	I6718-03	1	Solid	20.04	20	7.82	12/04/2017	19:00
12	I6718-06	1	Solid	20.02	20	8.40	12/04/2017	19:01
13	I6718-09	1	Solid	20.03	20	6.70	12/04/2017	19:02
14	I6718-12	1	Solid	20.05	20	6.55	12/04/2017	19:05
15	I6718-15	1	Solid	20.05	20	7.70	12/04/2017	19:06
16	CCV2	1	Water			2.02	12/04/2017	19:07
17	I6718-18	1	Solid	20.02	20	7.78	12/04/2017	19:09
18	I6721-05	1	Solid	20.04	20	5.60	12/04/2017	19:10
19	I6725-04	1	Solid	20.02	20	7.39	12/04/2017	19:11
20	I6725-05	1	Solid	20.03	20	6.06	12/04/2017	19:12

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	I6725-06	1	Solid	20.05	20	6.30	12/04/2017	19:13
22	I6725-07	1	Solid	20.04	20	6.38	12/04/2017	19:15
23	I6733-01	1	Solid	20.02	20	7.30	12/04/2017	19:16
24	I6734-02	1	Solid	20.03	20	6.99	12/04/2017	19:18
25	I6734-04	1	Solid	20.02	20	6.67	12/04/2017	19:20
26	I6734-04DUP	1	Solid	20.03	20	6.68	12/04/2017	19:21
27	CCV3	1	Water			12.01	12/04/2017	19:22

LB91940

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6734

WorkList ID : 106088

Date : 12/4/2017 6:31:29 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/04/2017	Solid	I6676-02	Corrosivity	Cool 4 deg C	PSEG05	M52	OK-02-120417	12/04/2017	9045C
12/04/2017	Solid	I6679-02	Corrosivity	Cool 4 deg C	PSEG05	M52	HR-02-120417	12/04/2017	9045C
11/30/2017	Solid	I6705-03	Corrosivity	Cool 4 deg C	WSP001	O41	B12(0-10)COMP	11/30/2017	9045C
12/01/2017	Solid	I6717-05	Corrosivity	Cool 4 deg C	WSP001	O22	PB-3(0-10)	12/01/2017	9045C
12/01/2017	Solid	I6717-06	Corrosivity	Cool 4 deg C	WSP001	O22	PB-3(10-20)	12/01/2017	9045C
12/04/2017	Solid	I6718-03	Corrosivity	Cool 4 deg C	PSEG01	M11	2041	12/04/2017	9045C
12/04/2017	Solid	I6718-06	Corrosivity	Cool 4 deg C	PSEG01	M11	RT-3992	12/04/2017	9045C
12/04/2017	Solid	I6718-09	Corrosivity	Cool 4 deg C	PSEG01	M11	RT-4651	12/04/2017	9045C
12/04/2017	Solid	I6718-12	Corrosivity	Cool 4 deg C	PSEG01	M11	RT-1500	12/04/2017	9045C
12/04/2017	Solid	I6718-15	Corrosivity	Cool 4 deg C	PSEG01	M11	RT-3026	12/04/2017	9045C
12/04/2017	Solid	I6718-18	Corrosivity	Cool 4 deg C	PSEG01	M11	RT-2898	12/04/2017	9045C
12/04/2017	Solid	I6721-05	Corrosivity	Cool 4 deg C	PSEG01	M52	TOTE-1-2	12/04/2017	9045C
11/28/2017	Solid	I6725-04	Corrosivity	Cool 4 deg C	ROYF02	M11	P001-TPI1-1218-01	11/28/2017	9045C
11/30/2017	Solid	I6725-05	Corrosivity	Cool 4 deg C	ROYF02	M11	P001-TPI1-SW01	11/30/2017	9045C
11/30/2017	Solid	I6725-06	Corrosivity	Cool 4 deg C	ROYF02	M11	P001-TPI1-SW02	11/30/2017	9045C
11/30/2017	Solid	I6725-07	Corrosivity	Cool 4 deg C	ROYF02	M11	P001-TPI1-SW03	11/30/2017	9045C
12/04/2017	Solid	I6733-01	Corrosivity	Cool 4 deg C	PSEG01	M52	OILY-SPEEDY-DRY	12/04/2017	9045C
12/04/2017	Solid	I6734-02	Corrosivity	Cool 4 deg C	PSEG01	M52	Driveway-Praving-1(0-12)	12/04/2017	9045C
12/04/2017	Solid	I6734-04	Corrosivity	Cool 4 deg C	PSEG01	M52	Driveway-Praving-1(0-12)	12/04/2017	9045C

Date/Time 12.04.17 6.31 PM
 Received by: JD
 Relinquished by: CR

Date/Time 12.04.17 6.31 PM
 Received by: CR
 Relinquished by: JP