



Analytical Summary Report

Analysis Method: SM2510 B

ANALYST: AMANDEEP

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB91949

Reagent/Standard	Lot/Log #
conductivity standard	WP59787

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB91949BL		1	23.8	0.09	0.09	12/05/2017	11:53
2	LB91949BS	1413	1	23.7	1342.00	1376.17	12/05/2017	11:59
3	I6729-01		1	24.8	2320.00	2328.90	12/05/2017	12:05

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$

CB91949

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CONDUCTANCE12517 WorkList ID : 106130 Date : 12/5/2017 10:10:08 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/05/2017	Water	I6729-01	Conductance	Cool 4 deg C	ECHE01	M62	C193-50	11/18/2017	SM2510 B

Date/Time 12/05/17 11:30
Received by: AC
Relinquished by: CP

Date/Time 12/05/17 12:20
Received by: CR
Relinquished by: AC