



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB91959

Temp.: 20.2 °C

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/05/2017	16:20
2	CAL2	1	Water			6.99	12/05/2017	16:21
3	CAL3	1	Water			10.03	12/05/2017	16:22
4	ICV	1	Water			7.00	12/05/2017	16:23
5	CCV1	1	Water			2.02	12/05/2017	16:25
6	I6742-01	1	Water			6.68	12/05/2017	16:26
7	I6742-01DUP	1	Water			6.69	12/05/2017	16:27
8	CCV2	1	Water			12.01	12/05/2017	16:30

1391959

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I6742

WorkList ID : 106156

Date : 12/5/2017 4:07:27 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/05/2017	Water	I6742-01 -F	pH	Cool 4 deg C	PSEG01	M52	TOTE-1-4	12/05/2017	SM 4500-PH B

Date/Time 12.05.17 4:07 PM
Received by: JP
Relinquished by: CP

Date/Time 12.05.17 5:00 PM
Received by: CP
Relinquished by: JP