



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91963

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/05/2017	16:20
2	CAL2	1	Water			6.99	12/05/2017	16:21
3	CAL3	1	Water			10.03	12/05/2017	16:22
4	ICV	1	Water			7.00	12/05/2017	16:23
5	CCV1	1	Water			2.02	12/05/2017	17:00
6	I6303-02	1	Solid	20.02	20	9.12	12/05/2017	17:01
7	I6675-02	1	Solid	20.04	20	7.24	12/05/2017	17:02
8	I6683-02	1	Solid	20.02	20	8.65	12/05/2017	17:03
9	I6683-04	1	Solid	20.05	20	9.20	12/05/2017	17:05
10	I6683-06	1	Solid	20.03	20	7.88	12/05/2017	17:06
11	I6685-02	1	Solid	20.01	20	6.92	12/05/2017	17:08
12	I6685-04	1	Solid	20.03	20	8.63	12/05/2017	17:10
13	I6738-03	1	Solid	20.02	20	5.58	12/05/2017	17:11
14	I6741-02	1	Solid	20.04	20	9.04	12/05/2017	17:12
15	I6741-02DUP	1	Solid	20.05	20	9.05	12/05/2017	17:13
16	CCV2	1	Water			12.01	12/05/2017	17:15

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6675

WorkList ID : 106160

Date : 12/5/2017 4:09:03 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/04/2017	Solid	I6303-02	Corrosivity	Cool 4 deg C	PSEG05	M52	HD-02-120417	12/04/2017	9045C
12/05/2017	Solid	I6675-02	Corrosivity	Cool 4 deg C	PSEG05	M52	SU-04-120517	12/05/2017	9045C
12/05/2017	Solid	I6683-02	Corrosivity	Cool 4 deg C	PSEG05	M52	EO-01-120517-A	12/05/2017	9045C
12/05/2017	Solid	I6683-04	Corrosivity	Cool 4 deg C	PSEG05	M52	EO-01-120517-B	12/05/2017	9045C
12/05/2017	Solid	I6683-06	Corrosivity	Cool 4 deg C	PSEG05	M52	EO-01-120517-C	12/05/2017	9045C
12/05/2017	Solid	I6685-02	Corrosivity	Cool 4 deg C	PSEG05	M52	AU-06-120517-A	12/05/2017	9045C
12/05/2017	Solid	I6685-04	Corrosivity	Cool 4 deg C	PSEG05	M52	AU-06-120517-B	12/05/2017	9045C
12/05/2017	Solid	I6738-03	Corrosivity	Cool 4 deg C	PSEG01	M52	88-20-90-20	12/05/2017	9045C
12/05/2017	Solid	I6741-02	Corrosivity	Cool 4 deg C	PSEG01	M52	DRUM-1-7	12/05/2017	9045C

Date/Time 12-05-17 4:09 PM
 Received by: JP
 Relinquished by: CA

Date/Time 12-05-17 6:30 PM
 Received by: CA
 Relinquished by: JP